



KENYATTA NATIONAL HOSPITAL
P.O. BOX 20723, 00202 Nairobi

Tel.: 2726300/2726450/2726550

Fax: 2725272

Email: knhadmin@knh.or.ke

Date: 9th August, 2017

Ref: KNH/SCM/ADM.33/3.

TO ALL BIDDERS

RE: ADDENDUM TO TENDER NO: KNH/T/60/2017-2018 SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF CSSD/TSSU AUTOCLAVES AND MEDICAL GRADE AIR COMPRESSOR

Pursuant to Clause 2.6 on Amendment of Tender Documents, the Hospital wishes to make the following amendments/Clarifications as indicated here below.

1. SECTION VI - TECHNICAL SPECIFICATIONS page 41 of the tender document

1. Item no. 7 Capacity should be between 800- 1000 liters, 3STU, has been Expunged And Replaced By:-

- a. Should be 12 STU Pass through Autoclave of not less than 800 litres capacity
- b. With double stage vacuum pump
- c. Steam generator should be equipped with an integrated water visual indicator
- d. Chamber, should be double mantle and steam generator should be entirely made of corrosion free stainless steel, 1.4571/AISI 316Ti
- e. Vendors to note that chamber should be made of AISI 316 Ti corrosion free stainless steel
- f. All pipes exposed to steam should be made of 1.4404/AISI 316L stainless steel and isolated particle free

2. Item No. 30 Factory training: Should have training component for end users and facility/Hospitals Technicians (2 operators, 2 Nursing staff, and 2 Biomedical Technicians/Technologist) Has Been Expunged And Replaced By:-

On-site training for operators and maintenance training for biomedical technologists within the country by a factory application expert

Please note the Tender Closing and Opening date remain as earlier communicated (16th August 2017 10:00Hrs).

**ROSE NJOROGE MRS
FOR: CHIEF EXECUTIVE OFFICER**