



KENYATTA NATIONAL HOSPITAL  
P.O. BOX 20723, 00202 Nairobi

Tel.: 2726300/2726450/2726550

Fax: 2725272

Email: [knhadmin@knh.or.ke](mailto:knhadmin@knh.or.ke)

Date: 11th August, 2017

Ref: KNH/SCM/ADM.33/4.

TO ALL BIDDERS

**RE: ADDENDUM TO TENDER NO: KNH/T/60/2017-2018 SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF CSSD/TSSU AUTOCLAVES AND MEDICAL GRADE AIR COMPRESSOR**

Pursuant to Clause 2.6 on Amendment of Tender Documents, the Hospital wishes to make the following amendments/Clarifications as indicated here below.

**1. SECTION VI - TECHNICAL SPECIFICATIONS page 41 of the tender document**

- i. The above tender shall be awarded as **ONE LOT**
- ii. **The awarded bidder shall;-**
  - a. Decommission, de-install and hand over the hospital the three existing autoclaves as they are.
  - b. Supply, delivery, installation, testing and commissioning of two autoclaves in TSSU space.
  - c. Supplier to lay terrazzo floor
  - d. Install all the required piping for the floor
  - e. Replace ceiling with non-porous ceiling material suitable for sterile area
  - f. To provide Air lock space and erect sterile barriers
  - g. To provide adequate lighting
  - h. To separate entries and exits for soiled, clean and sterile goods
  - i. To provide adequate room ventilation

Please note the Tender Closing and Opening date has been extended to **23rd August 2017 10:00Hrs.**

  
f  
ROSE NJOROGE MRS  
FOR: CHIEF EXECUTIVE OFFICER