



KENYATTA NATIONAL HOSPITAL
P.O. BOX 20723, 00202 Nairobi

Tel.: 2726300/2726450/2726550
Email: procurement@knh.or.ke

Date: November 16, 2022

Ref: KNH/SCM/ADM.43

ADDENDUM 1

RE: ADDENDUM TO TENDER NO: KNH/T/83/2022-2024, SUPPLY AND DELIVERY OF SURGICAL DRESSINGS AND APPLIANCES

To: ALL BIDDERS

Pursuant to the Public Procurement and Asset Disposal Act section 75, its attendant Regulations 2020 and Clause 5 on Amendment of Tender Documents, the Hospital wishes to make the following amendments/clarifications;

1. ADDITIONAL ITEMS

Technical Specifications	
Sterilization Steam Reel (Steam)	
Sizes:	
1. 75mm*200m	
2. 150mm*200m	
3. 250mm*200m	
4. 100mm*200m	
5. 300mm*200m	
Additional Information:	
1. Size as per order.	
2. Must have chemicals impregnated that change color on exposure to steam.	
3. Must have paper backing which is water resistant.	
4. The Top must be transparent plastic material which can stand high temperature of up to 150°C.	
5. Must be a continuous reel of 200m length.	
6. Sample must be provided.	
7. Date of manufacture and expiry must be indicated.	
8. Must have two years and above shelf life on delivery	

Sterilization Reel (Steam)								
No.	Item Description	Unit of Issue	Qty M/H	Qty P/Wing	Unit Cost	Total Cost	Delivery Period	Remarks
139. (a)	Sterilization Reel (Steam) - Sizes 75mm*200m	No	1	1				
139. (b)	Sterilization Reel (Steam) - Sizes 150mm*200m	No	1	1				
139. (c)	Sterilization Reel (Steam) - Sizes 250mm*200m	No	1	1				

Sterilization Reel (Steam)								
No.	Item Description	Unit of Issue	Qty M/H	Qty P/Wing	Unit Cost	Total Cost	Delivery Period	Remarks
139. (d)	Sterilization Reel (Steam) - Sizes 1000mm*200m	No	1	1				
139. (e)	Sterilization Reel (Steam) - Sizes 300mm*200m	No	1	1				

Technical Specifications								
Polythene Sheet (Pkts of 10)								
<ol style="list-style-type: none"> 1. Should be yellow in color. 2. Gauge should be 200mm. 3. Size should be 2m by 3m 4. Should be 10 pieces per pack. 5. Sample must be provided. 								

No	Item Description	Unit of Issue	Qty M/H	Qty P/Wing	Unit Cost	Total Cost	Delivery Period	Remarks
140.	Polythene Sheet	Pkts of 10	1	1				

2. CHANGE OF CLOSING DATE

The Tender closing date will be on 29th November, 2022 at 10:00am East African Time.

Original information literature, complete and in English language where applicable should be properly bound in the Tender Document and submitted before the Tender closing date 29th November, 2022 at 10:00Am East African Time



Rose Njoroge (Mrs)
FOR: CHIEF EXECUTIVE OFFICER