



KENYATTA NATIONAL HOSPITAL
P.O. BOX 20723, 00202 Nairobi

Tel.: 2726300/2726450/2726550
Email: procurement@knh.or.ke

Date: February 27, 2023

Ref: KNH/SCM/ADM.43

ADDENDUM 2

RE: ADDENDUM TO TENDER NO: KNH/T/84/2022-2023, SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF MEDICAL EQUIPMENT.

To: ALL BIDDERS

Pursuant to the Public Procurement and Asset Disposal Act section 75, its attendant Regulations 2020 and Clause 5 on Amendment of Tender Documents, the Hospital wishes to make the following amendments/clarifications;

1. CHANGE OF NEUROSURGERY OPERATING TABLE SPECIFICATIONS

The technical specifications captured in the tender document and published in the website have been expunged and replaced as indicated below:

NEUROSURGERY OPERATING TABLE SPECIFICATIONS

1. The table must be electro-hydraulic with manual override
2. Should have the following table top dimensions:
 - Table top length of at least 1960mm
 - Width of at least 550mm
 - Maximum table top total weight capacity up to 450kg
3. Must be height adjustable from 680-1180mm and have a longitudinal shift of 300mm
4. Should have the following power specifications:
 - Power supply 220-240 Volts
 - Consumption 600-700VA
 - Frequency 50-60Hz
 - Battery power DC 24 Volts
5. Must achieve the following inclination angles:
 - 30 degree Trendelenburg
 - 30 degree reverse Trendelenburg
 - Back section -40 degrees to +70 degrees
 - Head section -60 degrees to + 90 degrees
 - Lateral adjustment of +/-20 degrees on either side
6. Must have a full radiolucent table top for C-arm use
7. Must have a medical emergency manual override of all hydraulic functions
8. Must have brakes in all 4 casters
9. Should have a hand control unit and a foot control unit with Bluetooth function for both.
10. Should have the following accessories:
 - Detachable head plate with double adjustment
 - One section detachable leg section
 - An inclination back section extension with downward tilt up to 40 degrees

- 2 Arm rests with foam pad and antistatic cover, an strap, attachment clamp and made of stainless steel. Adjustable to any direction
 - Body restraint straps with stainless steel attachment elements which are washable with hook and loop fasteners X3
 - A Bluetooth integration module
 - A stainless steel universal adaptor for head clamp
 - A stainless steel fixture with central square mounting and three joint adjustments for holding horseshoe head rest. Compatible with universal adaptor
 - 2 section horseshoe head rest with detachable pad and restraint strap
 - Stainless steel crossbar attachment for seated positioning
 - Head clamp attachment to adaptor and clamp adaptor
 - Skull clamp with 3 point fixation
 - Skull clamp pins adult and paediatric 2 sets of each
 - Semi-circular cushion with antistatic covered foam pad 240x120x475mm x3
 - Adult size open gel head ring 200x54mm
 - Paediatric size open gel head ring 140x34mm
 - Adult size closed gel head ring 200x50mm
 - Paediatric size closed gel head ring 140x34mm
 - Gel head rest for Prone positioning adult size 280x240x150mm and paediatric size 260x230x140mm
11. Theatre table must made of stainless steel
 12. Must provide onsite training for users and technicians
 13. Table must come with a complete tool kit
 14. Must have a 2 year warranty
 15. Must provide original manufacturer's brochure
 16. Service and repair facilities must be available locally
 17. Must have ownership
 18. Company must have manufacturer's authorization
 19. Supplier must be ready to provide an alternative table should it need to go for repair outside the hospital

Original information literature, complete and in English language where applicable should be properly bound in the Tender Document and submitted.

NOTE: THE TENDER OPENING AND CLOSING DATE REMAINS AS COMMUNICATED ON 9TH MARCH, 2023 AT 10:00AM EAST AFRICAN TIME.



Rose M. Njoroge, OGW
FOR: CHIEF EXECUTIVE OFFICER