



KENYATTA NATIONAL HOSPITAL  
P.O. BOX 20723, 00202 Nairobi

Tel.: 2726300/2726450/2726550  
Email: [procurement@knh.or.ke](mailto:procurement@knh.or.ke)

Date: March 08, 2023

Ref: KNH/SCM/ADM.43

ADDENDUM 3

RE: ADDENDUM TO TENDER NO: KNH/T/84/2022-2023, SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF MEDICAL EQUIPMENT.

To: ALL BIDDERS

Pursuant to the Public Procurement and Asset Disposal Act section 75, its attendant Regulations 2020 and Clause 5 on Amendment of Tender Documents, the Hospital wishes to make the following amendments/clarifications;

**1. CLARIFICATION OF VENTILATOR SPECIFICATIONS**

The time taken to rinse at the various specified flow rate and at the selected mode eg. Adult, Paediatric have been indicated as shown below:

Rinse Time		
Adult Water Trap		
Numbering	Time	Flow Rate
9.1	400 ms	@70ml/min
9.2	330 ms	@100ml/min
9.3	300 ms	@120ml/min
9.4	24 ms	@150ml/min

**2. CHANGE OF OPENING DATE**

Following to the clarification the tender Opening and Closing date was indicated as 9<sup>th</sup> March, 2023 at 10:00am East African Time. The new date is 16<sup>th</sup> March, 2023 at 10:00am East African Time.

Original information literature, complete and in English language where applicable should be properly bound in the Tender Document and submitted.

Rose M. Njoroge, OGW  
FOR: CHIEF EXECUTIVE OFFICER