



INTERNAL MEMO

Kenyatta National Hospital, P. O. Box 20723-00202, Tel: 2726300-9

OFFICE OF THE DIRECTOR, SUPPLY CHAIN MANAGEMENT

Ref: KNH/SCM/ADM/18

To: All Bidders in Quotation No. KNH/Q/12/CS/MMUH/2023-2024

Date: 19/9/2023

RE: ADDENDUM ON MEASUREMENTS ON QUOTATION NO.
KNH/Q/12/CS/MMUH/2023-2024

The above-mentioned subject refers.

This is to clarify that the correct specification for the Stainless-Steel mortuary stretcher trays are as follows:

1. Length 78inches (1,981.2mm) by 21inches (533.4mm)
2. Weight Capacity 300Lbs
3. Hospital Grade Stainless Steel Gauge SWG 18mm
4. The Stainless Steel should be Easily Cleanable, Free from Corrosion and be able to withstand Low Temperatures

The previous specified width of 23inches has been expunged. The closing date remains the same i.e. 21st September 2023.

Rose M. Njoroge OGW

Director, Supply Chain Management
Kenyatta National Hospital
