



KENYATTA NATIONAL HOSPITAL
P.O. BOX 20723, 00202 Nairobi

Tel.2726300/2726450/2726550
Fax: 2725272
Email: knhadmin@knh.or.ke

Ref: KNH/SCM/ADM.43

Date: 24th May, 2024

RE: ADDENDUM TO TENDER NO: KNH/T/03/2024-2026- FRAMEWORK CONTRACT FOR SUPPLY AND DELIVERY OF LABORATORY REAGENTS & CONSUMABLES.

To: ALL BIDDERS

Pursuant to the Public Procurement and Asset Disposal Act 2015 section 75 and its attendant Regulations 2020 and Clause 7 on Amendment of Tender Documents, the Hospital wishes to make the following amendments/clarifications;

1. TECHNICAL SPECIFICATIONS FOR A PLACED GLUCOSE METER WITH STRIPS

The specifications for the Glucose Meter with Strips are as tabulated below:

CRITERIA	SPECIFICATIONS/REQUIREMENTS
General Descriptions	<ul style="list-style-type: none">i. Hand-held instrumentii. Capable of blood glucose level measurement on multiple patientsiii. Able to withstand repeated disinfection for operator and patient safety
Test Methodology	Electrochemistry
Sample types	Venous, arterial, capillary and neonatal
Sample Volume	0.5 - 15 μ L
Duration of test analysis	3 - 10 Seconds
Measurement Units	mmol/L
Measurement Range	0.5 - 38 mmol/L
Calibration	Plasma equivalent traceable to a higher standard
Data Storage capacity	Total for both patient & QC Tests: 350 to at least 2000
Operational features	<ul style="list-style-type: none">Automatic test strip/ sample detection and analysisInterference correction: Oxygen, HematocritStrip ejection mechanism (Safety)Quality control material: Minimum 2 levelsQC prompting or QC lock outAutomatic shut-off when not in useCritical value alert capabilityAutomatic sample counterNon-coding test strips

	CRITERIA	SPECIFICATIONS/REQUIREMENTS
		Data transfer capability
	Battery Information	Display: LCD touch screen with patient demographic data entry capability i. Replaceable or rechargeable (minimum battery life 500 tests) ii. Replace at supplier's cost (accessory) iii. Any reported faulty glucose meter will be replaced at the supplier's cost
	International Standards	FDA approve or CE mark of approval
	Training	Onsite training for the technical personnel
	Testing and handover	i. Sample instrument to come with at least a control ii. Selected equipment to be delivered with a packet/s of fifty (50) strips for verification before transition.

Tender Opening and Closing date **REMAINS** as earlier communicated i.e. 28th May, 2024.



Linet Adhiambo
FOR: CHIEF EXECUTIVE OFFICER