



KENYATTA NATIONAL HOSPITAL  
Tel.2726300/2726450/2726550  
P.O. BOX 20723, 00202 Nairobi

Fax: 2725272  
Email: [knhadmin@knh.or.ke](mailto:knhadmin@knh.or.ke)

Ref: KNH/SCM/ADM.43

Date: 23<sup>rd</sup> January, 2025

---

## ADDENDUM 2

RE: ADDENDUM TO TENDER NO: KNH/T/115/2024-2025 - SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF MEDICAL EQUIPMENT- LIVER TRANSPLANT PROJECT.

---

To: ALL BIDDERS

Pursuant to the Public Procurement and Asset Disposal Act 2015 section 75 and its attendant Regulations 2020 and Clause 7 on Amendment of Tender Documents, the Hospital wishes to make the following amendments/clarifications;

### GENERAL SURGERY OPERATING TABLES.

The following lithotomy stirrups specifications have been added as follows;

- The stirrups should be available in two sizes: for adults with a safe working load (SWL) of minimum 180 kg and Floating Plus for bariatric patients with a SWL of 200- 250 kg.
- It should have clearly visible scale simplifies symmetrical alignment.
- The table should have Floating lithotomy stirrups designed for effortless and optimal lithotomy positioning.
- It should have Non-Velcro silicone straps support improved infection control.
- It should have flexible boot to accommodate larger patient calves.
- It should have pressure-relieving padding to reduce the risk of pressure injuries and compartment syndrome.

Please note that the tender closing/opening date remains as earlier communicated i.e.30<sup>th</sup> January 2025 at 10:00am Kenyan time.

  
Simon Waithaka.  
FOR: CHIEF EXECUTIVE OFFICER