

ORIGINAL

Kenyatta National Hospital



Tender

Document For

SUPPLY AND DELIVERY OF CALIBRATION TOOLS

(Reserved for Youth, Women and People Living with Disability)

TENDER NO: KNH/43/2026- 2027

**The Chief Executive Officer
Kenyatta National Hospital
P.O Box 20723 – 00202
Nairobi**

Closing and opening Date: 29th June, 2026 @10:00am

APPENDIX TO THE PREFACE	iv
GUIDELINES FOR PREPARATION OF TENDER DOCUMENTS	iv
1. GENERAL	iv
2. PART 1 - TENDERING PROCEDURES.....	iv
3. PART 2 – PROCUREMENT ENTITY'S REQUIREMENTS.....	v
4. PART 3 – CONDITIONS OF CONTRACT AND CONTRACT FORMS.....	v
INVITATION TO TENDER.....	vii
Section I - Instructions to Tenderers	1
A General Provisions	1
1 Scope of Tender	1
.....
2 Fraud and Corruption.....	1
3 Eligible Tenderers.....	1
4 Eligible Goods and Related Services	3
B Contents of Request for Tenders Document.....	3
5. Sections of Tendering Document.....	3
6 Clarification of Tender Document.....	4
7 Amendment of Tender Document.....	4
C Preparation of Tenders.....	5
8 Cost of Tendering	5
9 Language of Tender	5
10 Documents Comprising the Tender.....	5
11 Form of Tender and Price Schedules.....	5
12 Alternative Tenders.....	6
13 Tender Prices and Discounts.....	6
14 Currencies of Tender and Payment	7
15 Documents Establishing the Eligibility and Conformity of the Goods and Related Services	7
16 Documents Establishing the Eligibility and Qualifications of the Tenderer.....	7
17 Period of Validity of Tenders.....	8
18 Tender Security.....	8
19 Format and Signing of Tender	9
D Submission and Opening of Tenders	9
20 Sealing and Marking of Tenders	9
21 Deadline for Submission of Tenders	10
22 Late Tenders	10
23 Withdrawal, Substitution, and Modification of Tenders.....	10
24 Tender Opening	10
E Evaluation and Comparison of Tenders.....	11
25 Confidentiality	11
26 Clarification of Tenders.....	11
27 Deviations, Reservations, and Omissions	12
28 Determination of Responsiveness	12
29 Non-conformities, Errors and Omissions	12
30 Correction of Arithmetical Errors	13
31 Conversion to Single Currency	13
32 Margin of Preference and Reservations.....	13
33 Evaluation of Tenders	13
34 Comparison of Tenders.....	14
35 Abnormally Low Tenders	14
36 Abnormally High Tenders.....	14
37. Post Qualification of the Tenderer.....	15



KENYATTA NATIONAL HOSPITAL

**THE CHIEF EXECUTIVE OFFICER
KENYATTA NATIONAL HOSPITAL
P.O BOX 20723- 00202,
NAIROBI.**

Email: procurement@knh.or.ke; procurementknh@gmail.com

Tender No: KNH/T/43/2026- 2027

**TENDER NAME: SUPPLY AND DELIVERY OF CALIBRATION TOOLS
(Reserved for Youth, Women and People Living with Disability)**

1.I

INVITATION TO TENDER

2.

P

PROCURING ENTITY: KENYATTA NATIONAL HOSPITAL PO BOX 20723- 00202 NAIROBI

KENYA

3.

CONTRACT NAME AND DESCRIPTION: SUPPLY AND DELIVERY OF CALIBRATION TOOLS (Reserved for Youth, Women and People Living with Disability)

KENYATTA NATIONAL HOSPITAL invites eligible bidders for the SUPPLY AND DELIVERY OF CALIBRATION TOOLS (Reserved for Youth, Women and People Living with Disability)

Tendering will be conducted under open competitive method National using a standardized tender document. Tendering is open to all qualified and interested Tenderers.

In case this tender is subject to a Reservation, specify the Group is eligible to tender, Insert e.g. "Tendering is open to all Small and Medium Enterprises registered

In case tender is subject to Multiple contracts/lots, insert "Tenderers will be allowed to tender for one or more lots". (Not applicable).

4.

Qualified and interested tenderers may obtain further information and inspect the Tender Documents during weekdays and office working hours [0900 to 1600 hours] at the address given below.

**THE CHIEF EXECUTIVE OFFICER
KENYATTA NATIONAL HOSPITAL
P.O BOX 20723- 00202,
NAIROBI.**

Email: procurement@knh.or.ke; procurementknh@gmail.com

5.

A complete set of tender documents may be purchased or obtained by interested tenders upon payment of a non-refundable fees of *Kenya shillings 1000* in cash or Banker's Cheque and payable to the address given below. Tender documents may be obtained electronically from the Website: www.knh.or.ke. Tender documents obtained electronically will be free of charge.

6.

Tender documents may be viewed and downloaded for free from the website : www.knh.or.ke Tenderers who download the tender document must forward their particulars immediately to facilitate any further clarification or addendum.

**DIRECTOR SUPPLY CHAIN MANAGEMENT
KENYATTA NATIONAL HOSPITAL
P.O BOX 20723- 00202,
NAIROBI.**

Email: procurement@knh.or.ke; procurementknh@gmail.com

7.

The Tenderer shall chronologically serialize all pages of the tender documents submitted.

8.

Completed tenders must be delivered to the address below on or before **29th June 2026 at 10:00 am**

9. Electronic Tenders will not be permitted.
10. Tenders will be opened immediately after the deadline date and time specified above or any dead line date and time specified later. Tenders will be publicly opened in the presence of the Tenderers' designated representatives who choose to attend at the address below.

**THE CHIEF EXXECUTIVE OFFICER
KENYATTA NATIONAL HOSPITAL,
Upper hill, off Hospital Road
Administration block, Supply Chain Management Entrance,
P.O BOX 20723- 00202, NAIROBI.
Email: procurement@knh.or.ke; procurementknh@gmail.com**

11. Late tenders will be rejected.
12. The addresses referred to above are:
- a. Address for obtaining further information and for purchasing tender documents**
- i. Kenyatta National Hospital
 - ii. Physical address: Nairobi City, Hospital Road, Kenyatta National Hospital , Administration Block, Supply Chain Management Division , contracts office room No. 6
 - iii. P.O. Box 20723-00202 Nairobi
 - iv. Director supply Chain management, Tel. 2726300, procurementknh@gmail.com or procurement@knh.or.ke.

A. Address for Submission of Tenders.

- 1) Kenyatta National Hospital
- 2) Director, Supply Chain Management
- 3) Nairobi City, Hospital Road, Kenyatta National Hospital, Administration Block , Supply Chain Management Division, Contracts office room No.6

B. Address for Opening of Tenders.

- 1) Kenyatta National Hospital
 - 2) Nairobi City, Hospital, Kenyatta National Hospital, Administration Block, Supply Chain Management Division, Contracts office room No.6
-

PART 1 – TENDERING PROCEDURES

SECTION I: INSTRUCTIONS TO TENDERERS

A General Provisions

1. Scope of Tender

The Procuring Entity as defined in the **TDS** invites tenders for supply of goods and, if applicable, any Related Services incidental thereto, as specified in Section V, Supply Requirements. The name, identification, and number of lots (contracts) of this Tender Document are specified in the **TDS**.

Throughout this tendering document:

- a) the term “in writing” means communicated in written form (e.g. by mail, e-mail, fax, including if specified in the **TDS**, distributed or received through the electronic-procurement system used by the Procuring Entity) with proof of receipt;
- b) if the context so requires, “singular” means “plural” and vice versa;
- c) “Day” means calendar day, unless otherwise specified as “Business Day”. A Business Day is any day that is an official working day of the Procuring Entity. It excludes official public holidays.

2. Fraud and Corruption

The Procuring Entity requires compliance with the provisions of the Public Procurement and Asset Disposal Act, 2015, Section 62 “Declaration not to engage in corruption”. The tender submitted by a person shall include a declaration that the person shall not engage in any corrupt or fraudulent practice and a declaration that the person or his or her sub-contractors are not debarred from participating in public procurement proceedings.

The Procuring Entity requires compliance with the provisions of the Competition Act 2010, regarding collusive practices in contracting. Any tenderer found to have engaged in collusive conduct shall be disqualified and criminal and/or civil sanctions may be imposed. To this effect, Tenders shall be required to complete and sign the “Certificate of Independent Tender Determination” annexed to the Form of Tender.

Unfair Competitive Advantage – Fairness and transparency in the tender process require that the firms or their Affiliates competing for a specific assignment do not derive a competitive advantage from having provided consulting services related to this tender. To that end, the Procuring Entity shall indicate in the **Data Sheet** and make available to all the firms together with this tender document all information that would in that respect give such firm any unfair competitive advantage over competing firms.

3. Eligible Tenderers

A Tenderer may be a firm that is a private entity, an individual, a state-owned enterprise or institution subject to ITT3.7, or any combination of such entities in the form of a joint venture (JV) under an existing agreement or with the intent to enter into such an agreement supported by a letter of intent. Public employees and their close relatives (*spouses, children, brothers, sisters and uncles and aunts*) are not eligible to participate in the tender.

In the case of a joint venture, all members shall be jointly and severally liable for the execution of the entire Contract in accordance with the Contract terms. The JV shall nominate a Representative who shall have the authority to conduct all business for and on behalf of any and all the members of the JV during the Tendering process and, in the event the JV is awarded the Contract, during contract execution. The maximum number of JV members shall be specified in the **TDS**.

Public Officers of the Procuring Entity, their Spouses, Child, Parent, Brothers or Sister. Child, Parent, Brother or Sister of a Spouse their business associates or agents and firms/organizations in which they have a substantial or controlling interest shall not be eligible to tender or be awarded a contract. Public Officers are also not allowed to participate in any procurement proceedings.

A Tenderer shall not have a conflict of interest. Any Tenderer found to have a conflict of interest shall be disqualified. A Tenderer may be considered to have a conflict of interest for the purpose of this Tendering process, if the Tenderer:

- a) directly or indirectly controls, is controlled by or is under common control with another Tenderer; or
- b) receives or has received any direct or indirect subsidy from another Tenderer; or
- c) has the same – representative or ownership as another Tenderer; or
- d) has a relationship with another Tenderer, directly or through common third parties, that puts it in a

position to influence the Tender of another Tenderer, or influence the decisions of the Procuring Entity regarding this Tendering process; or

- e) or any of its affiliates participated as a consultant in the preparation of the design or technical specifications of the goods that are the subject of the Tender; or
- f) or any of its affiliates has been hired (or is proposed to be hired) by the Procuring Entity or Procuring Entity for the Contract implementation; or
- g) would be providing goods, works, or non-consulting services resulting from or directly related to consulting services for the preparation or implementation of the project specified in the **TDS** ITT 1.1 that it provided or were provided by any affiliate that directly or indirectly controls, is controlled by, or is under common control with that firm; or has a close business or family relationship with a professional staff of the Procuring Entity (or of the project implementing agency, who: (i) are directly or indirectly involved in the preparation of the tendering document or specifications of the Contract, and/or the Tender evaluation process of such Contract; or (ii) would be involved in the implementation or supervision of such Contract unless the conflict stemming from such relationship has been resolved in a manner acceptable to the Procuring Entity throughout the Tendering process and execution of the Contract.

A tenderer shall not be involved in corrupt, coercive, obstructive, collusive or fraudulent practice. A tenderer that is proven to have been involved in any of these practices shall be automatically disqualified.

A firm that is a Tenderer (either individually or as a JV member) shall not submit more than one Tender, except for permitted alternative Tenders. This includes participation as a subcontractor. Such participation shall result in the disqualification of all Tenders in which the firm is involved. A firm that is not a Tenderer or a JV member, may participate as a subcontractor in more than one Tender. Members of a joint venture may not also make an individual tender, be a subcontractor in a separate tender or be part of another joint venture for the purposes of the same Tender.

A Tenderer may have the nationality of any country, subject to the restrictions pursuant to ITT3.9. A Tenderer shall be deemed to have the nationality of a country if the Tenderer is constituted, incorporated or registered in and operates in conformity with the provisions of the laws of that country, as evidenced by its articles of incorporation (or equivalent documents of constitution or association) and its registration documents, as the case may be. This criterion also shall apply to the determination of the nationality of proposed subcontractors or sub consultants for any part of the Contract including related Services.

A Tenderer that has been debarred by the PPRA from participating in public procurement shall be ineligible to tender or be awarded a contract. The list of debarred firms and individuals is available from the PPRA's website www.ppra.go.ke

Tenderers that are state-owned enterprises or institutions may be eligible to compete and be awarded a Contract(s) only if they are (i) a legal public entity of the state Government and/or public administration, (ii) financially autonomous and not receiving any significant subsidies or budget support from any public entity or Government, and (iii) operating under commercial law and vested with legal rights and liabilities similar to any commercial enterprise to enable it compete with firms in the private sector on an equal basis. Public employees and their close relatives are not eligible to participate in the tender.

Tenderers may be ineligible if their countries of origin (a) as a matter of law or official regulations, Kenya prohibits commercial relations with that country, or (b) by an act of compliance with a decision of the United Nations Security Council taken under Chapter VII of the Charter of the United Nations, Kenya prohibits any import of goods or contracting for supply of goods or services from that country, or any payments to any country, person, or entity in that country. A tenderer shall provide such documentary evidence of eligibility satisfactory to the Procuring Entity, as the Procuring Entity shall reasonably request.

Tenderers shall provide the qualification information statement that the tenderer (including all members of a joint venture and subcontractors) is not associated, or have been associated in the past, directly or indirectly, with a firm or any of its affiliates which have been engaged by the Procuring entity to provide consulting services for the preparation of the design, specifications, and other documents to be used for the procurement of the goods under this Invitation for tenders.

3.1 Where the law requires tenderers to be registered with certain authorities in Kenya, such registration requirements shall be defined in the **TDS**

The Competition Act of Kenya requires that firms wishing to tender as Joint Venture undertakings which may prevent, distort or lessen competition in provision of services are prohibited unless they are exempt in accordance with the provisions of Section 25 of the Competition Act, 2010. JVs will be required to seek for exemption from the Competition Authority. Exemption shall not be a condition for tender, but it shall be a condition of contract award and signature. A JV tenderer shall be given opportunity to seek such exemption as a condition of award and signature of contract. Application for exemption from the Competition Authority of Kenya may be accessed from the website www.cak.go.ke.

A Kenyan tenderer shall provide evidence of having fulfilled his/her tax obligations by producing a current tax clearance certificate or tax exemption certificate issued by the Kenya Revenue Authority.

4. Eligible Goods and Related Services

All the Goods and Related Services to be supplied under the Contract shall have their origin in any country that is eligible in accordance with ITT 3.9.

For purposes of this ITT, the term “goods” includes commodities, raw material, machinery, equipment, and industrial plants; and “related services” include services such as insurance, installation, training, and initial maintenance.

The term “origin” means the country where the goods have been mined, grown, cultivated, produced, manufactured or processed; or, through manufacture, processing, or assembly, another commercially recognized article results that differs substantially in its basic characteristics from its components.

A procuring entity shall ensure that the items listed below shall be sourced from Kenya and there shall be no substitutions from foreign sources. The affected items are:

- a) motor vehicles, plant and equipment which are assembled in Kenya;
- b) furniture, textile, foodstuffs, oil and gas, information communication technology, steel, cement, leather, agro-processed products, sanitary products, and other goods made in Kenya; or
- c) goods manufactured, mined, extracted or grown in Kenya.

Any goods, works and production processes with characteristics that have been declared by the relevant national environmental protection agency or by other competent authority as harmful to human beings and to the environment shall not be eligible for procurement.

5. Sections of Tendering Document

The tendering document consist of Parts 1, 2, and 3, which include all the sections indicated below, and should be read in conjunction with any Addenda issued in accordance with ITT8.

PART 1: Tendering Procedures

- i) Section I – Instructions to Tenderers (ITT)
- ii) Section II – Tendering Data Sheet (TDS)
- iii) Section III – Evaluation and Qualification Criteria
- iv) Section IV – Tendering Forms

PART 2: Supply Requirements

- v) Section V – Schedule of Requirements

PART 3: Contract

- vi) Section VI – General Conditions of Contract (GCC)
- vii) Section VII – Special Conditions of Contract (SCC)
- viii) Section VIII- Contract Forms

The notice of Invitation to Tender or the notice to the prequalified Tenderers issued by the Procuring Entity is not part of the tendering document.

Unless obtained directly from the Procuring Entity, the Procuring Entity is not responsible for the completeness of the document, responses to requests for clarification, the minutes of the pre-tender meeting (if any), or addenda to the tendering document in accordance with ITT7.

The Tenderer is expected to examine all instructions, forms, terms, and specifications in the tendering document and to furnish with its Tender all information or documentation as is required by the tendering document.

6. Clarification of Tendering Document

A Tenderer requiring any clarification of the Tender Document shall contact the Procuring Entity in writing at the Procuring Entity’s address specified in the **TDS** or raise its enquiries during the pre-Tender meeting if provided for in accordance with ITT 6.4. The Procuring Entity will respond in writing to any request for clarification, provided that such request is received no later than the period specified in the **TDS** prior to the deadline for submission of tenders. The Procuring Entity shall forward copies of its response to all tenderers who have acquired the Tender documents in accordance with ITT 5.3, including a description of the inquiry but without identifying its source. If so specified in the **TDS**, the Procuring Entity shall also promptly publish its response at the web page identified in the **TDS**. Should the clarification result in changes to the

essential elements of the Tender Documents, the Procuring Entity shall amend the Tender Documents following the procedure under ITT 7.

The Procuring Entity shall specify in the **TDS** if a pre-tender conference will be held, when and where. The Tenderer's designated representative is invited to attend a pre-Tender meeting. The purpose of the meeting will be to clarify issues and to answer questions on any matter that may be raised at that stage.

The Tenderer is requested to submit any questions in writing, to reach the Procuring Entity not later than the period specified in the **TDS** before the meeting.

Minutes of the pre-Tender meeting, if applicable, including the text of the questions asked by Tenderers and the responses given, together with any responses prepared after the meeting, will be transmitted promptly to all Tenderers who have acquired the Tender Documents in accordance with ITT 6.3. Minutes shall not identify the source of the questions asked.

The Procuring Entity shall also promptly publish anonymized (*no names*) Minutes of the pre-Tender meeting at the web page identified in the **TDS**. Any modification to the Tender Documents that may become necessary as a result of the pre-Tender meeting shall be made by the Procuring Entity exclusively through the issue of an Addendum pursuant to ITT 7 and not through the minutes of the pre-Tender meeting. Nonattendance at the pre-Tender meeting will not be a cause for disqualification of a Tenderer.

7. Amendment of Tendering Document

At any time prior to the deadline for submission of Tenders, the Procuring Entity may amend the tendering document by issuing addenda.

Any addendum issued shall be part of the tendering document and shall be communicated in writing to all who have obtained the tender document from the Procuring Entity in accordance with ITT 6.3. The Procuring Entity shall also promptly publish the addendum on the Procuring Entity's web page in accordance with ITT 7.1.

To give prospective Tenderers reasonable time in which to take an addendum into account in preparing their Tenders, the Procuring Entity may, at its discretion, extend the deadline for the submission of Tenders, pursuant to ITT 21.2.

C. Preparation of Tenders

8. Cost of Tendering

The Tenderer shall bear all costs associated with the preparation and submission of its Tender, and the Procuring Entity shall not be responsible or liable for those costs, regardless of the conduct or outcome of the Tendering process.

9. Language of Tender

The Tender, as well as all correspondence and documents relating to the Tender exchanged by the Tenderer and the Procuring Entity, shall be written in English Language. Supporting documents and printed literature that are part of the Tender may be in another language provided they are accompanied by an accurate translation of the relevant passages into the English Language, in which case, for purposes of interpretation of the Tender, such translation shall govern.

10. Documents Comprising the Tender

The Tender shall comprise the following:

- a) Form of Tender prepared in accordance with ITT11;
- b) Price Schedules: completed in accordance with ITT 11 and ITT 13;
- c) Tender Security or Tender-Securing Declaration, in accordance with ITT 18.1;
- d) Alternative Tender: if permissible, in accordance with ITT12;
- e) Authorization: written confirmation authorizing the signatory of the Tender to commit the Tenderer, in accordance with ITT19.3;
- f) Qualifications: documentary evidence in accordance with ITT 16.2 establishing the Tenderer qualifications to perform the Contract if its Tender is accepted;

- g) Tenderer Eligibility: documentary evidence in accordance with ITT16.1 establishing the Tenderer eligibility to tender;
- h) Eligibility of Goods and Related Services: documentary evidence in accordance with ITT 15, establishing the eligibility of the Goods and Related Services to be supplied by the Tenderer;
- i) Conformity: documentary evidence in accordance with ITT15.2 that the Goods and Related Services conform to the tender document; and
- j) any other document required in the **TDS**.

In addition to the requirements under ITT 10.1, Tenders submitted by a **JV** shall include a copy of the Joint Venture Agreement entered into by all members. Alternatively, a letter of intent to execute a Joint Venture Agreement in the event of a successful Tender shall be signed by all members and submitted with the Tender, together with a copy of the proposed Agreement.

The Tenderer shall furnish in the Form of Tender information on commissions gratuities, and fees, if any, paid or to be paid to agents or any other party relating to this Tender.

11. Form of Tender and Price Schedules

The Form of Tender and Price Schedules shall be prepared using the relevant forms furnished in Section IV, Tendering Forms. The forms must be completed without any alterations to the text. All blank spaces shall be filled in with the information requested. The Tenderer shall chronologically serialize pages of all tender documents submitted.

12. Alternative Tenders

Unless otherwise specified **in the TDS**, alternative Tenders shall not be considered.

13. Tender Prices and discounts

The prices quoted by the Tenderer in the Form of Tender and in the Price, Schedules shall conform to the requirements specified below.

All lots (contracts) and items must be listed and priced separately in the Price Schedules.

The price to be quoted in the Form of Tender in accordance with ITT10.1 shall be the total price of the Tender, including any discounts offered.

The Tenderer shall quote any discounts and indicate the methodology for their application in the form of tender. Conditional discounts will be rejected.

Prices quoted by the Tenderer shall be fixed during the performance of the Contract and not subject to variation on any account, unless otherwise specified **in the TDS**. A Tender submitted with an adjustable price quotation shall be treated as non-responsive and shall be rejected, pursuant to ITT 28. However, if in accordance with **the TDS**, prices quoted by the Tenderer shall be subject to adjustment during the performance of the Contract, a Tender submitted with a fixed price quotation shall not be rejected, but the price adjustment shall be treated as zero.

If specified in ITT 1.1, Tenders are being invited for individual lots (contracts) or for any combination of lots (packages). Unless otherwise specified **in the TDS**, prices quoted shall correspond to 100 % of the items specified for each lot and to 100% of the quantities specified for each item of a lot. Tenderers wishing to offer discounts for the award of more than one Contract shall specify in their Tender the price reductions applicable to each package, or alternatively, to individual Contracts within the package. Discounts shall be submitted in accordance with ITT 13.4 provided the Tenders for all lots (contracts) are opened at the same time.

The terms EXW, CIP, CIF, DDP and other similar terms shall be governed by the rules prescribed in the current edition of Incoterms, published by the International Chamber of Commerce.

Prices shall be quoted as specified in each Price Schedule included in Section IV, Tendering Forms. The disaggregation of price components is required solely for the purpose of facilitating the comparison of Tenders by the Procuring Entity. This shall not in any way limit the Procuring Entity's right to contract on any of the terms offered. In quoting prices, the Tenderer shall be free to use transportation through carriers registered in any eligible country. Similarly, the Tenderer may obtain insurance services from any eligible country in accordance with ITT 3.6, Eligible Tenders. Prices shall be entered in the following manner:

- a) For Goods manufactured in Kenya:

- i) the price of the Goods quoted EXW (ex-works, ex-factory, ex warehouse, ex showroom, or off-the-shelf, as applicable) final destination point indicated in the **TDS**, including all customs duties and sales and other taxes already paid or payable on the components and raw material used in the manufacture or assembly of the Goods;
 - ii) any sales tax and other taxes which will be payable in Kenya on the Goods if the Contract is awarded to the Tenderer; and
 - iii) the price for inland transportation, insurance, and other local services required to convey the Goods to their final destination specified **in the TDS**.
- b) For Goods manufactured outside Kenya, to be imported:
- i) the price of the Goods, quoted CIP named place of destination, in Kenya, as specified **in the TDS**;
 - ii) the price for inland transportation, insurance, and other local services required to convey the Goods from the named place of destination to their final destination specified **in the TDS**;
- c) For Goods manufactured outside Kenya, already imported:
- i) the price of the Goods, including the original import value of the Goods; plus, any mark-up (or rebate); plus, any other related local cost, and custom duties and other import taxes already paid or to be paid on the Goods already imported;
 - ii) the custom duties and other import taxes already paid (need to be supported with documentary evidence) or to be paid on the Goods already imported;
 - iii) any sales and other taxes levied in Kenya which will be payable on the Goods if the Contract is awarded to the Tenderer; and
 - iv) the price for inland transportation, insurance, and other local services required to convey the Goods from the named place of destination to their final destination (Project Site) specified **in the TDS**.
- d) for Related Services, other than inland transportation and other services required to convey the Goods to their final destination, whenever such Related Services are specified in the Schedule of Requirements, the price of each item comprising the Related Services (inclusive of any applicable taxes).

14. Currencies of Tender and Payment

The currency (ies) of the Tender, the currency (ies) of award and the currency (ies) of contract payments shall be the same.

The Tenderer shall quote in Kenya shillings. If allowed in the **TDS**, the Tenderer may express the Tender price in any currency, provided it shall use no more than two foreign currencies in addition to the Kenya Shilling.

The rates of exchange to be used by the Tenderer shall be based on the exchange rates provided by the Central Bank of Kenya on the date 30 days prior to the actual date of tender opening.

15. Documents Establishing the Eligibility and Conformity of the Goods and Related Services

To establish the eligibility of the Goods and Related Services in accordance with ITT 15, Tenderers shall complete the country of origin declarations in the Price Schedule Forms, included in Section IV, Tendering Forms.

To establish the conformity of the Goods and Related Services to the tendering document, the Tenderer shall furnish as part of its Tender the documentary evidence that the Goods conform to the technical specifications and standards specified in Section VII, Schedule of Requirements.

The documentary evidence may be in the form of literature, drawings or data, and shall consist of a detailed item by item description of the essential technical and performance characteristics of the Goods and Related Services, demonstrating substantial responsiveness of the Goods and Related Services to the technical specification, and if applicable, a statement of deviations and exceptions to the provisions of the Section VII, Schedule of Requirements.

The Tenderer shall also furnish a list giving full particulars, including available sources and current prices of spare parts, special tools, etc., necessary for the proper and continuing functioning of the Goods during the period **specified in the TDS** following commencement of the use of the goods by the Procuring Entity.

Standards for workmanship, process, material, and equipment, as well as references to brand names or catalogue numbers specified by the Procuring Entity in the Schedule of Requirements, are intended to be descriptive only and not restrictive. The Tenderer may offer other standards of quality, brand names, and/or catalogue numbers, provided that it demonstrates, to the Procuring Entity's satisfaction, that the substitutions ensure substantial equivalence or are superior to those specified in the Section VII, Schedule of Requirements.

16. Documents Establishing the Eligibility and Qualifications of the Tenderer

To establish Tenderer eligibility in accordance with ITT 4, Tenderers shall complete the Form of Tender, included in Section IV, Tendering Forms.

The documentary evidence of the Tenderer qualifications to perform the Contract if its Tender is accepted shall establish to the Procuring Entity's satisfaction:

- a) that, if required **in the TDS**, a Tenderer that does not manufacture or produce the Goods it offers to

supply shall submit the Manufacturer's Authorization using the form included in Section IV, Tendering Forms to demonstrate that it has been duly authorized by the manufacturer or producer of the Goods to supply these Goods in Kenya;

- b) that, if required **in the TDS**, in case of a Tenderer not doing business within the Kenya, the Tenderer is or will be (if awarded the Contract) represented by an Agent in the country equipped and able to carry out the Supplier's maintenance, repair and spare parts-stocking obligations prescribed in the Conditions of Contract and/or Technical Specifications; and
- c) that the Tenderer meets each of the qualification criterion specified in Section III, Evaluation and Qualification Criteria.

17. Period of Validity of Tenders

Tenders shall remain valid for the Tender Validity period specified **in the TDS**. The Tender Validity period starts from the date fixed for the Tender submission deadline (as prescribed by the Procuring Entity in accordance with ITT 21.1). A Tender valid for a shorter period shall be rejected by the Procuring Entity as non-responsive.

In exceptional circumstances, prior to the expiration of the Tender validity period, the Procuring Entity may request Tenderers to extend the period of validity of their Tenders. The request and the responses shall be made in writing. If a Tender Security is requested in accordance with ITT 18, it shall also be extended for a corresponding period. A Tenderer may refuse the request without forfeiting its Tender Security. A Tenderer granting the request shall not be required or permitted to modify its Tender, except as provided in ITT 17.3.

If the award is delayed by a period exceeding the number of days to be specified in the **TDS** days beyond the expiry of the initial tender validity period, the Contract price shall be determined as follows:

- a) in the case of **fixed price** contracts, the Contract price shall be the tender price adjusted by the factor specified **in the TDS**;
- b) in the case of **adjustable price** contracts, no adjustment shall be made; or in any case, tender evaluation shall be based on the tender price without taking into consideration the applicable correction from those indicated above.

18. Tender Security

The Tenderer shall furnish as part of its Tender, either a Tender-Securing Declaration or a Tender Security, as specified **in the TDS**, in original form and, in the case of a Tender Security, in the amount and currency specified **in the TDS**.

A Tender Securing Declaration shall use the form included in Section IV, Tendering Forms.

If a Tender Security is specified pursuant to ITT 18.1, the Tender Security shall be a demand guarantee in any of the following forms at the Tenderer option:

- i) cash;
- ii) a bank guarantee;
- iii) a guarantee by an insurance company registered and licensed by the Insurance Regulatory Authority listed by the Authority; or
- iv) a letter of credit; or
- v) Guarantee by a deposit taking micro-finance institution, Sacco society, the Youth Enterprise Development Fund or the Women Enterprise Fund.

If an unconditional guarantee is issued by a non-Bank financial institution located outside Kenya, the issuing non-Bank financial institution shall have a correspondent financial institution located in Kenya to make it enforceable unless the Procuring Entity has agreed in writing, prior to tender submission, that a correspondent financial institution is not required. In the case of a bank guarantee, the Tender Security shall be submitted either using the Tender Security Form included in Section IV, Tendering Forms, or in another substantially similar format approved by the Procuring Entity prior to Tender submission. The Tender Security shall be valid for thirty (30) Days beyond the original validity period of the Tender, or beyond any period of extension if requested under ITT 17.2.

If a Tender Security is specified pursuant to ITT 18.1, any Tender not accompanied by a substantially responsive Tender Security shall be rejected by the Procuring Entity as non-responsive.

If a Tender Security is specified pursuant to ITT 18.1, the Tender Security of unsuccessful Tenderers shall be returned as promptly as possible upon the successful tenderer signing the Contract and furnishing the Performance Security pursuant to ITT 46. The Procuring Entity shall also promptly return the tender security

The Tender Security of the successful Tenderer shall be returned as promptly as possible once the successful Tenderer has signed the Contract and furnished the required Performance Security.

The Tender Security may be forfeited or the Tender Securing Declaration executed:

- a) if a Tenderer withdraws its Tender during the period of Tender validity specified by the Tenderer in the Form of Tender, or any extension thereto provided by the Tenderer; or
- b) if the successful Tenderer fails to:
 - i) sign the Contract in accordance with ITT 45; or
 - ii) Furnish a Performance Security in accordance with ITT 46.

Where tender securing declaration is executed, the Procuring Entity shall recommend to the PPRA that PPRA debar the Tenderer from participating in public procurement as provided in the law.

The Tender Security or Tender- Securing Declaration of a JV must be in the name of the JV that submits the Tender. If the JV has not been legally constituted into a legally enforceable JV at the time of Tendering, the Tender Security or Tender-Securing Declaration shall be in the names of all future members as named in the letter of intent referred to in ITT3.1 and ITT 10.2.

A tenderer shall not issue a tender security to guarantee itself.

19. Format and Signing of Tender

The Tenderer shall prepare one original of the documents comprising the Tender as described in ITT 11 and clearly mark it "ORIGINAL." Alternative Tenders, if permitted in accordance with ITT 12, shall be clearly marked "ALTERNATIVE." In addition, the Tenderer shall submit copies of the Tender, in the number **specified in the TDS** and clearly mark them "COPY." In the event of any discrepancy between the original and the copies, the original shall prevail.

Tenderers shall mark as "CONFIDENTIAL" information in their Tenders which is confidential to their business. This may include proprietary information, trade secrets, or commercial or financially sensitive information.

The original and all copies of the Tender shall be typed or written in indelible ink and shall be signed by a person duly authorized to sign on behalf of the Tenderer. This authorization shall consist of a written confirmation **as specified in the TDS** and shall be attached to the Tender. The name and position held by each person signing the authorization must be typed or printed below the signature. All pages of the Tender where entries or amendments have been made shall be signed or initialed by the person signing the Tender.

In case the Tenderer is a JV, the Tender shall be signed by an authorized representative of the JV on behalf of the JV, and so as to be legally binding on all the members as evidenced by a power of attorney signed by each members' legally authorized representatives.

Any inter-lineation, erasures, or overwriting shall be valid only if they are signed or initialed by the person signing the Tender.

D. Submission and Opening of Tenders

20 Sealing and Marking of Tenders

Depending on the sizes or quantities or weight of the tender documents, a tenderer may use an envelope, package or container. The Tenderer shall deliver the Tender in a single sealed envelope, or in a single sealed package, or in a single sealed container bearing the name and Reference number of the Tender, addressed to the Procuring Entity and a warning not to open before the time and date for Tender opening date. Within the single envelope, package or container, the Tenderer shall place the following separate, sealed envelopes:

- a) in an envelope or package or container marked "ORIGINAL", all documents comprising the Tender, as described in ITT 11; and
- b) in an envelope or package or container marked "COPIES", all required copies of the Tender; and
- c) if alternative Tenders are permitted in accordance with ITT 12, and if relevant:
 - i) in an envelope or package or container marked "ORIGINAL –ALTERNATIVE TENDER", the alternative Tender; and
 - ii) in the envelope or package or container marked "COPIES- ALTERNATIVE TENDER", all required copies of the alternative Tender.

The inner envelopes or packages or containers shall:

- a) Bear the name and address of the Procuring Entity.
- b) bear the name and address of the Tenderer; and
- c) Bear the name and Reference number of the Tender.

Where a tender package or container cannot fit in the tender box, the procuring entity shall:

- a) Specify in the **TDS where** such documents should be received.
- b) Maintain a record of tenders received and issue acknowledgement receipt note to each tenderer specifying time and date of receipt.
- c) Ensure all tenders received are handed over to the tender opening committee for opening at the specified opening place and time.

If an envelope or package or container is not sealed and marked as required, the *Procuring Entity* will assume no responsibility for the misplacement or premature opening of the Tender. Tenders misplaced or opened prematurely will not be accepted.

21. Deadline for Submission of Tenders

Tenders must be received by the Procuring Entity at the address and no later than the date and time specified **in the TDS**. When so specified **in the TDS**, Tenderers shall have the option of submitting their Tenders electronically. Tenderers submitting Tenders electronically shall follow the electronic Tender submission procedures **specified in the TDS**.

The Procuring Entity may, at its discretion, extend the deadline for the submission of Tenders by amending the tendering document in accordance with ITT7, in which case all rights and obligations of the Procuring Entity and Tenderers previously subject to the deadline shall thereafter be subject to the deadline as extended.

22. Late Tenders

The Procuring Entity shall not consider any Tender that arrives after the deadline for submission of Tenders. Any Tender received by the Procuring Entity after the deadline for submission of Tenders shall be declared late, rejected, and returned unopened to the Tenderer.

23. Withdrawal, Substitution, and Modification of Tenders

A Tenderer may withdraw, substitute, or modify its Tender after it has been submitted by sending a written notice, duly signed by an authorized representative, and shall include a copy of the authorization (the power of attorney) in accordance with ITT19.3, (except that withdrawal notices do not require copies). The corresponding substitution or modification of the Tender must accompany the respective written notice. All notices must be:

- a) prepared and submitted in accordance with ITT 20 and 21 (except that withdrawal notices do not require copies), and in addition, the respective envelopes shall be clearly marked “WITHDRAWAL,” “SUBSTITUTION,” or “MODIFICATION;” and
- b) Received by the Procuring Entity prior to the deadline prescribed for submission of Tenders, in accordance with ITT 22.

Tenders requested to be withdrawn in accordance with ITT 23.1 shall be returned unopened to the Tenderers.

No Tender may be withdrawn, substituted, or modified in the interval between the deadline for submission of Tenders and the expiration of the period of Tender validity specified by the Tenderer on the Form of Tender or any extension thereof.

24. Tender Opening

Except as in the cases specified in ITT 23, the Procuring Entity shall, at the Tender opening, publicly open and read out all Tenders received by the deadline at the date, time and place specified **in the TDS** in the presence of Tenderers’ designated representatives who choose to attend, including to attend any specific electronic tender opening procedures if electronic tendering is permitted in accordance with ITT 21.1, shall be as specified **in the TDS**.

First, envelopes marked “WITHDRAWAL” shall be opened and read out and the envelope with the corresponding Tender shall not be opened, but returned to the Tenderer. If the withdrawal envelope does not contain a copy of the “power of attorney” confirming the signature as a person duly authorized to sign on behalf of the Tenderer, the corresponding Tender will be opened. No Tender withdrawal shall be permitted unless the corresponding withdrawal notice contains a valid authorization to request the withdrawal and is read out at Tender opening.

Next, envelopes marked “SUBSTITUTION” shall be opened and read out and exchanged with the corresponding Tender being substituted, and the substituted Tender shall not be opened, but returned to the Tenderer. No Tender substitution shall be permitted unless the corresponding substitution notice contains a valid authorization to request the substitution and is read out at Tender opening.

Next, envelopes marked “MODIFICATION” shall be opened and read out with the corresponding Tender. No

Tender modification shall be permitted unless the corresponding modification notice contains a valid authorization to request the modification and is read out at Tender opening.

Next, all remaining envelopes shall be opened one at a time, reading out: the name of the Tenderer and whether there is a modification; the total Tender Prices, per lot (contract) if applicable, including any discounts and alternative Tenders; the presence or absence of a Tender Security, if required; and any other details as the Procuring Entity may consider appropriate.

Only Tenders, alternative Tenders and discounts that are opened and read out at Tender opening shall be considered further for evaluation. The Form of Tender and pages of the Bills of Quantities are to be initialed by the members of the tender opening committee attending the opening. The number of representatives of the Procuring Entity to sign shall be specified in the **TDS**.

The Procuring Entity shall neither discuss the merits of any Tender nor reject any Tender (except for late Tenders, in accordance with ITT 22.1).

The Procuring Entity shall prepare a record of the Tender opening that shall include, as a minimum:

- a) the name of the Tenderer and whether there is a withdrawal, substitution, or modification;
- b) the Tender Price, per lot (contract) if applicable, including any discounts;
- c) any alternative Tenders;
- d) the presence or absence of a Tender Security or Tender-Securing Declaration, if one was required;
- e) Number of pages of each tender document submitted.

The Tenderers' representatives who are present shall be requested to sign the record. The omission of a Tenderer signature on the record shall not invalidate the contents and effect of the record. A copy of the tender opening register shall be issued to a Tenderer upon request.

E. Evaluation and Comparison of Tenders

25. Confidentiality

Information relating to the evaluation of Tenders and recommendation of contract award, shall not be disclosed to Tenderers or any other persons not officially concerned with the tendering process until the information on Intention to Award the Contract is transmitted to all Tenderers in accordance with ITT 41.

Any effort by a Tenderer to influence the Procuring Entity in the evaluation or contract award decisions may result in the rejection of its Tender.

Notwithstanding ITT 25.2, from the time of Tender opening to the time of Contract Award, if any Tenderer wishes to contact the Procuring Entity on any matter related to the Tendering process, it should do so in writing.

26. Clarification of Tenders

To assist in the examination, evaluation, comparison of the Tenders, and qualification of the Tenderers, the Procuring Entity may, at its discretion, ask any Tenderer for a clarification of its Tender. Any clarification submitted by a Tenderer in respect to its Tender and that is not in response to a request by the Procuring Entity shall not be considered. The Procuring Entity's request for clarification and the response shall be in writing. No change, including any voluntary increase or decrease, in the prices or substance of the Tender shall be sought, offered, or permitted except to confirm the correction of arithmetic errors discovered by the Procuring Entity in the Evaluation of the Tenders, in accordance with ITT 30.

If a Tenderer does not provide clarifications of its Tender by the date and time set in the Procuring Entity's request for clarification, its Tender may be rejected.

27. Deviations, Reservations, and Omissions

During the evaluation of Tenders, the following definitions apply:

- a) "Deviation" is a departure from the requirements specified in the Tendering document;
- b) "Reservation" is the setting of limiting conditions or withholding from complete acceptance of the requirements specified in the tendering document; and
- c) "Omission" is the failure to submit part or all of the information or documentation required in the tendering document.

28. Determination of Responsiveness

The Procuring Entity's determination of a Tender's responsiveness is to be based on the contents of the Tender itself, as defined in ITT28.2.

- 28 A substantially responsive Tender is one that meets the requirements of the tendering document without material deviation, reservation, or omission. A material deviation, reservation, or omission is one that:
- a) if accepted, would:
 - i) affect in any substantial way the scope, quality, or performance of the Goods and Related Services specified in the Contract; or
 - ii) limit in any substantial way, inconsistent with the tendering document, the Procuring Entity's rights or the Tenderer obligations under the Contract; or
 - b) if rectified, would unfairly affect the competitive position of other Tenderers presenting substantially responsive Tenders.

The Procuring Entity shall examine the technical aspects of the Tender submitted in accordance with ITT 15 and ITT 16, in particular, to confirm that all requirements of Section VII, Schedule of Requirements have been met without any material deviation or reservation, or omission.

If a Tender is not substantially responsive to the requirements of tendering document, it shall be rejected by the Procuring Entity and may not subsequently be made responsive by correction of the material deviation, reservation, or omission.

29. Non-conformities, Errors and Omissions

Provided that a Tender is substantially responsive, the Procuring Entity may waive any non-conformities in the Tender.

Provided that a Tender is substantially responsive, the Procuring Entity may request that the Tenderer submit the necessary information or documentation, within a reasonable period of time, to rectify nonmaterial non-conformities or omissions in the Tender related to documentation requirements. Such omission shall not be related to any aspect of the price of the Tender. Failure of the Tenderer to comply with the request may result in the rejection of its Tender.

Provided that a Tender is substantially responsive, the Procuring Entity shall rectify quantifiable nonmaterial non-conformities related to the Tender Price. To this effect, the Tender Price shall be adjusted, for comparison purposes only, to reflect the price of a missing or non-conforming item or component in the manner specified **in the TDS**. The adjustment shall be based on the *average* price of the item or component as quoted in other substantially responsive Tenders. If the price of the item or component cannot be derived from the price of other substantially responsive Tenders, the Procuring Entity shall use its best estimate.

30. Arithmetical Errors

The tender sum as submitted and read out during the tender opening shall be absolute and final and shall not be the subject of correction, adjustment or amendment in any way by any person or entity.

Provided that the Tender is substantially responsive, the Procuring Entity shall handle errors on the following basis:

- a) Any error detected if considered a major deviation that affects the substance of the tender, shall lead to disqualification of the tender as non-responsive.
- b) Any errors in the submitted tender arising from a miscalculation of unit price, quantity, and subtotal and total bid price shall be considered as a major deviation that affects the substance of the tender and shall lead to disqualification of the tender as non-responsive. And
- c) if there is a discrepancy between words and figures, the amount in words shall prevail.

Tenderers shall be notified of any error detected in their bid during the notification of a ward.

31. Conversion to Single Currency

- 31.1 For evaluation and comparison purposes, the currency (ies) of the Tender shall be converted in a single currency as specified **in the TDS**.

32. Margin of Preference and Reservations

A margin of preference may be allowed on locally manufactured goods only when the contract is open to

International tendering, where the tender is likely to attract foreign goods and where the contract exceeds the threshold specified in the Regulations.

For purposes of granting a margin of preference on locally manufactured goods under international competitive tendering, a procuring entity shall not subject the items listed below to international tender and hence no margin of preference shall be allowed. The affected items are:

- a) motor vehicles, plant and equipment which are assembled in Kenya;
- b) furniture, textile, foodstuffs, oil and gas, information communication technology, steel, cement, leather agro-processing, sanitary products, and other goods made in Kenya; or
- c) Goods manufactured, mined, extracted or grown in Kenya.

A margin of preference shall not be allowed unless it is specified so in the **TDS**.

Contracts procured on basis of international competitive tendering shall not be subject to reservations to specific group's as provided in ITT 32.5.

Where it is intended to reserve a contract to a specific group of businesses (these groups are Small and Medium Enterprises, Women Enterprises, Youth Enterprises and Enterprises of persons living with disability, as the case may be), and who are appropriately registered as such by the authority to be specified in the **TDS**, a procuring entity shall ensure that the invitation to tender specifically indicates that only businesses or firms belonging to the specified group are eligible to tender as specified in the **TDS**. No tender shall be reserved to more than one group. If not so stated in the Tender documents, the invitation to tender will be open to all interested tenderers.

3.Evaluation of Tenders

The Procuring Entity shall use the criteria and methodologies listed in this ITT and Section III, Evaluation and Qualification criteria. No other evaluation criteria or methodologies shall be permitted. By applying the criteria and methodologies, the Procuring Entity shall determine the Lowest Evaluated Tender. This is the Tender of the Tenderer that meets the qualification criteria and who's Tender has been determined to be:

- a) substantially responsive to the tender documents; and
- b) The lowest evaluated price.

Price evaluation will be done for Items or Lots (contracts), as specified **in the TDS**; and the Tender Price as quoted in accordance with ITT 14. To evaluate a Tender, the Procuring Entity shall consider the following:

- a) price adjustment due to unconditional discounts offered in accordance with ITT 13.4;
- b) converting the amount resulting from applying (a) and (b) above, if relevant, to a single currency in accordance with ITT 31;
- c) price adjustment due to quantifiable nonmaterial non-conformities in accordance with ITT 29.3; and
- d) Any additional evaluation factors specified **in the TDS** and Section III, Evaluation and Qualification Criteria.

The estimated effect of the price adjustment provisions of the Conditions of Contract, applied over the period of execution of the Contract, shall not be considered in Tender evaluation.

Where the tender involves multiple lots or contracts, the tenderer will be allowed to tender for one or more lots (contracts). Each lot or contract will be evaluated in accordance with ITT 33.2. The methodology to determine the lowest evaluated tenderer or tenderers based one lot (contract) or based on a combination of lots (contracts), will be specified in Section III, Evaluation and Qualification Criteria. In the case of multiple lots or contracts, tenderer will be will be required to prepare the Eligibility and Qualification Criteria Form for each Lot.

The Procuring Entity's evaluation of a Tender will include and consider:

- a) in the case of Goods manufactured in Kenya, sales and other similar taxes, which will be payable on the goods if a contract is awarded to the Tenderer;
- b) in the case of Goods manufactured outside Kenya, already imported or to be imported, customs duties and other import taxes levied on the imported Good, sales and other similar taxes, which will be payable on the Goods if the contract is awarded to the Tenderer;

The Procuring Entity's evaluation of a Tender may require the consideration of other factors, in addition to the Tender Price quoted in accordance with ITT 14. These factors may be related to the characteristics, performance, and terms and conditions of purchase of the Goods and Related Services. The effect of the factors selected, if any, shall be expressed in monetary terms to facilitate comparison of Tenders, unless otherwise specified in the **TDS** from amongst those set out in Section III, Evaluation and Qualification

Criteria. The additional criteria and methodologies to be used shall be as specified in ITT 33.2(d).

34 Comparison of Tenders

The Procuring Entity shall compare the evaluated costs of all substantially responsive Tenders established in accordance with ITT 33.2 to determine the Tender that has the lowest evaluated cost. The comparison shall be on the basis of total cost (place of final destination) prices for all goods and all prices, plus cost of inland transportation and insurance to place of destination, for goods manufactured within the Kenya, together with prices for any required installation, training, commissioning and other services.

35 Abnormally Low Tenders

An Abnormally Low Tender is one where the Tender price, in combination with other constituent elements of the Tender, appears unreasonably low to the extent that the Tender price raises material concerns with the Procuring Entity as to the capability of the Tenderer to perform the Contract for the offered Tender price.

In the event of identification of a potentially Abnormally Low Tender by the evaluation committee, the Procuring Entity shall seek written clarification from the Tenderer, including a detailed price analyses of its Tender price in relation to the subject matter of the contract, scope, delivery schedule, allocation of risks and responsibilities and any other requirements of the tendering document.

After evaluation of the price analysis, in the event that the Procuring Entity determines that the Tenderer has failed to demonstrate its capability to perform the contract for the offered Tender price, the Procuring Entity shall reject the Tender.

36 Abnormally High Tenders

An abnormally high price is one where the tender price, in combination with other constituent elements of the Tender, appears unreasonably too high to the extent that the Procuring Entity is concerned that it (the Procuring Entity) may not be getting value for money or it may be paying too high a price for the contract compared with market prices or that genuine competition between Tenderers is compromised.

In case of an abnormally high tender price, the Procuring Entity shall make a survey of the market prices, check if the estimated cost of the contract is correct and review the Tender Documents to check if the specifications, scope of work and conditions of contract are contributory to the abnormally high tenders. The Procuring Entity may also seek written clarification from the tenderer on the reason for the high tender price. The Procuring Entity shall proceed as follows:

- i) If the tender price is abnormally high based on wrong estimated cost of the contract, the Procuring Entity may accept or not accept the tender depending on the Procuring Entity's budget considerations.
- ii) If specifications, scope of work and/or conditions of contract are contributory to the abnormally high tender prices, the Procuring Entity shall reject all tenders and may retender for the contract based on revised estimates, specifications, scope of work and conditions of contract, as the case may be.

If the Procuring Entity determines that the Tender Price is abnormally too high because genuine competition between tenderers is compromised (*often due to collusion, corruption or other manipulations*), the Procuring Entity shall reject all Tenders and shall institute or cause relevant Government Agencies to institute an investigation on the cause of the compromise, before retendering.

37 Post-Qualification of the Tenderer

The Procuring Entity shall determine, to its satisfaction, whether the eligible Tenderer that is selected as having submitted the lowest evaluated cost and substantially responsive Tender, meets the qualifying criteria specified in Section III, Evaluation and Qualification Criteria.

The determination shall be based upon an examination of the documentary evidence of the Tenderer qualifications submitted by the Tenderer, pursuant to ITT 15 and 16. The determination shall not take into consideration the qualifications of other firms such as the Tenderer subsidiaries, parent entities, affiliates, subcontractors (other than specialized subcontractors if permitted in the tendering document), or any other firm(s) different from the Tenderer.

An affirmative determination shall be a prerequisite for award of the Contract to the Tenderer. A negative determination shall result in disqualification of the Tender, in which event the Procuring Entity shall proceed to the Tenderer who offers a substantially responsive Tender with the next lowest evaluated cost to make a similar determination of that Tenderer qualifications to perform satisfactorily.

38 Lowest Evaluated Tender

Having compared the evaluated prices of Tenders, the Procuring Entity shall determine the Lowest Evaluated Tender. The Lowest Evaluated Tender is the Tender of the Tenderer that meets the Qualification Criteria and who's Tender has been determined to be:

- a) most responsive to the Tender document; and
- b) The lowest evaluated price.

39. Procuring Entity's Right to Accept Any Tender, and to reject any or All Tenders.

The Procuring Entity reserves the right to accept or reject any tender, and to annul the Tendering process and reject all Tenders at any time prior to notification Award, without thereby incurring any liability to Tenderers. In case of annulment, all Tenderers shall be notified with reasons and all Tenders submitted and specifically, tender securities, shall be promptly returned to the Tenderers.

F. Award of Contract

40. Award Criteria

The Procuring Entity shall award the Contract to the successful tenderer whose tender has been determined to be the Lowest Evaluated Tender in accordance with procedures in Section 3: Evaluation and Qualification Criteria.

41. Procuring Entity's Right to Vary Quantities at Time of Award

The Procuring Entity reserves the right at the time of Contract award to increase or decrease, by the percentage (s) for items as indicated in the TDS.

42. Notice of Intention to enter into a Contract

Upon award of the contract and Prior to the expiry of the Tender Validity Period the Procuring Entity shall issue a Notification of Intention to Enter into a Contract / Notification of award to all tenderers which shall contain, at a minimum, the following information:

- a) the name and address of the Tenderer submitting the successful tender;
- b) the Contract price of the successful tender;
- c) a statement of the reason(s) the tender of the unsuccessful tenderer to whom the letter is addressed was unsuccessful, unless the price information in I above already reveals the reason;
- d) the expiry date of the Standstill Period; and
- e) instructions on how to request a debriefing and/or submit a complaint during the standstill period;

43. Standstill Period

The Contract shall not be awarded earlier than the expiry of a Standstill Period of 14 days to allow any dissatisfied candidate to launch a complaint. Where only one Tender is submitted, the Standstill Period shall not apply.

Where standstill period applies, it shall commence when the Procuring Entity has transmitted to each Tenderer the Notification of Intention to Enter into a Contract to the successful Tenderer.

4. Debriefing by the Procuring Entity

On receipt of the Procuring Entity's Notification of Intention to Enter into a Contract referred to in ITT 41, an unsuccessful tenderer may make a written request to the Procuring Entity for a debriefing on specific issues or concerns regarding their tender. The Procuring Entity shall provide the debriefing within five days of receipt of the request.

Debriefings of unsuccessful Tenderers may be done in writing or verbally. The Tenderer shall bear its own costs of attending such a debriefing meeting.

45. Letter of Award

Prior to the expiry of the Tender Validity Period and upon expiry of the Standstill Period specified in ITT 42, upon addressing a complaint that has been filed within the Standstill Period, the Procuring Entity shall transmit the Letter of Award to the successful Tenderer. The letter of award shall request the successful tenderer to furnish the Performance Security within 21 days of the date of the letter.

46. Signing of Contract

Upon the expiry of the fourteen days of the Notification of Intention to enter into contract and upon the parties meeting their respective statutory requirements, the Procuring Entity shall send the successful Tenderer the Contract Agreement.

Within fourteen (14) days of receipt of the Contract Agreement, the successful Tenderer shall sign, date, and return it to the Procuring Entity.

The written contract shall be entered into within the period specified in the notification of award and before expiry of the tender validity period.

47. Performance Security

Within twenty-one (21) days of the receipt of Letter of Acceptance from the Procuring Entity, the successful Tenderer, if required, shall furnish the Performance Security in accordance with the GCC 18, using for that purpose the Performance Security Form included in Section X, Contract Forms. If the Performance Security furnished by the successful Tenderer is in the form of a bond, it shall be issued by a bonding or insurance company that has been determined by the successful Tenderer to be acceptable to the Procuring Entity. A foreign institution providing a bond shall have a correspondent financial institution located in Kenya, unless the Procuring Entity has agreed in writing that a correspondent financial institution is not required.

Failure of the successful Tenderer to submit the above-mentioned Performance Security or sign the Contract shall constitute sufficient grounds for the annulment of the award and forfeiture of the Tender Security. In that event the Procuring Entity may award the Contract to the Tenderer offering the next lowest Evaluated Tender.

Performance security shall not be required for a contract, if so specified in the **TDS**.

48. Publication of Procurement Contract

Within fourteen days after signing the contract, the Procuring Entity shall publish and publicize the awarded contract at its notice boards, entity website; and on the Website of the Authority in manner and format prescribed by the Authority. At the minimum, the notice shall contain the following information:

- a) name and address of the Procuring Entity;
- b) name and reference number of the contract being awarded, a summary of its scope and the selection method used;
- c) the name of the successful Tenderer, the final total contract price, the contract duration.
- d) dates of signature, commencement and completion of contract;
- e) names of all Tenderers that submitted Tenders, and their Tender prices as read out at Tender opening;

49. Procurement Related Complaints and Administrative Review

The procedures for making a Procurement-related Complaint are as specified in the **TDS**.

A request for administrative review shall be made in the form provided under contract forms.

SECTION II – TENDER DATA SHEET (TDS)

The following specific data shall complement, supplement, or amend the provisions in the Instructions to Tenderers (ITT). Whenever there is a conflict, the provisions herein shall prevail over those in ITT.

ITT Reference	Particulars Of Appendix To Instructions To Tenders
A. General	
ITT 1.1	The reference number of the Invitation for Tenders is: KNH/T43/2026-2027 The Procuring Entity is: KENYATTA NATIONAL HOSPITAL The name of the Contract is: SUPPLY AND DELIVERY OF CALIBRATION TOOLS (<i>Reserved for Youth, Women and People Living with Disability</i>) The number and identification of lots (contracts) comprising this Invitation for Tenders is: KNH/T/43/2026- 2027
ITT 1.2(a)	<i>[delete if not applicable]</i> Electronic –Procurement System The Procuring Entity shall use the following electronic-procurement system to manage this Tendering process: Not applicable <i>[insert name of the e-system and full address or link]</i> The electronic-procurement system shall be used to manage the following aspects of the Tendering process: <i>[list aspects here and modify the relevant parts of the TDS accordingly e.g., issuing Tendering document, submissions of Tenders, opening of Tenders]- not applicable</i>
ITT 2.3	The Information made available on competing firms is as follows:
	The firms that provided consulting services for the contract being tendered for are: Not applicable
ITT 3.1	Maximum number of members in the Joint Venture (JV) shall be: <i>[Not applicable]</i>
ITT 3.7	A list of debarred firms and individuals is available on the PPRA’s website: www.ppra.go.ke
ITT 3.11	Tenderers shall be required to be to be registered with – Not applicable
B. Contents of Tendering Document	
ITT 6.1	(a) Address where to send enquiries is P.O box 20723-00202 Nairobi and procurementknh@gmail.com or procurement@knh.or.ke to reach the Procuring Entity not later than 22 /06/2026 10:00hrs (Kenyan time) (b) The Procuring Entity publish its response at the website www.knh.or.ke
ITT 6.2	A pre-tender conference will be held on- NOT APPLICABLE
ITT 6.3	The questions to reach the Procuring Entity not later than 22/06/2026 10:00hrs
ITT 6.5	The Minutes of the Pre-Tender meeting shall be published on the at the website: www.knh.or.ke
C. Preparation of Tenders	

ITT Reference	Particulars Of Appendix To Instructions To Tenders
ITT 10 (j)	The Tenderer shall submit the following additional documents in its Tender: <i>[list any additional documents not already listed in ITT 11.1 that must be submitted with the Tender]- not applicable</i>
ITT 12.1	Alternative Tenders “ <i>shall not be</i> ” considered. <i>[If alternatives shall be considered, the methodology shall be defined in Section III – Evaluation and Qualification Criteria. See Section III for further details]</i>
ITT 13.5	The prices quoted by the Tenderer “ shall not ” be subject to adjustment during the performance of the Contract.
ITT 13.6	Prices quoted for each lot (contract) shall correspond at least to [100%] percent of the items specified for each lot (contract). Prices quoted for each item of a lot shall correspond at least to [100%] percent of the quantities specified for this item of a lot.
ITT 13.8 (a) (i) and (iii)	Place of final destination: is Kenyatta National Hospital
ITT 13.8 (a) (iii)	Final Destination (Project Site): <i>[insert final destination/project site, if different from named place of destination]- Not applicable</i>
ITT 13.8 (b) (i)	Named place of destination, in Kenya is Kenyatta National Hospital
ITT 13.8 (b) (ii)	The price for inland transportation, insurance, and other local services required to convey the Goods from the named place of destination to their final destination which is inclusive of price quoted.
13.8 I (iv)	The place of final destination (Project Site) is Kenyatta National Hospital.
ITT 14.2	Foreign currency requirements Not allowed.
ITT 15.4	Period of time the Goods are expected to be functioning (for the purpose of spare parts): <i>[2/3 of its lifespan] N/A</i>
ITT 16.2 (a)	Manufacturer’s authorization is: not “ required ”
ITT 16.2 (b)	After sales service is: “ not required ”
ITT 17.1	The Tender validity period shall be 119 days.
ITT 17.3	(a) The Number of days beyond the expiry of the initial tender validity period will be 30days. (b) The Tender price shall be adjusted by the following percentages of the tender price: (i) _____% of the local currency portion of the Contract price adjusted to reflect local inflation during the period of extension, and By _____ (ii) _____% the foreign currency portion of the Contract price adjusted to reflect the international inflation during the period of extension.- not applicable
ITT 18.1	<i>[If a Tender Security shall be required, a Tender-Securing Declaration shall not be required, and vice versa.]</i>

ITT Reference	Particulars Of Appendix To Instructions To Tenders
	<p>A Tender-Securing Declaration SHALL BE required. Bidders are required to completely fill, sign and stamp the Tender Securing Declaration Form as provided in the tender document.</p>
ITT 19.1	<p>In addition to the original of the Tender, the number of copies is: <i>[ONE (1) of copies of the Original bid document]</i></p>
ITT 19.3	<p>The written confirmation of authorization to sign on behalf of the Tenderer shall consist of: <i>[insert the name and description of the documentation required to demonstrate the authority of the signatory to sign the Tender].</i></p>
	<p>D. Submission and Opening of Tenders</p>
ITT 20.3	<p>A tender package or container that cannot fit in the tender box shall be received as follows: Physical delivery at below address for registration</p> <p>Director supply Chain management KNH</p> <p>Po Box 20723-00202 Nairobi</p> <p>Administration Block; Supply chain Management Division Room 6_____</p>
ITT 21.1	<p>For Tender submission purposes only, the Procuring Entity’s address is: <i>[This address may be the same as or different from that specified under provision ITT 7.1 for clarifications]</i></p> <p>Attention: <i>To Chief executive officer]</i></p> <p>Postal Address: <i>[20723-00202 Nairobi Kenya]</i></p> <p>Physical Address: Nairobi City County Upper hill off Hospital, Kenyatta National Hospital Administration block, supply Chain Management Office Entrance.</p> <p>Telephone: <i>[2726300-9]</i></p> <p>Electronic mail address: <i>[procurement@knh.or.ke, procurementknh@gmail.com]</i></p> <p>The deadline for Tender submission is: Date: 29th June,2026</p> <p>tenderers “shall not” have the option of submitting their Tenders electronically. <i>[Note: The following provision should be included and the required corresponding information inserted <u>only</u> if tenderers have the option of submitting their Tenders electronically. Otherwise omit.]</i></p> <p>The electronic Tendering submission procedures shall be: <i>N/A</i></p>
ITT 24.1	<p>The Tender opening shall take place at:</p> <p>Attention: <i>The Chief Executive Officer</i></p> <p>Postal Address: <i>[20723-00200 Nairobi]</i></p> <p>Physical Address: Nairobi City County Upper hill off Hospital, Kenyatta National Hospital Administration block, supply Chain Management office Entrance</p> <p>Date: 29th June 2026 at 10:00 am</p>

ITT Reference	Particulars Of Appendix To Instructions To Tenders
	The electronic Tender opening procedures shall be: <i>N/A</i>
ITT 24.6	The number of representatives of the Procuring Entity to sign is three.
E. Evaluation and Comparison of Tenders	
ITT 29.3	The manner of rectify quantifiable non material non conformities described below: -----
ITT 31.1	The currency that shall be used for Tender evaluation and comparison purposes to convert at the selling exchange rate all Tender prices expressed in various currencies into a single currency is: <i>[Kenya Shilling]</i> The source of exchange rate shall be: <i>(the Central Bank in Kenya)</i> The date for the exchange rate shall be : <i>N/ A</i>
ITT 32.3	A margin of preference and/or reservation shall apply and specify the details. If a margin of preference applies, the application methodology shall be defined in Section III – Evaluation and Qualification Criteria.
ITT 32.5	The invitation to tender is extended to the following group that qualify for Reservations __ who shall be duly registered with _____ Not applicable
ITT 33.2	Price evaluation will be done for the Items
ITT 33.2 (d)	Additional evaluation factors are – <i>N/A</i>
ITT 33.6	The adjustments shall be determined using the following criteria, from amongst those set out in Section III, Evaluation and Qualification Criteria: <i>[refer to Section III, Evaluation and Qualification Criteria; insert complementary details if necessary]</i> <ul style="list-style-type: none"> a) Deviation in Delivery schedule: <i>[No. If yes insert the adjustment factor in Section III, Evaluation and Qualification Criteria]</i> b) Deviation in payment schedule: <i>[No. If yes insert the adjustment factor in Section III, Evaluation and Qualification Criteria]</i> c) the cost of major replacement component, mandatory spare parts, and service: <i>[No. If yes, insert the Methodology and criteria in Section III, Evaluation and Qualification Criteria]</i> d) the availability in Kenya of spare parts and after-sales services for the equipment offered in the Tender <i>[No. If yes, insert the Methodology and criteria in Section III, Evaluation and Qualification Criteria]</i> e) Lifecycle costs: the costs during the life of the goods or equipment <i>[No. If yes, insert the Methodology and criteria in Section III, Evaluation and Qualification Criteria]</i> f) The performance and productivity of the equipment offered; <i>[No. If yes, insert the Methodology and criteria]</i>

ITT Reference	Particulars Of Appendix To Instructions To Tenders
	i. <i>[insert any other specific criteria in Section III, Evaluation and Qualification Criteria]</i>
	F. Award of Contract
ITT 41.1	The maximum percentage by which quantities may be increased is: <i>[15%]</i> The maximum percentage by which quantities may be decreased is: <i>[100%]</i>
ITT 41.1	The Procuring Entity shall increase or decrease the quantity of Goods and Related Services by an amount not exceed 15% and without any change in the unit prices or other terms and conditions of the Tender and the tendering document.
ITT 47.3	Performance security if so required shall be in the sum of 1% where applicable
ITT 49.1	<p>The procedures for making a Procurement-related Complaint are detailed in the “Notice of Intention to Award the Contract” herein and are also available from the PPRA Website www.ppra.go.ke.</p> <p>If a Tenderer wishes to make a Procurement-related Complaint, the Tenderer should submit its complaint following these procedures, in writing (by the quickest means available, that is either by email or fax), to:</p> <p>For the attention: [Dr. Richard Lesiyampe] Title/position: [Ag Chief Executive Officer] Procuring Entity: [Kenyatta national Hospital] Email address: [knhadmin@knh.or.ke]</p> <p>In summary, a Procurement-related Complaint may challenge any of the following:</p> <ol style="list-style-type: none"> 1. the terms of the Tendering Documents; and 2. the Procuring Entity’s decision to award the contract.

SECTION III – EVALUATION AND QUALIFICATION CRITERIA

1. General Provisions

Wherever a Tenderer is required to state a monetary amount, Tenderers should indicate the Kenya Shilling equivalent using the rate of exchange determined as follows:

- a) For business turnover or financial data required for each year - Exchange rate prevailing on the last day of the respective calendar year (in which the amounts for that year is to be converted) was originally established.
- b) Value of single contract – Exchange rate prevailing on the date of the contract signature.
- c) Exchange rates shall be taken from the publicly available source identified in **the ITT 14.3**. Any error in determining the exchange rates in the Tender may be corrected by the Procuring Entity.

This section contains the criteria that the Procuring Entity shall use to evaluate tender and qualify tenderers. No other factors, methods or criteria shall be used other than those specified in this tender document. The Tenderer shall provide all the information requested in the forms included in Section IV, Tendering Forms. The Procuring Entity should use the Standard Tender Evaluation Report for Goods and Works for evaluating Tenders.

2. Evaluation of Tenders (ITT 33)

Successful Tender or Tenders

The Procuring Entity shall use the criteria and methodologies listed in this Section to evaluate Tenders. By applying these criteria and methodologies, the Procuring Entity shall determine the successful Tender or Tenders which has/have been determined to:

- a) be substantially responsive to the tender documents;
- b) offer the lowest evaluated cost to the Procuring Entity for all items of Goods to be procured based on either a single Contract or all multiple Contracts combined, as the case may be, in accordance with the ITT 13.6 inviting Tender prices and discounts, and provisions made of the Tender Document for evaluation of tenders and award of contract (s); and
- c) be offered by Tenderer or Tenderers that substantially meet the qualification criteria applicable for Contract or combined Contracts for which they are selected.

Evaluation of Tenders

Preliminary examination for Determination of Responsiveness

The Procuring Entity will start by examining all tenders to ensure they meet in all respects the eligibility criteria and other mandatory requirements in the ITT, and that the tender is complete in all aspects in meeting the requirements provided for in the preliminary evaluation criteria outlined below. The Standard Tender Evaluation Report Document for Goods and Works for evaluating Tenders provides very clear guide on how to deal with review of these requirements. Tenders that do not pass the Preliminary Examination will be considered non-responsive and will not be considered further.

[The Procuring Entity will provide the preliminary evaluation criteria. To facilitate, a template may be attached or clearly described all information and list of documentation to be submitted by Tenderers to enable preliminary evaluation of the Tender]

MANDATORY REQUIREMENTS

S/No.	Completeness and Responsiveness Criteria	Requirement
MR 1.	Form of Tender	Must submit dully filled form of tender on company letterhead, signed and stamped in the prescribed format in the tender document.
MR 2.	Confidential Business Questionnaire	Duly Filled, Stamped and Signed
MR 3.	Certificate of Independent Tender Determination	Duly Filled, Stamped and Signed
MR 4.	Self-Declaration on debarment (PPADA 2015)	Duly Filled, Stamped and Signed
MR 5.	Self-Declaration on Corruption / Fraudulent Practices	Duly Filled, Stamped and Signed
MR 6.	Declaration and Commitment to the Code of Ethics	Duly Filled, Stamped and Signed
MR 7.	Tenderer Information Form	Dully filled and stamped (organizational chart not required for this tender - bidders to attach list of board of Directors (CR12 or CR13 or copy of National ID of sole proprietor)
MR 8.	Serialization	Must be chronologically and sequentially serialized i.e. 1, 2, 3,4.....on every page including the original tender document.
MR 9.	Tax Compliance Certificate	Provide valid tax compliance certificate
MR 10.	Certificate of Incorporation/ Registration	Must Submit a copy of the Certificate of incorporation or Registration Certificate and corresponding CR12, CR13 where applicable
MR 11.	Original/Copy of Bid Document	Must submit two Tender Documents (Original and Copy) spiral/book bound no stapled documents will be accepted
MR 12.	Trade License	Attach Valid Copy of Trade License or Evidence of renewal from relevant County Government
MR 13.	Tender Securing Declaration Form	Attach Tender Securing Declaration Form duly completed, signed and stamped
MR 14.	Access to Government Procurement Opportunities Certificate (AGPO)	Attach Valid Certificate for access to government procurement opportunities
MR 15.	Written Declaration by all Companies/Institutions that neither of their Directors have participated in the same Tender as Individual Tenderers, Joint Venture, Sole Proprietor or as a subcontractor	Attach copy of declaration signed and stamped by the person authorized to sign the Tender
MR 16.	Power of Attorney	Attach Power of Attorney for company with more than one director
MR 18.	Bank Details	Attach Duly filled , signed and stamped bank details
MR 19.	Tenderer Data Consent Form	Duly Filled, Stamped and Signed as per attached form

MR 20.	Inspection and Certificate of Goods	Must submit a duly filled signed and stamped commitment to comply with Inspection and Acceptance requirements
MR 21.	Supplier Performance	Must submit duly filled signed and stamped commitment to comply with the supplier evaluation requirements

Failure to provide ANY of the above mandatory requirements will lead to disqualification

Mandatory Requirement Evaluation

Evaluation Method: PASS/FAIL

Note: Tenderers must meet all the mandatory requirements to qualify for Technical Evaluation. Failure to satisfy any mandatory requirement shall result in disqualification and the tender shall not proceed to Technical or Financial Evaluation.

STAGE 2: TECHNICAL EVALUATION

The technical evaluation shall be conducted on a score-based system with a maximum score of 100 marks.

No.	Evaluation Criteria	Assessment Parameters	Maximum Marks
1	Delivery Schedule Commitment	Submit a signed undertaking confirming ability to deliver goods within the stipulated LPO period. <ul style="list-style-type: none"> • Delivery period not exceeding 30 days - 20 Marks • Delivery period of 31-40 days - 10 Marks • Delivery period exceeding 40 days - 5 Marks • No undertaking submitted - 0 Marks 	20
2	Price Stability Commitment	Submit a written undertaking confirming that quoted prices shall remain fixed throughout the contract period. <ul style="list-style-type: none"> • Fixed prices for the entire contract period (12 months) - 20 Marks • Prices subject to adjustment during contract period - 5 Marks • No undertaking submitted - 0 Marks 	20
3	Relevant Experience	Submit recommendation/reference letters from institutions where similar goods/services have been supplied. References must be completed in the prescribed format, signed, and stamped. <ul style="list-style-type: none"> • One satisfactory reference - 15 Marks • Two or more satisfactory references - 30 Marks • No reference provided - 0 Marks 	30

No.	Evaluation Criteria	Assessment Parameters	Maximum Marks
4	Financial Capacity / Bank Reference	Submit a recent bank reference letter indicating account performance and availability of credit facilities. <ul style="list-style-type: none"> • Satisfactory account operation with confirmed credit facility - 30 Marks • Satisfactory account operation without credit facility - 15 Marks • Unsatisfactory account operation with credit facility - 5 Marks • No bank reference submitted - 0 Marks 	30

Technical Evaluation Threshold:

Tenderers scoring at least **80/100** and submitting valid Manufacturer's Authorization for the applicable item(s) shall qualify for Financial Evaluation. **Only authorized item(s) shall proceed to Financial Evaluation.** While Tenderers scoring below **80/100** shall be technically non-responsive and shall not proceed to Financial Evaluation.

Evaluation of Commercial Terms and Conditions of the Tender (ITT 33.1(a)):

The Procuring Entity shall determine whether the Tenders are substantially responsive to the Commercial and Contractual Terms and Conditions (e.g. Performance securities, Payment and delivery schedules).

[The Procuring Entity will highlight herein any particular requirements under the Contract which the Tenderer is required to specifically confirm or provide information to enable evaluation of Commercial Terms and Conditions of the Tender]

Evaluation Criteria (Other Factors) (ITT 33.6)

The Procuring Entity's evaluation of a Tender may take into account, in addition to the Tender Price quoted in accordance with ITT 13.8, one or more of the following factors as specified in ITT 33.2(d) and in TDS ITT 33.6, using the following criteria and methodologies.

a) Delivery schedule.

The Goods specified in the List of Goods are required to be delivered within the acceptable time range (after the earliest and before the final date, both dates inclusive) specified in Section V, Schedule of Requirements. No credit will be given to deliveries before the earliest date, and Tenders offering delivery after the final date shall be treated as non-responsive. Within this acceptable period, an adjustment of [insert the adjustment factor], will be added, for evaluation purposes only, to the Tender price of Tenders offering deliveries later than the "Earliest Delivery Date" specified in Section V, Schedule of Requirements.

[An adjustment factor of 0.5% per week of delay would be reasonable. However, the adjustment factor should not be more than the rate of Liquidated Damages to be applied in case of delay in delivery of Goods and Services under the Contract conditions.]

b) Deviation in payment schedule. [insert one of the following]

- i. tenderers shall state their Tender price for the payment schedule outlined in the SCC. Tenders shall be evaluated on the basis of this base price. tenderers are, however, permitted to state an alternative payment schedule and indicate the reduction in Tender price they wish to offer for such alternative payment schedule. The Procuring Entity may consider the alternative payment schedule and the reduced Tender price offered by the tenderer selected on the basis of the base price for the payment schedule outlined in the SCC.
or
- ii. The SCC stipulates the payment schedule specified by the Procuring Entity. If a Tender deviate from the schedule and if such deviation is considered acceptable to the Procuring Entity, the Tender will be evaluated by calculating interest earned for any earlier payments involved in the terms outlined in the Tender as compared with those stipulated in the SCC, at the rate per annum [insert adjustment rate].

c) Cost of major replacement components, mandatory spare parts, and service. [insert one of the followings]

The list of items and quantities of major assemblies, components, and selected spare parts, likely to be required during the initial period of operation specified in the TDS 15.4, is in the List of Goods. An adjustment equal to the total cost of these items, at the unit prices quoted in each Tender, shall be added to the Tender price, for evaluation purposes only.

or

The Procuring Entity will draw up a list of high-usage and high-value items of components and spare parts, along with estimated quantities of usage in the initial period of operation specified in the TDS 15.4. The total cost of these items and quantities will be computed from spare parts unit prices submitted by the tenderer and added to the Tender price, for evaluation purposes only.

or

Tenderer shall provide along with its Tender, the list of recommended spare parts for Goods offered indicating for each item of spare part the recommended quantity and unit, and total CIP final destination prices required during the initial period of operation specified in the TDS 15.4. The prices offered shall not exceed the prevailing prices charged to other parties by the Tenderer. The cost of such spare parts will not be taken into account for tender evaluation. The Procuring Entity may award the contract for spare parts to the Tenderer that is successful for the supply of Goods, by selecting at its option, from the Tender's list of recommended spare parts, such items and quantities against each as the Procuring Entity may deem appropriate at the unit prices indicated by the Tenderer but not exceeding % (present) of the cost of Goods [normally not more than 10% or 15%.]

- d) **Availability in Kenya** of spare parts and after sales services for equipment offered in the Tender.

An adjustment equal to the cost to the Procuring Entity of establishing the minimum service facilities and parts inventories if quoted separately, shall be added to the Tender price, for evaluation purposes only.

- e) **Life Cycle Costs**

If specified in TDS 33.6, an adjustment to consider the additional life cycle costs for the period specified below, such as the operating and maintenance costs of the Goods, will be added to the Tender price, for evaluation purposes only. The adjustment will be evaluated in accordance with the methodology specified below and the following information: not applicable

[Note to Procuring Entity: Life cycle costing should be used when the costs of operation and/or maintenance over the specified life of the goods are estimated to be considerable in comparison with the initial cost and may vary among different Tenders. Life cycle costs shall be evaluated on a net present value basis. If life cycle costs apply, then specify the factors required to determine them for evaluation purposes.]

[Either amend the following text as required, or delete if life cycle cost is not applicable]

- i) number of years for life cycle cost determination *[insert the number of years of economic life of Goods];*
- ii) the discount rate to be applied to determine the net present value of the life-cycle-cost is *[insert the discount rate];*
- iii) the annual operating and maintenance costs (recurrent costs) shall be determined on the basis of the following methodology: *[insert methodology E.G. This should include factors that will be used for determination of life-cycle- cost such as costs of operation and maintenance, residual value at the end of economic life of Goods, major elements that will be used for determination of cost of operation and maintenance such as fuel, power, labor, spare parts, etc. unit prices of elements such as fuel, power, etc., quantity of annual usage such as Kms or Hours of operation of Goods, Formula for calculation of LCC, etc];*
- iv) and the following information is required from tenderers *[insert any information required from tenderers, including prices e.g. Guaranteed fuel and/or power consumption, cost of labour, spare parts, etc].*

- f) **Performance and productivity of the equipment:** *[insert one of the followings]*

- i) Performance and productivity of the equipment. An adjustment representing the capitalized cost of additional operating costs over the life of the goods will be added to the Tender price, for evaluation purposes if specified in the TDS 33.6. The adjustment will be evaluated based on the drop in the guaranteed performance or efficiency offered in the Tender below the norm of 100, using the methodology specified below.
[Insert the methodology and criteria if applicable e.g. The Following aspects could be considered in the formulation of this methodology and criteria: (i) Tender price for the equipment; ii) Price of spare parts required for AAA years of operations, iii) Adjustments to tender price for omissions, deviations and exceptions to technical and commercial conditions in the tender documents; iv)

Capitalized cost savings due to the equipment efficiency at the rate of XXX (specify currency and amount) for each YYY % (percent) above the minimum ZZZ % (percent) efficiency; v) Capitalized cost for the auxiliary power consumption at PPP (specify currency and amount) per KW for AAA years; and vi) Applicable discount rate of BBB%.]

or

- ii) An adjustment to consider the productivity of the goods offered in the Tender will be added to the Tender price, for evaluation purposes only, if specified in ITT 33.6. The adjustment will be evaluated based on the cost per unit of the actual productivity of goods offered in the Tender with respect to minimum required values, using the methodology specified below.

[Insert the methodology and criteria if applicable E.G. The evaluation and comparison of responsive tenders shall be based on the total life cycle cost for XXX years, per unit of output. The life cycle cost shall be the sum of the initial purchase price of the equipment and the cost of operation in electric energy for XXX years of operation at unit cost of AAA (specify currency and amount) per kwh, discounted to net present value at YYY percent.]

g) Specific additional criteria

[Other specific additional criteria to be considered in the evaluation, and the evaluation method shall be detailed in TDS 34.6][If specific **sustainable procurement technical requirements** have been specified in Section VII- Specification, **either** state that (i) those requirements will be evaluated on a pass/fail (compliance basis) **or** otherwise (ii) in addition to evaluating those requirements on a pass/fail (compliance basis), if applicable, specify the monetary adjustments to be applied to Tender Prices for comparison purposes on account of Tenders that exceed the specified minimum sustainable procurement technical requirements.]

Multiple Contracts (ITT 33.4)

Multiple contracts will be permitted in accordance with ITT 33.4. Tenderers are evaluated on basis of Lots and the lowest evaluated tenderer identified for each Lot. The Procuring Entity will select one Option of the two Options listed below for award of Contracts.

OPTION 1

- i) If a tenderer wins only one Lot, the tenderer will be awarded a contract for that Lot, provided the tenderer meets the Eligibility and Qualification Criteria for that Lot.
- ii) If a tenderer wins more than one Lot, the tender will be awarded contracts for all won Lots, provided the tenderer meets the aggregate Eligibility and Qualification Criteria for all the Lots. The tenderer will be awarded the combination of Lots for which the tenderer qualifies and the others will be considered for award to second lowest the tenderers.

OPTION 2

The Procuring Entity will consider all possible combinations of won Lots [contract(s)] and determine the combinations with the lowest evaluated price. Tenders will then be awarded to the Tenderer or Tenderers in the combinations provided the tenderer meets the aggregate Eligibility and Qualification Criteria for all the won Lots.

Alternative Tenders

(ITT 13.1) *An alternative if permitted under ITT*

13.1, will be evaluated as follows: [insert one of the following]

“A Tenderer may submit an alternative Tender only with a Tender for the base case. The Procuring Entity shall only consider the alternative Tenders offered by the Tenderer whose

Tender for the base case was determined to be the Lowest Evaluated Tender.”

or

“A Tenderer may submit an alternative Tender with or without a Tender for the base case. The Procuring Entity shall consider Tenders offered for alternatives as specified in the Technical Specifications of Section V, Schedule of Requirements. All Tenders received, for the base case, as well as alternative Tenders meeting the specified requirements, shall be evaluated on their own merits in accordance with the same procedures, as specified in the ITT 33.”

3. MARGIN OF PREFERENCE

If the TDS so specifies, the Procuring Entity will grant a margin of preference of 15% (fifteen percent) to Tenderers offering goods manufactured, mined, extracted, grown, assembled or semi-processed in Kenya. Goods assembled or semi-processed in Kenya shall have a local content of not less than 40%.

The margin of preference will be applied in accordance with, and subject to, the following provisions:

- a) Tenderers applying for such preference on goods offered shall provide, as part of the data for qualification, such information, including details of the goods produced in Kenya, so as to determine whether, according to the classification established by the Procuring Entity, a particular category of goods or group of goods qualifies for a margin of preference.
- b) After Tenders have been received and reviewed by the Procuring Entity, goods offered in the responsive Tenders shall be assessed to ascertain they are manufactured, mined, extracted, grown, assembled or semi-processed in Kenya. Responsive tenders shall be classified into the following groups:
 - i) **Group A:** Tenders offering goods manufactured in Kenya, for which (a) labour, raw materials, and components from within Kenya account for more than forty (40) percent of the Ex-Works price; and (b) the production facility in which they will be manufactured or assembled has been engaged in manufacturing or assembling such goods at least since the date of Tender Submission date;
 - ii) **Group B:** All other Tenders offering Goods manufactured in Kenya;
 - iii) **Group C:** Tenders offering Goods manufactured outside Kenya that have been already imported or that will be imported.
- c) To facilitate this classification by the Procuring Entity, the Tenderer shall complete whichever version of the Price Schedule furnished in the Tender Documents is appropriate. Incorrect classification may render the Tender non-responsive as no reclassification will be permitted after Tender opening. Tenderers shall provide correct information especially with respect to duties, taxes etc. paid on previously imported Goods and percentage of local labour, materials and components for Goods manufactured in Kenya as any false information which cannot be supported by documentation may render the Tender non-responsive besides other sanctions for providing falsified information.
- d) The Procuring Entity will first review the Tenders to confirm the appropriateness of the Tender group classification to which Tenderers assigned their Tenders in preparing their Tender Forms and Price Schedules.
- e) All evaluated Tenders in each group will then be compared to determine the lowest evaluated Tender of each group. Such lowest evaluated Tenders shall be compared with each other and if as a result of this comparison a Tender from Group A or Group Bis the lowest, it shall be selected for the award.
- f) If as a result of the preceding comparison, the lowest evaluated Tender is a Tender from Group C, all Tenders from Group C shall be further compared with the lowest evaluated Tender from Group A after adding to the evaluated price of goods offered in each Tender from Group C, for the purpose of this further comparison only, an amount equal to 15% (fifteen percent) of the respective CIP Tender price for goods to be imported and already imported goods. Both prices shall include unconditional discounts and be corrected for arithmetical errors. If the Tender from Group A is the lowest, it shall be selected for award. If not, the lowest evaluated Tender from GroupC shall be selected as per paragraph (e) above.”

4. Post-Qualification of Tenderers (ITT 37) – not applicable

[Note for Procuring Entity to be deleted before issuing the tender documents.]

This STD for Procurement of Goods assumes that no Prequalification has taken place before tendering. However, if a Prequalification process is undertaken, the Qualification Criteria stipulated in this Section III, Evaluation and Qualification Criteria must be updated to ensure that the Tenderer and any Sub- Suppliers shall meet or continue to meet the Criteria used at the time of Prequalification.]

Post-Qualification Criteria (ITT 37.1)

In case the tender was not subject to pre-qualification, the tender that has been determined to be the lowest evaluated tenderer shall be considered for contract award, subject to meeting each of the following conditions (post qualification Criteria applied on a GO/NO GO basis). The Procuring Entity shall carry out the post- qualification of the Tenderer in accordance with ITT 37, using only the requirements specified herein. Requirements not included in the text below shall not be used in the evaluation of the Tenderer's qualifications. The minimum qualification requirements for multiple contracts will be the sum of the minimum requirements for respective individual contracts, unless otherwise specified.

[Note for Procuring Entity to be deleted before issuing the tender documents. Select requirements (criteria) for post qualification from below as relevant and appropriate for the nature, size and type of Goods and Services to be procured. Generally, for procurement of Goods, unless the value of the item is very large, the criteria for assessment of Manufacturer's technical capability should always be considered more important than its financial resources. For very small value items, the criteria for financial capability may even be omitted].

If the Tenderer is a manufacturer

a) Financial Capability

- i) The Tenderer shall demonstrate that it has access to, or has available, liquid assets, unencumbered real assets, lines of credit, and other financial means (independent of any contractual advance payment) sufficient to meet the supply cash flow of Kenya Shillings _____ [or e
- ii) Minimum average annual supply turnover of Kenya Shillings _____ [insert amount, specify a figure about 2.5 times the total Tender price)] or equivalent calculated as total certified payments received for contracts of goods manufactured and supplied within the last _____ [insert number of years). In case of multiple contracts, limitation will be placed on the number of item(s) that will be awarded to the Tenderer.

b) Experience and Technical Capacity

The Tenderer shall furnish documentary evidence to demonstrate that it meets the following experience requirement(s) using the form provided in Section IV. In case the Tenderer is a JV, experience and demonstrated technical capacity of only the JV shall be taken into account and not of individual members nor their individual experience/capacity will be aggregated unless all members of the JV have been manufacturing and supplying Goods offered in the Tender to the same technology, processing, design, materials, specifications, model number, etc. in all respects such that Goods manufactured have the same functional characteristics, performance parameters, outputs and other guarantees and fully interchangeable which shall be documented along with other required documents demonstrating capacity to the satisfaction of the Procuring Entity in case individual members claim experience. Otherwise, documents evidencing experience and technical capacity shall be in the name of the JV that submitted the Tender. Wherever the Words "Similar Goods" have been used it includes upgrades, latest and improved versions or models of similar specifications and technology. Refer to Form Exp-1 to provide the required information.

[list the requirement(s), including experience in successfully implementing sustainable procurement requirements, if specified in the tender document.] Samples of Experience Requirements:

- i) The Tenderer shall be manufacturing similar Goods for the last _____ (specify the number of years to cover a sufficiently long period ranging from 2 to 5 years depending upon the Goods to be procured).
- ii) The Tenderer shall furnish documentary evidence to demonstrate successful completion of at least _____ (Insert number) of contracts of similar Goods in the last _____ (specify number) each

contract costing at least Kenya shillings _____ equivalent and involving a supply of at least _____ percentage of required quantity (*usually the percentage is about 70-80%*) in some cases where Procuring Entity requires deliveries in a scheduled manner over a specified time, include item (iii) below.

- iii) _____ (Optional)
The installed capacity to manufacture _____ number of items (*specify the relevant item number*) shall not be less than _____ units per _____ (*specify week or month*).

- c) **(Optional) Documentary Evidence of Usage of Goods (When appropriate)**
The Tenderer shall furnish documentary evidence satisfactory to the Procuring Entity to demonstrate that similar Goods as offered in the Tender have been in successful use or operation for the last _____ years. If the Tenderer is a JV, the evidence of demonstrated usage of Goods supplied in the past shall be in the name of the JV.

If Tenderer is a Supplier:

If a Tenderer is a Supplier offering the Goods on behalf of or from a Manufacturer under Manufacturer's Authorization Form (Section IV, Tendering Forms), the Manufacturer shall demonstrate the above qualifications 4.2 (b) (i), (ii), and (iii) and the Tenderer shall demonstrate it meets the following criteria.

- i) The Tenderer shall demonstrate that it has access to, or has available, liquid assets, unencumbered real assets, lines of credit, and other financial means (independent of any contractual advance payment) sufficient to meet the supply cash flow of Kenya Shillings _____
- ii) _____ of Kenya Shillings [in sert amount] or equivalent calculated as total certified payments received for contracts in progress and/or completed within the last [insert of year] years, divided by [insert number of years] years.
- iii) Has satisfactorily and substantially completed at least _____ (*specify number*) contract(s) of a similar nature either within Kenya, the East African Community or abroad, as a prime supplier or a joint venture member, each of a minimum value in Kenya shillings _____ equivalent.

History of non-performing contracts:

Tenderer (Supplier or/and manufacturer, and each member of JV in case the Tenderer is a JV, shall demonstrate that Non-performance of a contract did not occur as a result of the default of the Tenderer, manufacturer or the member of JV as the case may be, in the last one years). The required information shall be furnished as per form CON-2].

Pending Litigation

Financial position and prospective long-term profitability of the Single Tenderer, and in the case the Tenderer is a JV, of each member of the JV, shall remain sound according to criteria established with respect to Financial Capability under paragraph I (i) above assuming that all pending litigation will be resolved against the Tenderer. Tenderer shall provide information on pending litigations as per Form CON-2.

Litigation History

There shall be no consistent history of court/arbitral award decisions against the Tenderer, in the last one years). All parties to the contract shall furnish the information on the related Form (CON-2) about any litigation or arbitration resulting from contracts completed or ongoing under its execution over the years specified. A consistent history of awards against the Tenderer or any member of a JV may result in rejection of the tender.

SECTION IV - TENDERING FORMS

Form of Tender Tenderer Information Form Tenderer JV Members Information Form

Price Schedule: Goods Manufactured Outside Kenya, to be Imported Price Schedule: Goods

Manufactured Outside Kenya, already imported Price Schedule: Goods Manufactured in Kenya Price and

Completion Schedule – Related Services Form of Tender Security – Demand Guarantee Form of

Tender Security (Tender Bond)

Form of Tender-Securing Declaration Manufacturer's Authorization Form

FORM OF TENDER

INSTRUCTIONS TO TENDERERS

INSTRUCTIONS TO TENDERERS

- i) *The Tenderer must prepare this Form of Tender on stationery with its letterhead clearly showing the Tenderer's complete name and business address.*
- ii) *All italicized text is to help Tenderer in preparing this form.*
- iii) *Tenderer must complete and sign CERTIFICATE OF INDEPENDENT TENDER DETERMINATION and the SELF DECLARATION OF THE TENDERER and TENDERER'S ELIGIBILITY-CONFIDENTIAL BUSINESS QUESTIONNAIRE all attached to this Form of Tender.*
- iv) *The Form of Tender shall include the following Forms duly completed and signed by the Tenderer.*
 - a) *Tenderer's Eligibility-Confidential Business Questionnaire*
 - b) *Certificate of Independent Tender Determination*
 - c) *Self-Declaration of the Tenderer*

Date of this Tender submission:[insert date (as day, month and year) of Tender

submission] **Tender Name and Identification**..... [insert

identification] **Alternative No.:**[insert identification No if this is a Tender

for an alternative]

To..... [Insert complete name of Procuring Entity]

- a) **No reservations:** We have examined and have no reservations to the Tendering document, including Addenda issued in accordance with Instructions to tenderers (ITT 7);
- b) **Eligibility:** We meet the eligibility requirements and have no conflict of interest in accordance with ITT 3;
- c) **Tender/Proposal-Securing Declaration:** We have not been suspended nor declared ineligible by the Procuring Entity based on execution of a Tender-Securing Declaration.
or
Proposal-Securing Declaration in Kenya in accordance with ITT 3.6;
- d) **Conformity:** We offer to supply in conformity with the Tendering document and in accordance with the Delivery Schedules specified in the Schedule of Requirements the following Goods: [insert a brief description of the Goods and Related Services];
- e) **Tender Price:** The total price of our Tender, excluding any discounts offered in item (f) below is:

Option 1, in case of one lot: Total price is: [insert the total price of the Tender in words and figures, indicating the various amounts and the respective currencies];

or

Option 2, in case of multiple lots: (a) Total price of each lot [insert the total price of each lot in words and figures, indicating the various amounts and the respective currencies]; and (b) Total price of all lots (sum of all lots) [insert the total price of all lots in words and figures, indicating the various amounts and the respective currencies];

- f) **Discounts:** The discounts offered and the methodology for their application are:

- i) The discounts offered are: *[Specify in detail each discount offered.]*
- ii) The exact method of calculations to determine the net price after application of discounts are shown below: *[Specify in detail the method that shall be used to apply the discounts];*
- g) **Tender Validity Period:** Our Tender shall be valid for the period specified in TDS 17.1 (as amended, if applicable) from the date fixed for the Tender submission deadline specified in TDS 21.1 (as amended, if applicable), and it shall remain binding upon us and may be accepted at any time before the expiration of that period;
- (h) **Performance Security:** If our Tender is accepted, we commit to obtain a performance security in accordance with the Tendering document;
- i) **One Tender per tenderer:** We are not submitting any other Tender(s) as an individual tenderer, and we are not participating in any other Tender(s) as a Joint Venture member, or as a subcontractor, and meet the requirements of ITT 3.9, other than alternative Tenders submitted in accordance with ITT 12;
- j) **Suspension and Debarment:** We, along with any of our subcontractors, suppliers, consultants, manufacturers, or service providers for any part of the contract, are not subject to, and not controlled by any entity or individual that is subject to, a temporary suspension or a debarment imposed by the Procuring Entity. Further, we are not ineligible under the Kenya laws or official regulations or pursuant to a decision of the United Nations Security Council;
- k) **State-owned enterprise or institution:** *[select the appropriate option and delete the other] [We are not a state-owned enterprise or institution] / [We are a state-owned enterprise or institution but meet the requirements of ITT 3.7];*
- l) **Commissions, gratuities, fees:** We have paid, or will pay the following commissions, gratuities, or fees with respect to the Tendering process or execution of the Contract: *[insert complete name of each Recipient, its full address, the reason for which each commission or gratuity was paid and the amount and currency of each such commission or gratuity]*

Name of Recipient	Address	Reason	Amount

(If none has been paid or is to be paid, indicate “none.”)

- m) **Binding Contract:** We understand that this Tender, together with your written acceptance thereof included in your Letter of Acceptance, shall constitute a binding contract between us, until a formal contract is prepared and executed;
- n) **Procuring Entity Not Bound to Accept:** We understand that you are not bound to accept the lowest evaluated cost Tender, the Best Evaluated Tender or any other Tender that you may receive; and
- o) **Fraud and Corruption:** We hereby certify that we have taken steps to ensure that no person acting for us or on our behalf engages in any type of Fraud and Corruption.
- p) **Code of Ethical Conduct:** We undertake to adhere by the Code of Ethics for Persons Participating in Public Procurement and Asset Disposal, copy available from___*(specify website)* during the procurement process and the execution of any resulting contract.
- q) **Collusive practices:** We hereby certify and confirm that the tender is genuine, non-collusive and made with the intention of accepting the contract if awarded. To this effect we have signed the “Certificate of Independent tender Determination” attached below.
- (r) We, the Tenderer, have completed fully and signed the following Forms as part of our

Tender:

- a) Tenderer's Eligibility; Confidential Business Questionnaire – to establish we are not in any conflict to interest.
- b) Certificate of Independent Tender Determination – to declare that we completed the tender without colluding with other tenderers.
- c) Self-Declaration of the Tenderer – to declare that we will, if awarded a contract, not engage in any form of fraud and corruption.
- d) Declaration and commitment to the Code of Ethics for Persons Participating in Public Procurement and Asset Disposal.

Further, we confirm that we have read and understood the full content and scope of fraud and corruption as informed in “**Appendix 1- Fraud and Corruption**” attached to the Form of Tender.

Name of the tenderer: **[insert complete name of the tenderer]*

Name of the person duly authorized to sign the Tender on behalf of the tenderer:

***[insert complete name of person duly authorized to sign the Tender]*

Title of the person signing the Tender: *[insert complete title of the person signing the Tender]*

Signature of the person named above: *[insert signature of person whose name and capacity are shown*

above] **Date signed** *[insert date of signing]* **day of** *[insert month], [insert year]*

*: In the case of the Tender submitted by a Joint Venture specify the name of the Joint Venture as tenderer.

** : Person signing the Tender shall have the power of attorney given by the tenderer. The power of attorney shall be attached with the Tender Schedules.

CERTIFICATE OF INDEPENDENT TENDER DETERMINATION

I, the undersigned, in submitting the accompanying Letter of Tender to the _____ [Name of Procuring Entity] for: _____ [Name and number of tender] in response to the request for tenders made by: _____ [Name of Tenderer] do hereby make the following statements that I certify to be true and complete in every respect:

I certify, on behalf of _____ [Name of Tenderer] that:

1. I have read and I understand the contents of this Certificate;
2. I understand that the Tender will be disqualified if this Certificate is found not to be true and complete in every respect;
3. I am the authorized representative of the Tenderer with authority to sign this Certificate, and to submit the Tender on behalf of the Tenderer;
4. For the purposes of this Certificate and the Tender, I understand that the word "competitor" shall include any individual or organization, other than the Tenderer, whether or not affiliated with the Tenderer, who:
 - a) has been requested to submit a Tender in response to this request for tenders;
 - b) could potentially submit a tender in response to this request for tenders, based on their qualifications, abilities or experience;
5. The Tenderer discloses that [check one of the following, as applicable]:
 - a) The Tenderer has arrived at the Tender independently from, and without consultation, communication, agreement or arrangement with, any competitor;
 - b) the Tenderer has entered into consultations, communications, agreements or arrangements with one or more competitors regarding this request for tenders, and the Tenderer discloses, in the attached document(s), complete details thereof, including the names of the competitors and the nature of, and reasons for, such consultations, communications, agreements or arrangements;
6. In particular, without limiting the generality of paragraphs (5)(a) or (5)(b) above, there has been no consultation, communication, agreement or arrangement with any competitor regarding:
 - a) prices;
 - b) methods, factors or formulas used to calculate prices;
 - c) the intention or decision to submit, or not to submit, a tender; or
 - d) the submission of a tender which does not meet the specifications of the request for Tenders; except as specifically disclosed pursuant to paragraph (5)(b) above;
7. In addition, there has been no consultation, communication, agreement or arrangement with any competitor regarding the quality, quantity, specifications or delivery particulars of the works or services to which this request for tenders relates, except as specifically authorized by the procuring authority or as specifically disclosed pursuant to paragraph (5)(b) above;
8. the terms of the Tender have not been, and will not be, knowingly disclosed by the Tenderer, directly or indirectly, to any competitor, prior to the date and time of the official tender opening, or of the awarding of the Contract, whichever comes first, unless otherwise required by law or as specifically disclosed pursuant to paragraph (5)(b) above.

Name.....Title.....Date..... Stamp.....

SELF-DECLARATION FORMS

FORM SD1

SELF DECLARATION THAT THE PERSON/TENDERER IS NOT DEBARRED IN THE MATTER OF THE PUBLIC PROCUREMENT AND ASSET DISPOSAL ACT 2015.

I of Post Office Box.....being a resident of in the Republic of.....do hereby make a statement as follows:-

1. THAT I am the Company Secretary/ Chief Executive/Managing Director/Principal Officer/Director of (*insert name of the Company*) who is a Bidder in respect of **Tender No.** for..... (*insert tender title/description*) for.....(*insert name of the Procuring entity*) and duly authorized and competent to make this statement.
2. THAT the aforesaid Bidder, its Directors and subcontractors have not been debarred from participating in procurement proceeding under Part IV of the Act.
3. THAT what is deponed to herein above is true to the best of my knowledge, information and belief.

.....
(Title)

.....
(Signature)

.....
(Date)

Bidder Official Stamp

FORM SD2

SELF DECLARATION THAT THE PERSON/TENDERER WILL NOT ENGAGE IN ANY CORRUPT OR FRAUDULENT PRACTICE

I, of P.O. Box.....being a resident of..... inthe Republic of..... do hereby make a statement as follows: -

1. THAT I am the Chief Executive/Managing Director/Principal Officer/Director of..... *(insert name of the Company)* who is a Bidder in respect of **Tender No.** for..... *(Insert tender title/description)* for..... *(insert name of the Procuring entity)* and duly authorized and competent to make this statement.
2. THAT the aforesaid Bidder, its servants and/or agents /subcontractors will not engage in any corrupt or fraudulent practice and has not been requested to pay any inducement to any member of the Board, Management, Staff and/or employees and/or agents of*(insert name of the Procuring entity)* which is the procuring entity.
3. THAT the aforesaid Bidder, its servants and/or agents /subcontractors have not offered any inducement to any member of the Board, Management, Staff and/or employees and/or agents of*(name of the procuring entity)*.
4. THAT the aforesaid Bidder will not engage/has not engaged in any corrosive practice with other bidders participating in the subject tender.
5. THAT what is deponed to herein above is true to the best of my knowledge information and belief.

.....
(Title)

.....
(Signature)

.....
(Date)

Bidder's Official Stamp

DECLARATION AND COMMITMENT TO THE CODE OF ETHICS

I..... (Person) on behalf of (*Name of the Business/ Company/Firm*).....declare that I have read and fully understood the contents of the Public Procurement & Asset Disposal Act, 2015, Regulations and the Code of Ethics for persons participating in Public Procurement and Asset Disposal and my responsibilities under the Code.

I do hereby commit to abide by the provisions of the Code of Ethics for persons participating in Public Procurement and Asset Disposal.

Name of Authorized signatory.....

Sign.....

Position.....

Office address..... Telephone.....

E-mail.....

Name of the Firm/Company.....

Date.....

(Company Seal/ Rubber Stamp where applicable)

Witness

Name

Sign.....

Date.....

APPENDIX 1- FRAUD AND CORRUPTION

(Appendix 1 shall not be modified)

1. Purpose

The Government of Kenya's Anti-Corruption and Economic Crime laws and their sanction's policies and procedures, Public Procurement and Asset Disposal Act (*no. 33 of 2015*) and its Regulation, and any other Kenya's Acts or Regulations related to Fraud and Corruption, and similar offences, shall apply with respect to Public Procurement Processes and Contracts that are governed by the laws of Kenya.

2. Requirements

The Government of Kenya requires that all parties including Procuring Entities, Tenderers, (applicants/proposers), Consultants, Contractors and Suppliers; any Sub-contractors, Sub-consultants, Service providers or Suppliers; any Agents (whether declared or not); and any of their Personnel, involved and engaged in procurement under Kenya's Laws and Regulation, observe the highest standard of ethics during the procurement process, selection and contract execution of all contracts, and refrain from Fraud and Corruption and fully comply with Kenya's laws and Regulations as per paragraphs 1.1 above.

Kenya's public procurement and asset disposal act (*no. 33 of 2015*) under Section 66 describes rules to be followed and actions to be taken in dealing with Corrupt, Coercive, Obstructive, Collusive or Fraudulent practices, and Conflicts of Interest in procurement including consequences for offences committed. A few of the provisions noted below highlight Kenya's policy of no tolerance for such practices and behavior:

- 1) a person to whom this Act applies shall not be involved in any corrupt, coercive, obstructive, collusive or fraudulent practice; or conflicts of interest in any procurement or asset disposal proceeding;
- 2) A person referred to under subsection (1) who contravenes the provisions of that subsection commits an offence;
- 3) Without limiting the generality of the subsection (1) and (2), the person shall be—
 - a) disqualified from entering into a contract for a procurement or asset disposal proceeding; or
 - b) if a contract has already been entered into with the person, the contract shall be voidable;
- 4) The voiding of a contract by the procuring entity under subsection (7) does not limit any legal remedy the procuring entity may have;
- 5) An employee or agent of the procuring entity or a member of the Board or committee of the procuring entity who has a conflict of interest with respect to a procurement:-
 - a) shall not take part in the procurement proceedings;
 - b) shall not, after a procurement contract has been entered into, take part in any decision relating to the procurement or contract; and
 - c) shall not be a subcontractor for the bidder to whom was awarded contract, or a member of the group of bidders to whom the contract was awarded, but the subcontractor appointed shall meet all the requirements of this Act.
- 6) An employee, agent or member described in subsection (1) who refrains from doing anything prohibited under that subsection, but for that subsection, would have been within his or her duties shall disclose the conflict of interest to the procuring entity;
- 7) If a person contravenes subsection (1) with respect to a conflict of interest described in subsection (5)(a) and the contract is awarded to the person or his relative or to another person in whom one of them had a direct or indirect pecuniary interest, the contract shall be terminated and all costs incurred by the public entity shall be made good by the awarding officer. Etc.

In compliance with Kenya's laws, regulations and policies mentioned above, the Procuring Entity:

- a) Defines broadly, for the purposes of the above provisions, the terms set forth below as follows:
 - i) "corrupt practice" is the offering, giving, receiving, or soliciting, directly or indirectly, of anything of value to influence improperly the actions of another

party;

- ii) “fraudulent practice” is any act or omission, including misrepresentation, that knowingly or recklessly misleads, or attempts to mislead, a party to obtain financial or other benefit or to avoid an obligation;
- iii) “collusive practice” is an arrangement between two or more parties designed to achieve an improper purpose, including to influence improperly the actions of another party;
- iv) “coercive practice” is impairing or harming, or threatening to impair or harm, directly or indirectly, any party or the property of the party to influence improperly the actions of a party;
- v) “obstructive practice” is:
 - deliberately destroying, falsifying, altering, or concealing of evidence material to the investigation or making false statements to investigators in order to materially impede investigation by Public Procurement Regulatory Authority (PPRA) or any other appropriate authority appointed by Government of Kenya into allegations of a corrupt, fraudulent, coercive, or collusive practice; and/or threatening, harassing, or intimidating any party to prevent it from disclosing its knowledge of matters relevant to the investigation or from pursuing the investigation; or
 - acts intended to materially impede the exercise of the PPRA's or the appointed authority's inspection and audit rights provided for under paragraph 2.3 e. below.

- b) Defines more specifically, in accordance with the above procurement Act provisions set forth for fraudulent and collusive practices as follows:

"fraudulent practice" includes a misrepresentation of fact in order to influence a procurement or disposal process or the exercise of a contract to the detriment of the procuring entity or the tenderer or the contractor, and includes collusive practices amongst tenderers prior to or after tender submission designed to establish tender prices at artificial non-competitive levels and to deprive the procuring entity of the benefits of free and open competition.

- c) Rejects a proposal for award¹ of a contract if PPRA determines that the firm or individual recommended for award, any of its personnel, or its agents, or its sub-consultants, sub-contractors, service providers, suppliers and/ or their employees, has, directly or indirectly, engaged in corrupt, fraudulent, collusive, coercive, or obstructive practices in competing for the contract in question;
- d) Pursuant to the Kenya's above stated Acts and Regulations, may sanction or debar or recommend to appropriate authority (ies) for sanctioning and debarment of a firm or individual, as applicable under the Acts and Regulations;
- e) Requires that a clause be included in Tender documents and Request for Proposal documents requiring (i) Tenderers (applicants/proposers), Consultants, Contractors, and Suppliers, and their Sub-contractors, Sub-consultants, Service providers, Suppliers, Agents personnel, permit the PPRA or any other appropriate authority appointed by Government of Kenya to inspect² all accounts, records and other documents relating to the procurement process, selection and/or contract execution, and to have them audited by auditors appointed by the PPRA or any other appropriate authority appointed by Government of Kenya; and
- f) Pursuant to Section 62 of the above Act, requires Applicants/Tenderers to submit along with their Applications/Tenders/Proposals a “Self-Declaration Form” as included in the procurement document declaring that they and all parties involved in the procurement process and contract execution have not engaged/will not engage in any corrupt or fraudulent practices.

¹For the avoidance of doubt, a party's ineligibility to be awarded a contract shall include, without limitation, (i) applying for pre-qualification, expressing interest in a consultancy, and tendering, either directly or as a nominated sub-contractor, nominated consultant, nominated manufacturer or supplier, or nominated service

provider, in respect of such contract, and (ii) entering into an addendum or amendment introducing a material modification to any existing contract.

² *Inspections in this context usually are investigative (i.e., forensic) in nature. They involve fact-finding activities undertaken by the Investigating Authority or persons appointed by the Procuring Entity to address specific matters related to investigations/audits, such as evaluating the veracity of an allegation of possible Fraud and Corruption, through the appropriate mechanisms. Such activity includes but is not limited to: accessing and examining a firm's or individual's financial records and information, and making copies thereof as relevant; accessing and examining any other documents, data and information (whether in hard copy or electronic format) deemed relevant for the investigation/audit, and making copies thereof as relevant; interviewing staff and other relevant individuals; performing physical inspections and site visits; and obtaining third party verification of information.*

TENDERER INFORMATION FORM

[The tenderer shall fill in this Form in accordance with the instructions indicated below. No alterations to its format shall be permitted and no substitutions shall be accepted.]

Date: *[insert date (as day, month and year) of Tender submission]*

Tender Name and Identification*[Insert identification*

Alternative No.: *[insert identification No if this is a Tender for an alternative]* Page _____ of _____ pages

1. Tenderer's Name <i>[insert Tenderer's legal name]</i>
2. In case of JV, legal name of each member: <i>[insert legal name of each member in JV]</i>
3. Tenderer's actual or intended country of registration: <i>[insert actual or intended country of registration]</i>
4. Tenderer's year of registration: <i>[insert Tenderer's year of registration]</i>
5. Tenderer's Address in country of registration: <i>[insert Tenderer's legal address in country of registration]</i>
6. Tenderer's Authorized Representative Information Name: <i>[insert Authorized Representative's name]</i> Address: <i>[insert Authorized Representative's Address]</i> Telephone/Fax numbers: <i>[insert Authorized Representative's telephone/fax numbers]</i> Email Address: <i>[insert Authorized Representative's email address]</i>
7. Attached are copies of original documents of <i>[check the box(es) of the attached original documents]</i> <input type="checkbox"/> For Kenyan Tenderers a current tax clearance certificate or tax exemption certificate issued by the Kenya Revenue Authority in accordance with ITT 3.14. <input type="checkbox"/> Articles of Incorporation (or equivalent documents of constitution or association), and/or documents of registration of the legal entity named above, in accordance with ITT 3.4. <input type="checkbox"/> In case of JV, letter of intent to form JV or JV agreement, in accordance with ITT 3.1. <input type="checkbox"/> In case of state-owned enterprise or institution, in accordance with ITT 4.6 documents establishing: (i) Legal and financial autonomy (ii) Operation under commercial law (iii) Establishing that the tenderer is not under the supervision of the Procuring Entity
2. Included are the organizational chart and a list of Board of Directors,

TENDERER'S ELIGIBILITY- CONFIDENTIAL BUSINESS QUESTIONNAIRE FORM

a) Instruction to Tenderer

Tender is instructed to complete the particulars required in this Form, *one form for each entity if Tender is a JV*. Tenderer is further reminded that it is an offence to give false information on this Form.

A. Tenderer's details

	ITEM	DESCRIPTION
1	Name of the Procuring Entity	
2	Name of the Tenderer	
3	Full Address and Contact Details of the Tenderer. 1. Country 2. City 3. Location 4. Building 5. Floor 6. Postal Address 7. Name and email of contact person.	
4	Reference Number of the Tender	KNH/T/43/2026- 2027
5	Date and Time of Tender Opening	29th June 2026 at 10:00 am
6	Current Trade License No and Expiring date	
7	Maximum value of business which the Tenderer handles.	
8		

General and Specific Details

b) Sole Proprietor, provide the following details.

Name in full _____

Age _____ Nationality _____

Country of Origin _____ Citizenship _____

c) Partnership, provide the following details.

	Names of Partners	Nationality	Citizenship	% Shares owned
1				
2				
3				

(d) Registered Company, provide the following details.

i) Private or public Company _____

ii) the Company- State the nominal and issued capital of

Nominal Kenya Shillings (Equivalent)

Issued Kenya Shillings (Equivalent)

iii) Give details of Directors as follows.

	Names of Director	Nationality	Citizenship	% Shares owned
1				
2				
3				

(e) Interest of the Firm in the Procuring Entity. DISCLOSURE OF INTEREST-

(i) Are there any person/persons in (Name of Procuring Entity) who has an interest or relationship in this firm? Yes/No.....

If yes, provide details as follows.

	Names of Person	Designation in the Procuring Entity	Interest Relationship or with Tenderer
1			
2			
3			

(ii) Conflict of interest disclosure

	Type of Conflict	Disclosure YES OR NO	If YES provide details of the relationship with Tenderer
1	Tenderer is directly or indirectly controlled by or is under common control with another tenderer.		
2	Tenderer receives or has received any direct or indirect subsidy from another tenderer.		
3	Tenderer has the same legal representative as another tenderer		
4	Tender has a relationship with another tenderer, directly or through common third parties that puts it in a position to influence the tender of another tenderer, or influence the decisions of the Procuring Entity regarding this tendering		

	Type of Conflict	Disclosure YES OR NO	If YES provide details of the relationship with Tenderer
	process.		
5	Any of the Tenderer's affiliates participated as a consultant in the preparation of the design or technical specifications of the works that are the subject of the tender.		
6	Tenderer would be providing goods, works, non-consulting services or consulting services during implementation of the contract specified in this Tender Document.		
7	Tenderer has a close business or family relationship with a professional staff of the Procuring Entity who are directly or indirectly involved in the preparation of the Tender document or specifications of the Contract, and/or the Tender evaluation process of such contract.		
8	Tenderer has a close business or family relationship with a professional staff of the Procuring Entity who would be involved in the implementation or supervision of the Contract.		
9	Has the conflict stemming from such relationship stated in item 7 and 8 above been resolved in a manner acceptable to the Procuring Entity throughout the tendering process and execution of the Contract?		

(f) Certification

On behalf of the Tenderer, I certify that the information

given above is correct. Full Name _____

Title or Designation _____

(Signature)

(Date)



KENYATTA NATIONAL HOSPITAL

Address: P.O BOX 20723-00202, Nairobi.

Telephone: 020 2726300-4 | 020 4243000 | 020 7244000

Cellphone Numbers: 0730 643 000 | 0709 854 000 | Email: knhadmin@knh.or.ke

TENDERER DATA CONSENT FORM

Tender Number: _____

Tender Description: _____

Kenya National Hospital is committed to processing your personal information in accordance with the Hospital's Data Protection Policy, Data Protection Act, 2019 and its Regulations.

The personal data submitted in the tender as detailed will therefore be processed in line with the relevant Data Protection, Policies, Laws and Regulations in the way(s) and purpose(s) detailed in this Tenderer Data Consent Form.

I/we _____ hereby give explicit consent to processing of my personal data by Kenya National Hospital for the purposes of compliance with the Data Protection Act, 2019.

Signed:

Name: (tenderers name): _____

Signature: _____

Date: _____

Stamp: _____



SUPPLIER PERFORMANCE EVALUATION AND COMMITMENT FORM

Procuring Entity: _____

Contract/Tender No.: _____

Tender Description: _____

Contract Period: From _____ To _____

Name of Supplier: _____

Supplier Address: _____

Contact Person: _____

Telephone: _____ Email: _____

SECTION A: SUPPLIER PERFORMANCE COMMITMENT

We, the undersigned supplier, hereby commit to providing goods/services/works in accordance with the contract terms, specifications, delivery timelines and applicable procurement laws.

1. Timely delivery of goods/services as per contract schedule.
2. Compliance with technical specifications and quality standards.
3. Prompt response to orders, communication and service requests.
4. Availability of adequate technical support and after-sales service where applicable.
5. Compliance with all statutory obligations including taxes, licenses, and regulatory requirements.
6. Acceptance to collect goods rejected when on use and refund of payments already made.
7. Immediate notification to the procuring entity of any challenges that may affect contract performance.

DECLARATION

Supplier Declaration

I certify and commit to fulfilling all contractual obligations as outlined above and the same shall form the basis for performance evaluation under **SECTION B** below.

Name: _____

Designation: _____

Signature: _____

Date: _____

Stamp: _____

SECTION B: PERFORMANCE EVALUATION BY PROCURING ENTITY

Evaluation Criteria	Score Guidance	Score Awarded (1-5)
Quality of Goods/Services	100% – All items accepted with no rejection.	
	40% – Goods rejected but replaced within agreed timelines.	
	0% – Goods rejected and no replacement supplied.	
Timeliness of Delivery	100% – All deliveries within agreed timelines.	
	40% – Delivery delayed leading to cancellation and replacement of expired LPO.	
	0% – Delivery failure resulting in cancellation of LPO without supply.	
Responsiveness & Communication	100% – Prompt response within 24 hours.	
	40% – Delayed response but issue resolved.	
	0% – No response or persistent failure to respond.	

Compliance with Contract Terms	100% – Full compliance with contract terms and specifications.	
	40% – Minor deviations corrected after notification.	
	0% – Major breach of contract terms.	
Acceptance to Collect Rejected Goods and Refund Payments Made	100% – Supplier collects rejected goods and refunds payments promptly.	
	40% – Goods collected but refund delayed.	
	0% – Supplier fails to collect rejected goods or refund payments.	
Value for Money	100% – Goods perform as per intended purpose.	
	40% – Goods rejected while on use but supplier accepts responsibility.	
	0% – Goods rejected while on use and supplier fails to accept responsibility.	

Additional Performance Incidents Affecting Scores

Rejection of goods without replacement – 0 score under Quality.

Rejection of goods with acceptable replacement – 40% score under Quality.

Cancellation and replacement of expired LPO due to supplier delay – 40% score under Delivery.

Cancellation of LPO due to supplier failure – 0 score under Delivery. Failure to respond to purchase order – 0 score under Responsiveness.

Issuing of credit note will be calculated based on value of credit note against the total value of the LPO.

Repeated delivery of substandard goods – Automatic contract performance review and possible termination.

Acceptance of termination of contract on goods rejected while on use. i.e. value for money.

Overall Rating Calculation

Rating	Percentage Score / Interpretation
5	90–100% – Excellent
4	80–89% – Very Good
3	70–79% – Satisfactory
2	60–69% – Poor
1	Below 60% – Unsatisfactory



KENYATTA NATIONAL HOSPITAL

Address: P.O BOX 20723-00202, Nairobi.

Telephone: 020 2726300-4 | 020 4243000 | 020 7244000

Cellphone Numbers: 0730 643 000 | 0709 854 000 | Email: knhadmin@knh.or.ke

INSPECTION AND ACCEPTANCE OF GOODS

The Supplier hereby acknowledges and agrees that all goods, works, or services supplied under the contract shall be subject to inspection and acceptance by the Procuring Entity through the Inspection and Acceptance Committee (IAC).

The Supplier further commits to the following;

1. To deliver goods, works, or services strictly in accordance with the specifications, quantities, and quality standards stipulated in the contract and tender documents.
2. To ensure that the goods supplied conform to the approved samples, where applicable, and meet all technical and regulatory requirements.
3. To accept that the Procuring Entity may subject the goods to independent testing by accredited third-party agencies where necessary to confirm compliance.
4. To bear the cost of testing and certification where such testing is required due to doubts regarding compliance with specifications.
5. To collect and replace rejected goods promptly where inspection confirms that the goods supplied do not meet the required specifications.
6. To refund any payments already made where goods supplied are rejected and cannot be replaced in accordance with the contract terms.
7. To acknowledge that failure to comply with the inspection and acceptance requirements may lead to contract enforcement measures, including rejection of goods, cancellation of LPO, or termination of the contract in accordance with applicable procurement laws and contract provisions.

Supplier Declaration

I/We hereby certify that we have read, understood, and agree to comply with the above Inspection and Acceptance requirements and commitments.

Name of Supplier: _____

Authorized Representative: _____

Designation: _____

Signature: _____

Date: _____

Company Stamp: _____

Procedure for Inspection and Acceptance

Upon delivery of goods, works, or services, the Inspection and Acceptance Committee (IAC) shall conduct inspection and verification before acceptance.

a) Documents to be used during Inspection

The Inspection and Acceptance Committee shall inspect the goods using the following documents;

1. Local Purchase Order (LPO)
2. Supplier Delivery Note
3. Tender Document as submitted by the Supplier
4. Contract Agreement (where applicable)
5. Approved Sample (where applicable)

b) Verification during Inspection

The Inspection and Acceptance Committee shall verify the following;

- i. Conformance to the awarded products as specified in the bid/tender document.
- ii. Conformance to the approved sample, where applicable.
- iii. Compliance with technical specifications indicated in the contract.
- iv. Quantity delivered against the quantities specified in the LPO.
- v. Country of origin, where applicable.
- vi. Product labeling, packaging and expiry dates where applicable.
- vii. Functional performance, where applicable.

c) Third-Party Testing and Certification

Where goods require specialized testing or regulatory verification, samples may be submitted to recognized third-party testing agencies for independent verification and certification at suppliers cost but limited to;

Such agencies may include;

- i. Kenya Bureau of Standards (KEBS)
- ii. Government Chemist Department
- iii. National Public Health Laboratories
- iv. Other accredited regulatory authorities depending on the nature of the goods.

The third-party testing agency shall issue a certificate or a report confirming compliance and conformance.

d) Rejection of Non-Compliant Goods

If the goods fail and are rejected;

- i. The Inspection and Acceptance Committee shall document the non-compliance.
- ii. The committee shall recommend rejection, replacement or corrective action in accordance with the contract terms.
- iii. The supplier shall be formally notified of the rejection and required to replace where applicable.

TENDERER'S JV MEMBERS INFORMATION FORM

[The tenderer shall fill in this Form in accordance with the instructions indicated below. The following table shall be filled in for the tenderer and for each member of a Joint Venture]].

Date:*[insert date (as day, month and year) of Tender submission]*.

Tender Name and Identification:*[insert identification Alternative No.:**[insert identification No if this is a Tender for an alternative].*

Page _____ of _____ pages

1.	Tenderer's Name: <i>[insert Tenderer's legal name]</i>
2.	Tenderer's JV Member's name: <i>[insert JV's Member legal name]</i>
3.	Tenderer's JV Member's country of registration: <i>[insert JV's Member country of registration]</i>
4.	Tenderer's JV Member's year of registration: <i>[insert JV's Member year of registration]</i>
5.	Tenderer's JV Member's legal address in country of registration: <i>[insert JV's Member legal address in country of registration]</i>
6.	Tenderer's JV Member's authorized representative information Name: <i>[insert name of JV's Member authorized representative]</i> Address: <i>[insert address of JV's Member authorized representative]</i> Telephone/Fax numbers: <i>[insert telephone/fax numbers of JV's Member authorized representative]</i> Email Address: <i>[insert email address of JV's Member authorized representative]</i>
7.	Attached are copies of original documents of <i>[check the box(es) of the attached original documents]</i> <input type="checkbox"/> Articles of Incorporation (or equivalent documents of constitution or association), and/or registration documents of the legal entity named above, in accordance with ITT 4.4. <input type="checkbox"/> In case of a state-owned enterprise or institution, documents establishing legal and financial autonomy, operation in accordance with commercial law, and that they are not under the supervision of the Procuring Entity, in accordance with ITT 4.6.
8.	Included are the organizational chart and a list of Board of Directors.

Price Schedule Forms

[The tenderer shall fill in these Price Schedule Forms in accordance with the instructions indicated. The list of line items in column 1 of the Price Schedules shall coincide with the List of Goods and Related Services specified by the Procuring Entity in the Schedule of Requirements.]

SECTION VI: SCHEDULE OF REQUIREMENTS

TECHNICAL SPECIFICATIONS FOR CALIBRATION TOOLS

ITEM 1: TECHNICAL SPECIFICATIONS FOR VITAL SIGNS SIMULATORS

General specifications	
Temperature	Operating 10 °C to 40 °C (50 °F to 104 °F)
	Storage -20 °C to +60 °C (-4 °F to 140 °F)
Humidity	10 % to 90 % non-condensing
Altitude	3,000 meters (9,843 ft)
Display	LCD colour display
Communication	USB device upstream port Mini-B connector for control by a Computer USB host controller port Type A, 5 V output, 0.5 A max load. Connector for keyboard, barcode reader, and printer Wireless IEEE 82.15.4 for control by a computer
Power	Lithium-ion rechargeable battery
Battery charger	100 V to 240 V input, 15 V/2.0 A output. For best performance, the battery charger should be connected to a properly-grounded ac receptacle
Battery life	9 hours (minimum), 100 NIBP cycles typical
Safety standards	IEC/EN 61010-1 3rd Edition; Pollution degree 2 CAT None
Certifications	CE, CSA, C-TICK N10140 , RoHS
Electromagnetic compatibility (EMC)	IEC 61326-1:2006
Normal-sinus-rhythm waveform	
ECG reference	The ECG amplitudes specified are for Lead II (calibration), from the Baseline to the peak of the R wave. All other leads are proportional
Normal sinus rhythm	12-lead configuration with independent outputs referenced to right leg (RL). Output to 10 universal ECG jacks, color-coded to AHA and IEC standards
High-level output	0.5 V/mV • } 5 % of the ECG amplitude setting available on a BNC connector
Amplitude	0.05 mV to 0.5 mV (0.05 mV steps); 0.5 mV to 5.0 mV (0.25 mV steps) Other leads are proportional to Lead II (reference lead) in percentage per: Lead I: 70 Lead V3: 100 Lead V4: 120

	Lead V5: 112 Lead V6: 80 Lead II: 100 Lead III: 30 Lead V1: 24 Lead V2: 48
Amplitude accuracy	• } (2 % of setting + 0.05 mV)
ECG rate	10 BPM to 360 BPM in 1 BPM steps
Rate accuracy	• } 1 % of setting
ECG waveform selection	Adult (80 ms) or pediatric (40 ms) QRS duration
ST-segment elevation	Adult mode only. -0.8 mV to +0.8 mV (0.1 mV steps). Additional steps: + 0.05 mV and - 0.05 mV
Power-on default	60 BPM, 1.0 mV, adult QRS and ST-segment elevation of 0 mV
Pacemaker waveform	
Pacer pulse	Amplitude 0 (off), ±2, ±4, ±6, ±8, ±10, ±12, ±14, ±16, ±18, ±20, ±50, ±100, ±200, ±500, and ±700 mV for lead II (reference lead) Accuracy Reference lead II: • } (5 % setting + 0.2 mV) All other leads: • } (10 % setting + 0.4 mV)
	MUST provide Technical and Operators manuals(Hard/Soft copies)
Pacer pulse width	0.1 ms, 0.2 ms, 0.5 ms, 1 ms, and 2 ms • } 5 %
Paced arrhythmias	Atrial 80 BPM Asynchronous 75 BPM Demand with frequent sinus beats Demand with occasional sinus beats Atrio-ventricular sequential Noncapture (one time) Nonfunction
Power-on default	Amplitude 5 mV, width 1 ms, atrial waveform
Arrhythmia	
Baseline NSR	80 BPM
PVC focus	Left focus, standard timing (except where specified)
Supraventricular arrhythmia	Atrial fibrillation (coarse or fine); atrial flutter; sinus arrhythmia; missed beat (one time); atrial tachycardia; paroxysmal atrial tachycardia; nodal rhythm; and supraventricular tachycardia
Premature arrhythmia	Premature atrial contraction (PAC); premature nodal contraction (PNC); PVC1 left ventricular; PVC1 left ventricular, early; PVC1 left ventricular, R on T; PVC2 right ventricular; PVC2 right ventricular, early; PVC2 right ventricular, R on T; and multifocal PVCs
Ventricular arrhythmia	PVCs 6, 12, or 24 per minute; frequent multifocal PVCs; bigeminy; trigeminy; multiple PVCs (one-time run of 2, 5, or 11 PVCs); monoventricular tachycardia (120 to 300 BPM)

	in 5 BPM steps); poly-ventricular tachycardia (5 types); ventricular fibrillation (coarse or fine); and a systole
Conduction defect	First-, second-, or third-degree heart block; and right- or left-bundle branch block
Advanced cardiac life support	Shock able pulse less arrest rhythms Ventricular fibrillation (coarse), ventricular fibrillation (fine), unstable polymorphic ventricular tachycardia Non-shock able pulse less arrest rhythms systole Symptomatic bradycardia Sinus bradycardia (< 60 BPM) 2nd degree AV block, mobitz type I 2nd degree AV block, mobitz type II Complete/3rd degree AV block Right bundle branch block Left bundle branch block
Advanced cardiac life support	Symptomatic tachycardia: regular narrow-complex tachycardia (QRS < 0.12 seconds) Sinus tachycardia > 150 BPM Supraventricular Tachycardia Symptomatic tachycardia: regular wide-complex tachycardias (QRS ≥ 0.12 seconds) Sinus tachycardia > 150 BPM Supraventricular tachycardia SVT with aberrancy Irregular tachycardia Atrial fibrillation (coarse and fine), atrial flutter, unstable monomorphic ventricular tachycardia (120 BPM to 300 BPM), torsade de pointes/polymorphic ventricular tachycardia (long QT interval)
Manuals	MUST provide Technical and Operators manuals(Hard/Soft copies
Training	Onsite Training for Biomedical Engineering Personnel on Application and Maintenance.
CALIBRATION	<ul style="list-style-type: none"> ➤ Valid Certificate of Calibration with Test Data ➤ Demonstrate Traceability to National / International Standards. ➤ Date Calibrated
Key Features	

	<ul style="list-style-type: none"> ➤ All-in-one complete monitor testing ➤ Stay-connected ECG posts for easy/secure ECG snap and lead connections ➤ Custom SpO2 r-curve for accurate testing of the latest oximetry technologies ➤ Static pressure linearity testing ➤ Repeatable NIBP simulation (± 2 mm Hg) for dynamic pressure ➤ repeatability testing ➤ Physiologically synchronized pulses across all parameters ➤ Barcode scanning and direct data capture, printing functionality ➤ Onboard, customizable patient pre-sets and auto sequences ➤ Multi-language user interface ➤ Integrated, easily replaceable, long-life battery ➤ Wireless communication for remote PC control of test device, as well as data transfer and automated regulatory reporting* ➤ .Ansur Test Executive version 2.9.6 or greater on your PC to communicate with the product.
--	--

Accessories

- CI-3 Cardiac Output Box.
- Mini-DIN to DIN IBP Adapter.
- NIBP Test Chamber 500 ML.
- Ansur Test SoftwarePlug-In.
- Cable kits
- HP/Philips Intellivue Cable Set.
- GE Marquette Eagle/ Dash/Solar Cable Set.
- SpaceLabs Ultraview Cable set

- Welch Allyn/Propaq Cable Set.
- Drager Infinity Cable Set.
- Nihon Kohden Cable Set.

- Accessory Kit
- Oximeter SPO2 optical Emitter and Detector.
- Two Cardiac Output Adapter Complete with Injectable temperature Connector and Blood Temperature Connector.
- Two Temperature cable.
- Two IBP cable.
- Two BP cable.
- Two Maternal Thigh reference Plate Complete with 4 jumper wire and Fetal Scalp Electrodes.

ITEM 2: TECHNICAL SPECIFICATIONS FOR ELECTROSURGICAL ANALYZER

Power	
Power	100 V ac, 115 V ac, 230 V ac, 50 Hz / 60 Hz, universal input, 100 V/115 V: 20 VA, 230 V: 30 VA
Backup power	
Measurements and tests specifications	
Generator Output	RF leakage: from active electrode or neutral plate with an open or closed load circuit
Mode of Operation	Manual or user-programmable. Control remotely with PC utilizing accessory Ansur QA-ES Plug-in software and RS232 communication cable connection.
Load resistance	Variable: 0 Ω , 10 Ω , 20 Ω , 25 Ω to 2500 Ω (by 25 Ω), 2500 Ω to 5200 Ω (by 100 Ω); DC Accuracy: $\pm 2.5\%$
Power (0 W to 9.9 W $\pm 5\%$ + 1W), 10 W to 500 W $\pm 5\%$	At 25 % duty cycle (10 seconds on, 30 seconds off): 10 Ω : 300 W, 20 Ω to 2900 Ω : 400 W, 3000 Ω to 5200 Ω : 200 W At 10 % duty cycle (5 seconds on, 45 seconds off): 10 Ω : 300 W, 20 Ω to 2400 Ω : 500 W, 2425 Ω to 2900 Ω : 400 W, 3000 Ω to 5200 Ω : 200 W
Current	RMS: 0 mA to 5,500 mA; Accuracy: $\pm (2.5\%$ of reading + 1 mA)
Voltage	Peak: 10 kV peak to peak; Accuracy: $\pm (10\%$ of reading + 50 V); Crest factor: 1.4 to 16.0 defined as the ratio of peak voltage to rms voltage (Vpk /Vrms), using the larger of the 2 peaks (positive or negative)
Vessel sealing measurement	Loop current, rms: 0 mA to 5500 mA; Accuracy: $\pm (2.5\%$ of reading + 1mA)
HF leakage current	Fixed load: 200 Ω V; V accuracy: $\pm 2.5\%$ Power rating: 400 W; Additional fixed load: 200 Ω ; Current, RMS: 0 mA to 5500 mA; Accuracy: $\pm (2.5\%$ of reading + 1 mA)
CQM test (contact quality monitor)	Resistances: 0 Ω to 475 Ω (by 1 Ω); Accuracy: 0 Ω to 10 Ω $\pm 0.5\%$, 11 Ω and above $\pm 5\%$; Power rating: 0.5 W; Auto time interval: 1 to 5 seconds
Oscilloscope output	1 V per ampere of input current, typical
Footswitch simulations	Cut and coag
Communications	USB device port: Micro B connector, full speed; Wireless port: 802.15, Speed: 115,200 baud
Memory	Test records: 5,000; Non-volatile: retained through power cycling
Safety	
IEC 61010-1	Overvoltage category none, pollution degree 2
IEC 61326-1	Basic
Wireless module listing	
FCC compliant (Class A)	FCC ID: X3ZBTMOD3
IC compliant	IC: 8828A-MOD3
CE (European) certified	CE0051
802.15.1 qualified	QD ID: B019224
Interface	

Push navigation buttons	Power on/off, test, select, back, and arrow keys
User Preferences	Adjust backlight, display brightness, and set clock
View verification history	
Recall and run templates on tester	General and personalized test templates allow for on-board test automation
Recall past saved and stored tests results	
Templates	Time duration, frequency of data capture and tests
Select user preferences	Units of measure, view and print test results on laptop
View battery life	Indicator bar shows life remaining
Service and Warranty	Free upgraded 24 months warranty
Manuals	MUST provide Technical and Operators manuals(Hard/Soft copies
Training	Onsite training for Biomedical Engineering Personnel on application and maintenance
CALIBRATION	<ul style="list-style-type: none"> ➤ Valid Certificate of calibration with test data ➤ Demonstrate Traceability to National / International Standards.
Key Features <ul style="list-style-type: none"> • Test ESU functions with precise power, current, frequency, crest factor, and load resistance ranges • Connect wirelessly via Bluetooth for easy record retrieval without interference or limitation by cables and wires* • All hardware and software necessary to complete preventive maintenance and troubleshooting is built in to the unit • User-friendly interface: large buttons and LCD screen guide the user through test sequences • Memory storage of up to 5,000 test records • Complies to global standards, including ANSI/ AAMI and IEC 	

ACCESSORIES

- Power Cord, country-specific 90-260v.
- Test Lead with Stackable Plugs.
- Suregrip Large Alligator Clip set.
- Test Lead set with Retractable Sheaths.
- ESU-Dispersive Safety Lead.
- ESU-CQM Safety Lead.
- ESU-Jumper Safety Lead.
- Users Manual & CD.
- Clamp (crocodile style), Grip C, black.

- Clamp (crocodile style), Grip C, red.
- Carrying Case.
- Calibration Manual.
- Ansur Test Software, QA-ES Plug-in License.
- Data Transfer Cable, RS232.
- 4635248 International dispersive lead (1/4 inch phono plug)
- 1909216 Test probe set—0.080 brass tip
- 4704312 Ansur Plug-in

ITEM 3: TECHNICAL SPECIFICATIONS FOR ELECTRICAL SAFETY ANALYZER

General specifications	
Display	LCD colour display
Communication	USB device upstream port Mini-B connector for control by a Computer
Mains Voltage (Range / Accuracy)	90V ac to 132V ac rms 180.0v ac to 264.0V ac rms \pm (2 % of reading + 1 V)
Mains Voltage Outlet	120 V ac : 230 V ac
Range (point-to-point voltage) Accuracy	0°0 V ac to 300 V ac rms \pm (2 % of reading + 2 LSD)
Voltage Tests	1Mains, Accessible, and Point to Point
Maximum Current	20A:16A
Hz	50 or 60 : 50 or 60
Storage Temperature	-20°C to 60°C
Operating Humidity	10 %-90 % RH (non-condensing)
Operating Altitude	<ul style="list-style-type: none"> ➤ To 2,000 meters @ 230 V ac (mains supply voltage). ➤ To 5,000 meters @ 120 V ac (mains supply voltage).
Earth Resistance	
Modes	Two terminal or four terminal
Test current	> 200 mA ac or 10 A ac to 25 A ac
Ranges Accuracy	0.000 Ω to 2.000 Ω \pm (2 % of reading 0.015 Ω)
Equipment current	
Patient load selection (input impedance)	AAMI ES1-1993 Fig 1 IEC 60601: Fig 15 IEC 61010: Fig A-1
Crest Factor	\leq 3
Ranges	0.0 μ A to 199.9 μ A 200 μ A to 1999 μ A 2.00 μ A to 10.00 mA
Frequency	DC to 1 kHz 1 kHz to 100 kHz 100 kHz to 1 MHz
Accuracy	\pm (1 % of reading + 1 μ A or 1 LSD, whichever is greater) \pm (2 % of reading + 1 μ A or 1 LSD, whichever is greater) \pm (5 % of reading + 1 μ A or 1 LSD, whichever is greater)
MAP Voltage	Additional residual leakage up to 4 μ A @ 120 V ac, and 8 μ A @ 240 V ac
Leakage Tests	<ul style="list-style-type: none"> ➤ Earth (ground wire). ➤ Chassis (enclosure). ➤ Patient (lead to ground). ➤ Patient Auxiliary (lead to lead). ➤ Mains on Applied Part (lead isolation). ➤ Direct Equipment. ➤ Direct Applied Part. ➤ Alternative Equipment.

	<ul style="list-style-type: none"> ➤ Alternative Applied Part. ➤ Accessible.
Differential Leakage	
Ranges	<ul style="list-style-type: none"> ➤ 50 μA to 199 μA. ➤ 200 μA to 1999 μA. ➤ 2 mA to 20 mA.
Accuracy	± 10 % of reading \pm (2 counts or 20 μ A, whichever is greater)
Insulation Resistance	
Ranges / Accuracy	<ul style="list-style-type: none"> ➤ 0.5 MΩ to 20 MΩ : \pm (2 % of reading + 2 counts). ➤ 20 MΩ to 100 MΩ : \pm (7.5 % of reading + 2 counts).
Source Test Voltage	<ul style="list-style-type: none"> ➤ 500 V dc ➤ 250 V dc
Insulation Resistance Tests	Mains-PE, AP-PE, Mains- PE, Mains-NE (non-earthed accessible conductive part) and AP- NE (non-earthed accessible conductive part)
ECG Performance Waveforms	
Waveforms / Rates	ECG Complex (BPM) : 30, 60, 120, 180, and 240
Accuracy	<ul style="list-style-type: none"> ➤ ± 2 %. ➤ ± 5 % for the amplitude of 2 Hz Square wave only, fixed @ 1 mV Lead II configuration.
Service and Warranty	Free upgraded 24 months warranty
Manuals	MUST provide Technical and Operators manuals(Hard/Soft copies
Training	Onsite training for Biomedical Engineering Personnel on application and maintenance
CALIBRATION	<ul style="list-style-type: none"> ➤ Valid Certificate of Calibration with Test Data ➤ Demonstrate Traceability to National / International Standards.

Key features

- Superior compliance with multiple standards: IEC60601- 1(partial), IEC62353, VDE 751, ANSI/AAMI ES1:1993, NFPA-99, AN/NZS 3551, IEC61010.
- Expanded leakage ranges through 10,000 μ A.
- Dual-lead resistance, leakage and voltage tests.
- AC only, dc only and true-rms leakage readings.
- 100 % and 110 % mains voltage for mains on applied parts (lead isolation) test.
- DSP filter technology.
- More applied parts selections.

- ECG and performance waveforms.
- Intuitive user interface.
- Easy-to-use applied parts (ECG) connections.
- Insulation posts on applied parts connections.
- Five different insulation tests.
- Varying insulation test voltage 500 V dc and 250 V dc.
- Two or four-wire ground wire resistance.
- Test- Automation Software plug-in
- USB connection.
- CE, C-TICK and CSA for USA and Canada certification.
- RoHS compliance.

Accessories

- CD-ROM Users Manual.
- Manual Multilingual Getting Started Guide.
- Accessory Kit (country specific).
- Soft Case Carry case Detachable Power Cord (country specific).
- Test- Automation Software plug-in.
- Cable Assembly Data Transfer Cable.
- Ground Adapter.
- Kelvin Test Lead Set for 4 Wire ground.
- Ultrasound Test Cable Adapter.
- Universal banana Adapter.
- NPA – null Post Assembly.
- 50' Test Lead.
- 75' Test Lead.

**ITEM 4: TECHNICAL SPECIFICATIONS FOR HUMAN BLACK BODY SOURCE (Precision
Infrared Calibrator)**

Power	
Power Adapter – Universal voltage	Input: 240V ac ($\pm 10\%$), 3.15A 50/60Hz, 630w
Backup power	Rechargeable battery internal 24 hour battery life with 30 second sample rate
Measurements and tests specifications	
Temperature range	At least 5 sensors Range: $-15\text{ }^{\circ}\text{C}$ to $120\text{ }^{\circ}\text{C}$ Accuracy: $\leq \pm 0.40\text{ }^{\circ}\text{C}$ at $-15\text{ }^{\circ}\text{C}$ $\pm 0.40\text{ }^{\circ}\text{C}$ at $0\text{ }^{\circ}\text{C}$ $\pm 0.50\text{ }^{\circ}\text{C}$ at $50\text{ }^{\circ}\text{C}$ $\pm 0.50\text{ }^{\circ}\text{C}$ at $100\text{ }^{\circ}\text{C}$ $\pm 0.55\text{ }^{\circ}\text{C}$ at $120\text{ }^{\circ}\text{C}$
Stability	$\pm 0.10\text{ }^{\circ}\text{C}$ at $-15\text{ }^{\circ}\text{C}$ $\pm 0.05\text{ }^{\circ}\text{C}$ at $0\text{ }^{\circ}\text{C}$ $\pm 0.10\text{ }^{\circ}\text{C}$ at $120\text{ }^{\circ}\text{C}$
Uniformity (5.0 in dia of centre of target)²	$\pm 0.15\text{ }^{\circ}\text{C}$ at $-15\text{ }^{\circ}\text{C}$ $\pm 0.10\text{ }^{\circ}\text{C}$ at $0\text{ }^{\circ}\text{C}$ $\pm 0.25\text{ }^{\circ}\text{C}$ at $120\text{ }^{\circ}\text{C}$
Uniformity (2.0 in dia of centre of target)²	$\pm 0.10\text{ }^{\circ}\text{C}$ at $-15\text{ }^{\circ}\text{C}$ $\pm 0.10\text{ }^{\circ}\text{C}$ at $0\text{ }^{\circ}\text{C}$ $\pm 0.20\text{ }^{\circ}\text{C}$ at $120\text{ }^{\circ}\text{C}$
Heating time	15 min: $-15\text{ }^{\circ}\text{C}$ to $120\text{ }^{\circ}\text{C}$ 14 min: $23\text{ }^{\circ}\text{C}$ to $120\text{ }^{\circ}\text{C}$
Cooling Time	15 min: $120\text{ }^{\circ}\text{C}$ to $23\text{ }^{\circ}\text{C}$ 20 min: $23\text{ }^{\circ}\text{C}$ to $-15\text{ }^{\circ}\text{C}$
Stabilization time	10 minutes
Nominal Emissivity	0.95
Thermometer emissivity compensation	0.9 to 1.0
Target diameter	152.4 mm (6 in)
Computer Interface	RS- 232
Safety	
IEC 61010-1	EN 61010-1:2001, CAN/CSA C22.2 No. 61010.1-04
Service and Warranty	Free upgraded 24 months warranty
Manuals	MUST provide Technical and Operators manuals(Hard/Soft copies)
Training	Onsite training for Biomedical Engineering Personnel on Application and Maintenance
CALIBRATION	<ul style="list-style-type: none"> ➤ Valid Certificate of calibration with test data ➤ Demonstrate Traceability to National / International Standards.
Key Features	

• Complies with all global standards, including IEC 60601-2-19, IEC 60601-2-20 and IEC 60601-2-21

ACCESSORIES

- 4180 IR Calibrator
- Report of Calibration (NVLAP Accredited)
- Power Cord
- User's Guide
- Documentation CD
- Target Cover
- Serial Cable
- Aperture, 50 mm (2 in)
- Case, Carrying,
- Carrying, w/Wheels

ITEM 5: TECHNICAL SPECIFICATION FOR BIOMETER

A highly versatile gauge used for measuring pressure in hydraulic and pneumatic sources.

- **Pressure Range:** 0 to 3000 psi (0 to 20.68 MPa)
- **Accuracy:** $\pm 0.05\%$ of full scale
- **Display:** 3.5-digit LCD with backlight
- **Resolution:** 0.1 psi (0.001 MPa)
- **Units of Measure:** psi, bar, kg/cm², kPa, MPa
- **Flow Measurement:** Can measure 10 to 250 liters per minute (lpm) with an optional parabolic flow adapter (accuracy: $\pm 5\%$)
- **Safety & Environment:** CAT III 300V, operating temp from 0°C to 50°C
- Three-year warranty
- **Onsite training** for Biomedical Engineering Personnel on application and maintenance
- Valid Certificate of Calibration with Test Data

**ITEM 6: TECHNICAL SPECIFICATIONS FOR BIOSAFETY CABINET TESTING /
CALIBRATION STANDARD TOOL**

SPECIFICATIONS	
Measurements and tests specifications	
Hotwire Anemometer Air Velocity Meter	<ul style="list-style-type: none"> ➤ Extended air velocity measurement range of 0 to 6000 ft/min. ➤ Articulating Probe ➤ Data logger ➤ Calculates volumetric flow, and actual/standard velocity. ➤ Valid Calibration certificate. ➤ Simultaneously measure temperature and velocity ➤ Probe holder stand with clamp ➤ Hard Carrying case ➤ USB Cable and downloading software
Air-Data Multimeter	
Air-Data Multimeter with 8400-flowhood kit	<p>AIR VELOCITY: ± 3% of reading ± 7 fpm from 50 to 8000 fpm pitot tube (29,000 fpm FS); 50 to 5000 fpm AirFoil; 50 to 2500 fpm VelGrid</p> <p>DIFFERENTIAL PRESSURE: ± 2% of reading ± 0.001 in wc from 0.0500 to 50.00 in wc, (0.0001 to 60 in wc FS); 20 psid safe pressure</p> <p>TEMPERATURE: ± 0.5°F accuracy from 32°F to 158°F using ADT440 Series TemProbes (-67°F to 250°F FS); 0.1°F resolution.</p> <p>AIRFLOW: Accuracy is ± 3% of reading ± 7 cfm from 100 to 2000 cfm; range is 25 to 2500 supply, 25 to 1500 exhaust with 8400 FlowHood.</p> <p>ABSOLUTE PRESSURE: ± 2% of reading ± 0.1 in Hg from 14 to 40 in Hg referenced to vacuum. 60 psia maximum safe pressure.</p> <p>OPERATIONAL TEMPERATURE LIMITS: 40°F to 140°F.</p> <p>AIR DENSITY CORRECTION: Local or standard (mass flow) air density correction range is 14 to 40 in Hg and 32°F to 158°F</p> <p>POSITION SENSITIVITY: Unaffected by position.</p> <p>MEMORY: 200 readings, sequence labelled, sum and average.</p> <p>CALIBRATION: Calibration certified NIST traceable</p> <p>READOUT: 10 digit, 0.4", high contrast, liquid crystal display. METER HOUSING: 6.0" x 6.4" x 2.7" high impact ABS. 36 oz. CONNECTIONS: 1/4" OD slip-on for 3/16" ID soft tubing. BATTERY LIFE: Up to 3000 readings/charge, 500 recharge cycles</p>
Digital Aerosol Photometer	
Digital Aerosol Photometer for HEPA filter integrity test	<ul style="list-style-type: none"> ➤ Power: 100 to 250 Volts AC, 50/60Hz automatic adjustment. ➤ Portable ➤ Auto Zero: Automatically establishes zero reading at startup. ➤ With an iprobe, handle scanner with local LCD and controls long life laser light source; ➤ High precision photoelectric multiplier tube detection; ➤ Can set PAO, DOP and other types of aerosols;

- Equipped with a dedicated handheld device to achieve control, display and sampling function;
- Large capacity data storage, real-time data storage;
- more than the setting alarm sound and light alarm;
- Can be obtained by disk or thermal printer to print historical data;
- Can print leakage rate and other monitoring data in real time;
- Through the special software, the sampling data can be imported into PC: Lap top Computer minimum requirements:
 - Processor & Core Logic: 11th Generation Intel® Core™ i7-1255U Processor
 - The laptop shall be 2 in 1.
 - System Memory: 16GB LPD4X-4266 RAM
 - Storage: 512GB SSD M.2 2280 PCIe 3.0x4 NVMe
 - Graphics: Integrated Intel Iris Xe Graphics
 - Display: 14-inch WUXGA (1920x1200) IPS 300nits Glossy, Glass, Touch
 - Audio: High Definition (HD) Audio, Realtek ALC3287 codec
 - Speakers: Front-facing stereo speakers, 2W x2, optimized with Dolby Audio
 - Wireless: Wi-Fi 6, 11ax 2x2 + Bluetooth 5.1
 - Battery: Integrated 52.5Wh
 - Security Chip: Firmware TPM 2.0
 - Fingerprint Reader: Touch Style, on Palm Rest
 - Power: 65W Round Tip (3-pin)
 - Microphone: 2x, Array
 - Operating System: Genuine Windows® 10 Professional 64-bit
 - Software: Latest Version, MS Office 2019 licensed
- Latest reputable Version of anti-virus
- Accessories: Executive carry case, Ethernet USB 3.0 Gigabit Lan & a wireless mouse
- Automatic fault detection.
- Dynamic Range-up to 600 micrograms per liter
- Reagent Settings-PAO-4, DOP, Ondina, Mineral Oil, Paraffin
- Sensitivity- 1% of readings $\geq 0.01\%$ to 100%
- Repeatability-0.5% of readings $\geq 0.01\%$ to 100%.
- Flow Control- Pulse Width Modulation (PWM) control ensures sampling rate of 1cfm (28.3lpm) +/- 10% in all applications.
- Light Source- Solid State, rated for life of instrument
- Alarm-User selectable audible, visual, and vibratory alarms notify the user when the user defined set point is exceeded.
- Power- 240VAC, 50/60HZ

	<ul style="list-style-type: none"> ➤ Standards compliance- NSF 49, IEST, ISO-14644
Aerosol Generator	
Aerosol Generator for air flow pattern test	<ul style="list-style-type: none"> ➤ Air flow range:50- 2000 cfm. ➤ Nozzle selection L(low) / H(high). ➤ ~240V , 50/60HZ. ➤ Aerosol Reagent: PAO-4, DOS, Ondina EL, DOP, Mineral oil, Paraffin or Corn Oil liquid aerosol reagent. (20Litres) ➤ Ambient Pressure: Standard Atmosphere ➤ Ambient Temperature: 5~40°C
Safety	
Service and Warranty	Free upgraded 24 months warranty
Manuals	<p>MUST provide Technical and Operators manuals(Hard/Soft copies)</p> <p>Attach the relevant brochures.</p>
Training	Onsite training for Biomedical Engineering Personnel on application and maintenance
CALIBRATION	<ul style="list-style-type: none"> ➤ Valid Calibration Certificate with test data ➤ Demonstrate Traceability to National / International Standards.
<p>Biosafety Cabinet Certification Program Training for All Classes of BSC-For Two (2) Biomedical Engineering personnel. (To include Maintenance that is Trouble Shooting, Repair, Service and Certification)</p>	
Overview	<ul style="list-style-type: none"> ➤ How BSCs are constructed and function. ➤ BSC types and exhaust requirements. ➤ How to use BSCs effectively. ➤ How to follow NSF 49 and EN 12469 test procedures for BSC certification. ➤ Developing a testing grid for BSCs. ➤ Performing inflow and down flow airflow tests. ➤ Evaluating BSC airflow using smoke pattern tests. ➤ Conducting a site assessment. ➤ Performing HEPA leak testing. ➤ Determining when and how to decontaminate a BSC. ➤ Troubleshooting BSCs. ➤ Balancing BSC airflows ➤ Repairing leaks in HEPA filters. ➤ Changing HEPA filters. ➤ Teaching others to use BSCs effectively.

- | | |
|--|---|
| | <ul style="list-style-type: none">➤ Clearly indicate Training institution, Duration and Detailed Training program |
|--|---|

Key Features

- Complies with all global standards, including IEC 60601-2-19, IEC 60601-2-20 and IEC 60601-2-21

ACCESSORIES

- Probe holder stand with clamp
- Hard Carrying case
- USB Cable and downloading software

ITEM 7: TECHNICAL SPECIFICATIONS FOR FETAL SIMULATOR

General Specifications	
Power:	9V battery/battery eliminator; low battery indication set at 6 V
Environmental	Operating conditions 10-40°C (50 -104°F) 0-90% RH -NC Storage environment -15 -60°C (5 -140°F) 0-90% RH -NC Environmental protection IP40 Impact Rating IK08
Display:	2-line x 16-character LCD with keypad
Housing:	Plastic case
RS232:	Bidirectional interface, 9600 Baud
Electrical Interfaces	ECG (& respiration) via 10 x 4 mm sockets BP 1 -4: 6 pin mini DIN TEMP 8 pin mini DIN USB Port micro
Fetal ECG	
Static Rates:	30, 60, 90, 120, 150, 180, 210, and 240 BPM
ECG Sensitivity:	50 µV, 100 µV, 200 µV, 0.5 mV, 1 mV, and 2 mV US-1 tracks primary fetal ECG rates. US-2 tracks secondary fetal activity for either independent "normal or "twins simulation, US-2 rate fixed at 140 BPM
Fetal Patterns:	Note: US-1 fetal ECG track these selections. US-2 is in normal pattern, except during TREND #1 selection.
Variability selections (added to fetal ECG)	Absent variability, low variability, mild variability, high variability, severe variability, and long-term variability
Optional mechanical fetal heart	Provides a mechanical interface to the ultrasound transducer; can be connected to either ultrasound channels. This option, due to its power consumption, requires an ac power adapter to be connected.
Twin Fetal Patterns	
NORMAL:	Normal pattern
TACHYCARDIA:	Tachycardia pattern
BRADYCARDIA:	Bradycardia pattern
ARRHYTHMIAS:	Arrhythmia pattern
LATE DECELERATION:	Late deceleration
EARLY DECELERATION:	Early deceleration
MODERATE DECELER:	Moderate variable deceleration
ACCELERATION #1:	Acceleration wave #1
ACCELERATION #2:	Acceleration wave #2
SINUSOIDAL (HIGH):	Sinusoidal pattern, large change
SINUSOIDAL (LOW):	Sinusoidal pattern, small change
SEVERE VAR. DEC. #1:	Severe deceleration wave #2
SEVERE VAR. DEC. #2:	Severe variable deceleration

PROLONGED DECELERATION:	Prolonged deceleration
BIPHASIC DECELERATION:	Biphasic deceleration
EXAGGERATED DECELERATION:	Exaggerated deceleration
NON-UNIFORM DECELERATION:	Non-uniform deceleration
VAR. DECELERATION (U):	Variable deceleration "V" shaped
VAR. DECELERATION TACH	Variable deceleration with high-rate BPM
VAR. DECELERATION (V)	Variable deceleration "V" shaped
VAR. DECELERATION (POST):	Variable deceleration post exaggerated
VAR. DECELERATION DECELERATION (POSITION):	Variable deceleration with position changes
LONG DECELERATION:	Long deceleration
COMPENSATORY ACCELERATION:	Compensatory acceleration
Maternal ECG	
ECG Static Rates:	60, 80, 100, 120, 140, 160 BPM
ECG Sensitivity:	0.5 mV, 1 mV, and 2 mV
Uterine Activity	
EXECUTE WAVEFORM:	Start toco waveform
UTERINE WAVE OFF:	Stop toco waveform
ANALOG 0 TO 1 VOLT:	Analog range 0 to 1 V (1 V = 100 mmHg)
ANALOG 0 TO -1 VOLT:	Analog range 0 to -1 V (1 V = 100 mmHg)
UTERINE WAVE 0 TO 25:	Range of toco waveform
UTERINE WAVE 0 TO 50:	Range of toco waveform
UTERINE WAVE 0 TO 100:	Range of toco waveform
SHORT DURATION:	Toco waveform of short duration
NORMAL DURATION:	Normal duration of toco waveform
INCREASED DURATION:	Long duration of toco waveform
UTERINE LEVEL = ZERO:	Zero toco channel (automatic on power up)
UTERINE STATIC + 20:	Increase toco static level by 20 mmHg (0 to 100 mmHg)
INCR. RESTING TONE:	Resting tone increases
COUPLING:	2 close toco waves
TRIPLING:	3 close toco waves
UTERINE PRESSURE SENSITIVITY:	5 or 40 μ V on power up
Service and Warranty	Free upgraded 24 months warranty
Manuals	MUST provide Technical and Operators manuals (Hard/Soft copies)
Training	Onsite training for Biomedical Engineering Personnel on application and maintenance
CALIBRATION	➤ Valid Certificate of calibration with test data

- | | |
|--|---|
| | ➤ Demonstrate Traceability to National / International Standards. |
|--|---|

Key Features

- Mechanical heart for ultrasound simulation
- TOCO simulation (External or IUP)
- Ultrasound simulation (including twins)
- Maternal ECG simulation
- Fetal ECG (tracks ultrasound #1)
- Internal (DECG) and external fetal ECG
- Uterine-activity selections
- Fetal beat-to-beat variability
- Periodic and non-periodic fetal ECG changes
- Arrhythmia selections
- Compact, lightweight, and pocket-size plastic housing
- Battery operated with status indications
- Avail special kits with all required accessories and cables to test fetal monitors for specified manufacturers

8. Technical Specification for Current Meter

- **AC Current Range:** 0 to 400.0 A
- **DC Current Range:** 0 to 400.0 A
- **Voltage (AC/DC):** up to 600.0 V
- **Accuracy:** 2.0% \pm 5 digits for AC current
- **Resistance:** up to 40.00 k Ω
- • **Additional:** Temperature measurement (-10.0°C to 400.0°C), capacitance, and frequency
- Three-year warranty
- **Onsite training** for Biomedical Engineering Personnel on application and maintenance
- Valid Certificate of Calibration with Test Data

ITEM 9: TECHNICAL SPECIFICATIONS FOR CONTACT AND NON CONTACT DUAL PURPOSE TACHOMETER

General specifications	
Display	LCD colour display
Communication	USB device upstream port Mini-B connector for control by a Computer
Battery life	(2) 1.5V LR6 (alkaline) AA battery 40 hours (minimum)
Operating Temperature	0°C to 50°C
Storage Temperature	-20°C to 70°C
Safety standards	IEC/EN 61010-1 3rd Edition; Pollution degree 2
Operating Humidity	10 %-90 % RH (non-condensing)
Operating Altitude	2000 M
Certifications	CE, CSA, C-TICK N10140 , RoHS
Electromagnetic compatibility (EMC)	IEC 61326-1:2006
Rotation Speed Measurement	
Range	Optical (non-contact) measurement: 1 to 99999 revolutions/minute Contact measurement: 1 to 19999 revolutions/minute
Precision	±0.02% RD+1
Sensing Distance	(optical) 500 mm
Service and Warranty	Free upgraded 24 months warranty
Manuals	MUST provide Technical and Operators manuals(Hard/Soft copies
Training	Onsite training for Biomedical Engineering Personnel on application and maintenance
CALIBRATION	<ul style="list-style-type: none"> ➤ Valid Certificate of calibration with test data ➤ Demonstrate Traceability to National / International Standards.

Key features

- 6-digit backlight display reads the data clearly.
- Signal detection reminder avoids invalid detection actions.
- Maximum/minimum/average value and final readings.
- Setting/Turning off the automatic power-off function.
- Small size and easy to hold.
- Easy switching of the contact and non-contact measurement modes.
- Different speed measuring heads and rotational speeds adapt to different measurement occasions.

- 2-year warranty

Accessories

- 0.1 m Contact wheel.
- 6 inch Contact wheel.
- Mechanical Adapter
- Shaft Contacts
- Shaft Extension
- Conical speed measuring head.
- Inner cone speed measuring head.
- Lengthening rod.
- 10 reflective stripes

ITEM 10: TECHNICAL SPECIFICATIONS FOR OXYGEN ANALYZER

General specifications	
Display	LCD display
Battery life	<ul style="list-style-type: none"> ➤ 5000 hrs with continuous use. ➤ Low Batt indication BAT icon displayed on LCD ➤ 2 AA Alkaline batteries.
Operating Temperature	15°C to 40°C
Storage Temperature	-20°C to 50°C
Safety standards	IEC/EN 61010-1 3rd Edition; Pollution degree 2
Operating Humidity	10 %-95 % RH (non-condensing)
Operating Altitude	2000 M
Measurement	
Measurement Range	0% to 100%
Resolution	0.1%
Accuracy and Linearity	1 % of full scale at constant temperature, RH and pressure when calibrated at full scale
Total Accuracy	± 3 % actual oxygen level overfull operating range
Response Time	90 % of final value in approx. 15 sec at 23° C
Oxygen Sensor	Galvanic fuel sensor 2 years in typical application
Service and Warranty	
	Free upgraded 24 months warranty

Manuals	MUST provide Technical and Operators manuals(Hard/Soft copies)
Training	Onsite training for Biomedical Engineering Personnel on application and maintenance
CALIBRATION	<ul style="list-style-type: none"> ➤ Valid Certificate of calibration with test data ➤ Demonstrate Traceability to National / International Standards.

Key features

- One- touch calibration with reminder.
- Long battery life (Approx 5000 hrs).
- Impact resistant and drip-proof
- 2-year warranty

Accessories

- External oxygen sensor MAX 250ESF
- Lanyard
- Carry case
- Threaded Adapter with Tygon Tubing
- Flow Diverter
- Blue Tee Adapter
- Perfusion Tee Adapter
- Long-Neck Tee Adapter
- Paediatric Tee Adapter
- MAX-Quick Connect
- Dovetail Bracket

ITEM 11: TECHNICAL SPECIFICATION FOR THERMO HYGROMETER

- **Sensors:** Detachable, interchangeable standard (“S”) or high-accuracy (“H”) sensors.
- **Temperature Accuracy (H-Model):** $\pm 0.125^{\circ}$ C (from 16° C to 24° C)
- **Temperature Accuracy (S-Model):** $\pm 0.25^{\circ}$ C (from 15° C to 35° C)
- **Humidity Accuracy:** $\pm 1.5\%$ RH (H-model) and $\pm 2.0\%$ RH (S-model) from 20% to 70% RH
- **Memory & Features:** Stores up to 400,000 date-stamped readings. Calculates dew point, heat index, and rate of change.
- **Connectivity:** Ethernet, RS-232, and wireless capabilities.
- Three-year warranty
- **Onsite training** for Biomedical Engineering Personnel on application and maintenance
- Valid Certificate of Calibration with Test Data

ITEM 12: TECHNICAL SPECIFICATION FOR HYDRAULIC TEST PRESSURE KIT

- **Pressure Range:** 0 to 700 bar [10, 000 PSI / 70 MPa]
- **Testing Medium:** Distilled water, mineral-based hydraulic oil, or transformer oil
- **Control Modes:** Manual (lever/hand pump) or Semi-Automated (pneumatic hydraulic/computer-controlled)
- **Measurement Accuracy:** Typically $\pm 0.5\%$ to $\pm 1\%$ of full scale
- **Drive/Power Supply:** Hand-operated (two-stage), pneumatic booster (drive air 3 - 8 bar), or electric-driven
- **Connections:** G 1/2" or 1/4" NPT female/male quick-release connections for Reference gauges and Devices Under Test (DUT)
- **Fine Adjustment:** Micro-metric spindle or fine adjustment valve.
- Three-year warranty
- **Onsite training** for Biomedical Engineering Personnel on application and maintenance
- Valid Certificate of Calibration with Test Data

ITEM 13: TECHNICAL SPECIFICATION FOR THERMO HYGROMETER

- **Accuracy:** ± 0.05 % of Full Scale
- **Pressure Ranges:** From ± 15 psi up to 10,000 psi (100 kPa to 70 MPa)
- **Media Compatibility:** Available for gases and liquids compatible with 316 stainless steel
- **Display:** $(5 \frac{1}{2})$ digits, 16.53 mm high with a 20-segment bar graph (0 to 100 %)
- **Sampling Rate:** User-configurable to 0.5, 1, 3, or 10 samples per second
- **Power & Battery Life:** Three AA alkaline batteries yield up to 1,500 hours of continuous use
- Combine with the 700PTPK or 700HTPK pump kits for a complete pressure testing solution for up to 600 psi (40 bar) with the PTP-1 pneumatic pump and up to 10,000 psi (690 bar) with the HTP-2 hydraulic pump
- Includes 700HTH-1 hose kit and adapters for connecting to the pressure device to be tested
- Log up to 8,493 pressure measurements to memory (requires 700G/TRACK software)
- Three-year warranty
- **Onsite training** for Biomedical Engineering Personnel on application and maintenance
- Valid Certificate of Calibration with Test Data
- Demonstrate Traceability to National / International Standards.

ITEM 14: TECHNICAL SPECIFICATIONS FOR ECG SIMULATOR

General Specifications	
Battery/Main Power Supply	3.7V 3900mAh 14.4WH Li-Ion battery 5V 1A USB micro-B power supply 100-240V ~ 50/60Hz 0.18A max.
Environmental	Operating conditions 10-40°C (50 -104°F) 0-90% RH -NC Storage environment -15 -60°C (5 -140°F) 0-90% RH -NC Environmental protection IP40 Impact Rating IK08
Electrical Interfaces	ECG (& respiration) via 10 x 4 mm sockets BP 1 -4: 6 pin mini DIN TEMP 8 pin mini DIN USB Port micro
Simulation Functions	
General ECG Simulation	Full 12 lead ECG with independent outputs for each signal lead
ECG Normal Sinus Rhythm	Heart rate 18 selectable values 30-300 BPM

	<p>Accuracy ± 1BPM or 1%</p> <p>Amplitude (lead II) 0.05 mV to 0.45 mV (0.05 mV steps) 0.5 mV to 5.5 mV (0.5 mV steps)</p> <p>Amplitude Accuracy $\pm 2\%$</p> <p>ST Segments 18 selectable values, 8 elevated & 8 depressed</p> <p>Neonatal Mode ECG R wave width is reduced to 40ms</p>
Arrhythmia Waveforms (Atrial)	
Amplitude (lead II)	<p>0.05 mV to 0.45 mV (0.05 mV steps)</p> <p>0.5 mV to 5.5 mV (0.5 mV steps)</p>
Amplitude Accuracy	$\pm 2\%$
Ventricular Waveforms	<p>A systole</p> <p>Bigeminy</p> <p>Trigeminy</p> <p>Ventricular Fibrillation (coarse)</p> <p>Ventricular Fibrillation (fine)</p> <p>Ventricular Tachycardia</p>
Supraventricular Waveforms	<p>Atrial Fibrillation coarse</p> <p>Atrial Fibrillation fine</p> <p>Atrial Flutter</p> <p>Atrial Tachycardia</p> <p>Missing beat</p> <p>Nodal rhythm</p> <p>Paroxysmal Atrial Tachycardia</p> <p>Sinus Arrhythmia</p> <p>Supraventricular Tachycardia</p>
Atrial Conduction Waveforms	<p>First Degree AV Block</p> <p>Left Bundle Branch Block</p> <p>Right Bundle Branch Block</p> <p>Second Degree AV Block -Mobitz I</p> <p>Second Degree AV Block -Mobitz II</p> <p>Third Degree AV Block</p>
Premature Waveforms	<p>Premature Atrial Contraction</p> <p>Premature Nodal Contraction</p> <p>Premature Left Ventricle Contraction</p> <p>Premature Left Ventricle Contraction -early</p> <p>Premature Left Ventricle Contraction -R on T</p> <p>Premature Right Ventricle Contraction</p> <p>Premature Right Ventricle Contraction -early</p> <p>Premature Right Ventricle Contraction -R on T</p> <p>Premature Ventricular Contraction -6 / min</p> <p>Premature Ventricular Contraction -12 / min</p> <p>Premature Ventricular Contraction -24 / min</p> <p>Premature Ventricular Contraction -frequent multifocal</p>

Performance Waveforms	
Square Waves	2 Hz, 0.125 Hz
Triangle Wave	2 Hz
Pulse	60bpm or 240bpm
Sine Waves	0.05, 0.5, 1, 10, 25, 30, 40, 50, 60, and 100 Hz
R-Wave Detector Test	60 BPM haver-triangle wave with selectable width and amplitude
Haver-triangle Width	12 selectable values between 8 and 200 msec
Performance amplitude	0.5 to 5.0 mV in 0.5 mV steps
Pacer Waveforms	
Simulated Rhythms	Asynchronous at 75 BPM Demand with frequent sinus beat Demand with occasional sinus beat A-V sequential Non-capture Non-function
Pulse amplitude	1.0, 2.0, 5.0, 10.0 mV
Accuracy	±10%
Width	5 selectable values 0.1-2.0 ms.
Accuracy	± 5%
R Wave Detection	
Heart Rate	6 selectable values 30-250 BPM
Amplitude	0.05 mV to 0.50mV (0.05 mV steps)
R wave width	13 selectable values 8-200ms
Respiration Simulation	
Rates	0, 5, 10, 15, 30, 40, 60, 80, 120, 180 BRPM
Resistance Variations	0.2, 0.5, 1.0, 3.0 Ω
Accuracy	±10%
Base resistances	500, 1000, 1500 and 2000 Ω
Accuracy	±5%
Lead selection	1 (LA), 2(LL) user selectable
Apnoea Simulation	Manual on/off
Temperature Simulation	
Simulation	2 channels
Static Pressure	-10,-5,0,20,40,50,60,80,100,150,160,200,240, 320, 400mmHg

Dynamic Simulation	Arterial [ART] 120/80 Radial Artery [RA] 120/80 Left Ventricle [LV] 120/00 Right Ventricle [RV] 25/00 Right Atrium (central venous) [CVP] 15/10 Pulmonary Artery [PA] 25/10 Pulmonary Artery Wedge [PAW] 10/2 Left Atrium [LA] 14/4
Auto sequence (C1 only)	Cycle through simulations with 15 second step duration: Arterial [A RT] 120/80 Radial Artery [RA] 120/80 Left Ventricle [LV] 120/00 Right Ventricle [RV] 25/00 Pulmonary Artery Wedge [PAW] 10/2 Pulmonary Artery [PA] 25/10 Right Atrium (central venous) [CVP] 15/10
Accuracy	± 1mmHg
Excitation voltage	2V to 16V
Impedance	350Ω Nominal
Simulated sensitivity	5μV/V/mmHg or 40μV/V/mmHg (user selectable)
Connector	mini DIN style
Service and Warranty	Free upgraded 24 months warranty
Manuals	MUST provide Technical and Operators manuals (Hard/Soft copies)
Training	Onsite training for Biomedical Engineering Personnel on application and maintenance
CALIBRATION	<ul style="list-style-type: none"> ➤ Certificate of calibration with test data ➤ Demonstrate Traceability to National / International Standards.

Key Features

- Large color display with intuitive navigation.
- Store and recall the 5 most used sequences with the touch of a button.
- Dedicated home navigation button.
- Rechargeable Li-Ion battery with up to 6 hours of operations between charges.

- High-contrast and low luminance color display for minimal disruption in critical or intensive care environments.
- Universal ECG studs to connect with the widest possible range of cables.
- Compatible with legacy temperature and IBP interconnect cables.
- Onscreen waveform display for quick reference.
- Robust carry case to protect your simulator. International localization options available to meet requirements across the globe.

Simulation Functions in summary

- 12 lead NSR ECG simulation.
- Over 40 selectable Arrhythmias.
- Performance waveforms.
- Respiration simulation on lead I & II.
- Universal temperature simulation.
- Two channel IBP simulation.
- Fetal simulation.

ITEM 15: TECHNICAL SPECIFICATIONS FOR INFUSION DEVICE ANALYZER

Flow and Volume	
Mean flow range	0 - 300mL/h (Steady flow),0 - 1500 ml/h (Peristaltic flow)
Mean flow resolution	0.001 mL.h
Mean flow accuracy (after > 0.1 mL and > 10s)	$\pm (1 \% + 0.005 \text{ mL/h})$ at < 500 mL/h, otherwise $\pm 2 \%$
Volume range	0 – 100 000 mL
Volume resolution	0.001 mL
Volume accuracy (after > 0.1 mL and > 10s)	$\pm (1 \% + 0.003 \text{ mL})$ at < 500 mL/h, otherwise $\pm 2 \%$
Time range	0 s – 1000 h
Time accuracy	$\pm (0.2 \% + 0.2 \text{ s})$
Time trig volume	$\leq 3 \mu\text{L}$
Graphs	Instant flow and mean flow with up to 1 s resolution (gradually reduced after 1 hour), IEC 60601-2-24 trumpet curve
Occlusion	
Peak pressure range	-200 – +2600 mmHg (-3.8 – +50 psi)
Peak pressure resolution	1 mmHg (0.01 psi)
Peak pressure accuracy	$\pm (1 \% + 5 \text{ mmHg (0.1 psi)})$, 50 ms moving average
Time to peak range	0 s – 1000 h
Time to peak accuracy	$\pm (0.2 \% + 0.05 \text{ s})$
Time to alarm	0 s – 1000 h
Time to alarm accuracy	Operator dependent, manual stop button
Residual volume accuracy	$\pm (2 \% + 0.01 \text{ mL})$
Graphs	Pressure with up to 50 ms resolution (gradually reduced after 3 minutes)
PCA bolus/dual flow measurement	
Bolus volume range	0.01 – 100 000 mL
Bolus volume accuracy (after > 1 s)	$\pm (2 \% + 0.01 \text{ mL})$
Bolus flow range	$10 \times \text{base flow} - 3000 \text{ mL/h}$ (steady flow), $10 \times \text{base flow} - 1500 \text{ mL/h}$ (peristaltic flow)
Bolus flow accuracy (after > 0.1 mL and > 10 s)	$\pm (1 \% + 0.005 \text{ mL/h})$ at < 500 mL/h, otherwise $\pm 2 \%$
Base flow range	1 – 100 mL/h
Base flow accuracy (after > 0.1 mL)	$\pm (1 \% + 0.005 \text{ mL/h})$
Bolus duration range	0 s – 1000 h
Bolus duration accuracy	$\pm (0.2 \% + 0.2 \text{ s})$
Graphs	Instant flow with up to 1 s resolution (gradually reduced after 1 hour), numbered boluses
Pressure measurement	
Range	≤ -500 to $\geq 2500 \text{ mmHg}$
Back pressure setting range	≤ -200 to $\geq 600 \text{ mmHg}$
Accuracy	-500 to 1000 mmHg $\pm 10 \text{ mmHg}$ 1000 to 2500 $\pm 1\%$ of reading

Measurement Units	
Pressure	mmHg, PSI, Bar, Kpa
Volume	mL, g (1 mL = 0.998 g)
General Specifications	
Operating voltage range	90 V ac to 264 V ac
Supply frequency	50/60 Hz
Supply power	Approx 60watts
Altitude	0 m to 3000 m (Nairobi 1680m)
Temperature	
Operating	15 °C to 40 °C (59 °F to 104 °F)
Storage	-20 °C to +70 °C , when drained
Display	At least 10-inch multi touch, 800 * 600 pixels
Data Storage	10000 Measurements
Channels	1 to 4, Depending on Configuration
Connectivity	1 × USB-C for PC communication, 3 × USB-A for peripherals
Safety Standard	Complies with IEC 61010-1:2010, pollution degree 2
Infusion Device Standard	For testing according to IEC 60601-2-24:1998
EMC Standard	Complies with IEC 61326-1:2012
Calibration	Device fully calibrated and traceability demonstrated
Service and Warranty	Free upgraded 24 months warranty
Manuals	MUST provide Technical and Operators manuals(Hard/Soft copies) in English
Training	Onsite training for Biomedical Engineering Personnel on Application and Maintenance
CALIBRATION	<ul style="list-style-type: none"> ➤ Valid Certificate of calibration with test data ➤ Demonstrate Traceability to National / International Standards.
Key features	
<ul style="list-style-type: none"> ➤ Installed with integrated OneQA-enabled workflow automation software which simplifies testing, enhancing accuracy and efficiency ➤ Achieve low-flow stable readings 100x faster than IDA-5 ➤ IntelliPump Technology offers no mess testing with a water recirculation system ➤ No priming required between tests or repriming after initial setup ➤ Test up to four infusion pumps simultaneously ➤ Modular design allows for flexibility; users may add additional channels as needed ➤ Removable channels simplify the calibration process and eliminate downtime ➤ Fully compliant with IEC60601-2-24 standard ➤ Meets or exceeds most OEM low-flow accuracy ➤ Enhanced resolution by 60x vs. IDA-5 ➤ Flow and volume are 100x more accurate vs. IDA-5 ➤ Simply and quickly customize pre-built test procedures for any infusion device 	

- Wireless capability allows you to run test procedures in the field and automatically sync results
- Large 10” touchscreen with an intuitive user interface
- Ability to upload pictures, instructions, and text to procedures
- Global support network to support customers anywhere, anytime
- Immediate results with as little as 0.003mL
- Verifies PCA, bolus, and occlusion functionality

ACCESSORIES

- Syringes
- 5-ft Plastic Drain Line
- USB-C Data Transfer Cable
- Integrated OneQA workforce automation Software License
- Adapter kit(US,EUROPE,AUSTRALIA AND UK)
- Silicone tubing
- Barcode scanner. C750, corded/Bluetooth/2.4wireless
- Wi-Fi Adapter Connector, Micro-Ribbon, USB Adapter, Wi-Fi USB Dongle
- 5-ft plastic Drain Line
- Micro-90 Bottle (225ml)

ITEM 16: TECHNICAL SPECIFICATIONS FOR INFUSION DEVICE ANALYZER

Flow and Volume	
Mean flow range	0 - 300mL/h (Steady flow),0 - 1500 ml/h (Peristaltic flow)
Mean flow resolution	0.001 mL.h
Mean flow accuracy (after > 0.1 mL and > 10s)	$\pm (1 \% + 0.005 \text{ mL/h})$ at < 500 mL/h, otherwise $\pm 2 \%$
Volume range	0 – 100 000 mL
Volume resolution	0.001 mL
Volume accuracy (after > 0.1 mL and > 10s)	$\pm (1 \% + 0.003 \text{ mL})$ at < 500 mL/h, otherwise $\pm 2 \%$
Time range	0 s – 1000 h
Time accuracy	$\pm (0.2 \% + 0.2 \text{ s})$
Time trig volume	$\leq 3 \mu\text{L}$
Graphs	Instant flow and mean flow with up to 1 s resolution (gradually reduced after 1 hour), IEC 60601-2-24 trumpet curve
Occlusion	
Peak pressure range	-200 – +2600 mmHg (-3.8 – +50 psi)
Peak pressure resolution	1 mmHg (0.01 psi)
Peak pressure accuracy	$\pm (1 \% + 5 \text{ mmHg (0.1 psi)})$, 50 ms moving average
Time to peak range	0 s – 1000 h
Time to peak accuracy	$\pm (0.2 \% + 0.05 \text{ s})$
Time to alarm	0 s – 1000 h
Time to alarm accuracy	Operator dependent, manual stop button
Residual volume accuracy	$\pm (2 \% + 0.01 \text{ mL})$
Graphs	Pressure with up to 50 ms resolution (gradually reduced after 3 minutes)
PCA bolus/dual flow measurement	
Bolus volume range	0.02 – 100 000 mL
Bolus volume accuracy (after > 1 s)	$\pm (2 \% + 0.01 \text{ mL})$
Bolus flow range	$10 \times \text{base flow} - 3000 \text{ mL/h}$ (steady flow), $10 \times \text{base flow} - 1500 \text{ mL/h}$ (peristaltic flow)
Bolus flow accuracy (after > 0.1 mL and > 10 s)	$\pm (1 \% + 0.005 \text{ mL/h})$ at < 500 mL/h, otherwise $\pm 2 \%$
Base flow range	1 – 100 mL/h
Base flow accuracy (after > 0.1 mL)	$\pm (1 \% + 0.005 \text{ mL/h})$
Bolus duration range	0 s – 1000 h
Bolus duration accuracy	$\pm (0.2 \% + 0.2 \text{ s})$
Graphs	Instant flow with up to 1 s resolution (gradually reduced after 1 hour), numbered boluses
Pressure measurement	
Range	≤ -500 to $\geq 2500 \text{ mmHg}$
Back pressure setting range	≤ -200 to $\geq 600 \text{ mmHg}$
Accuracy	-500 to 1000 mmHg $\pm 10 \text{ mmHg}$ 1000 to 2500 $\pm 1\%$ of reading

Measurement Units	
Pressure	mmHg, PSI, Bar, Kpa
Volume	mL, g (1 mL = 0.998 g)
General Specifications	
Operating voltage range	90 V ac to 264 V ac
Supply frequency	50/60 Hz
Supply power	Approx 60watts
Altitude	0 m to 3000 m (Nairobi 1680m)
Temperature	
Operating	15 °C to 40 °C (59 °F to 104 °F)
Storage	-20 °C to +70 °C , when drained
Display	At least 10-inch multi touch, 800 * 600 pixels
Data Storage	10000 Measurements
Channels	1 to 4, Depending on Configuration
Connectivity	1 × USB-C for PC communication, 3 × USB-A for peripherals
Safety Standard	Complies with IEC 61010-1:2010, pollution degree 2
Infusion Device Standard	For testing according to IEC 60601-2-24:1998
EMC Standard	Complies with IEC 61326-1:2012
Calibration	Device fully calibrated and traceability demonstrated
Service and Warranty	Free upgraded 24 months warranty
Manuals	MUST provide Technical and Operators manuals(Hard/Soft copies) in English
Training	Onsite training for Biomedical Engineering Personnel on Application and Maintenance
CALIBRATION	<ul style="list-style-type: none"> ➤ Valid Certificate of calibration with test data ➤ Demonstrate Traceability to National / International Standards.
Key features	
<ul style="list-style-type: none"> ➤ Installed with integrated OneQA-enabled workflow automation software which simplifies testing, enhancing accuracy and efficiency ➤ Achieve low-flow stable readings 100x faster than IDA-5 ➤ IntelliPump Technology offers no mess testing with a water recirculation system ➤ No priming required between tests or repriming after initial setup ➤ Test up to four infusion pumps simultaneously ➤ Modular design allows for flexibility; users may add additional channels as needed ➤ Removable channels simplify the calibration process and eliminate downtime ➤ Fully compliant with IEC60601-2-24 standard ➤ Meets or exceeds most OEM low-flow accuracy ➤ Enhanced resolution by 60x vs. IDA-5 ➤ Flow and volume are 100x more accurate vs. IDA-5 ➤ Simply and quickly customize pre-built test procedures for any infusion device 	

- Wireless capability allows you to run test procedures in the field and automatically sync results
- Large 10” touchscreen with an intuitive user interface
- Ability to upload pictures, instructions, and text to procedures
- Global support network to support customers anywhere, anytime
- Immediate results with as little as 0.003mL
- Verifies PCA, bolus, and occlusion functionality

ACCESSORIES

- Syringes
- 5-ft Plastic Drain Line
- USB-C Data Transfer Cable
- Integrated OneQA workforce automation Software License
- Adapter kit(US,EUROPE,AUSTRALIA AND UK)
- Silicone tubing
- Barcode scanner. C750, corded/Bluetooth/2.4wireless
- Wi-Fi Adapter Connector, Micro-Ribbon, USB Adapter, Wi-Fi USB Dongle
- 5-ft plastic Drain Line
- Micro-90 Bottle (225ml)

ITEM 17: TECHNICAL SPECIFICATION FOR DIGITAL TIMER

Key Technical Specifications

- **Frequency Ranges**
- **Input A:** 10^{-10} Hz to 300 MHz
- **Input B:** 10^{-10} Hz to 100 MHz
- **Input C (Optional Pre-scaler):** Up to 1.3 GHz, 2.7 GHz, or 4.2 GHz (via specific modules like PM9621, PM9624, or PM9625).

Measurement Resolution & Precision

- **Time Interval (A to B):** Single-shot resolution of 50 ps (1 ps with averaging).
- **Time Interval Range:** 0 ns to 10^{10} s.
- **Digits Displayed:** 11 digits in 1s measuring time.

Input Specifications

- **Coupling:** AC or DC
- **Impedance:** $1\text{ M}\Omega // 30\text{ pF}$ or 50Ω
- **Sensitivity:** 20 mV rms (<100 MHz); 60 mV rms (>250 MHz).
- **Trigger Level:** Range -5.1 V to +5.1 V (with $\times 1$ attenuation).

Measurement Functions

- Frequency, Period, Ratio, Time Interval, Pulse Width, Rise/Fall Time, Phase, Duty Factor, and Totalize.
- Burst frequency, PRF of burst signals, and Manual/Timed A-B event counting.

General Specs

- **Power Supply:** Wide range from 90 to 265 Vac (no manual switching needed).
- **Memory Storage:** Can save/recall 20 instrument setups (10 user-protected).
- Three-year warranty
- **Onsite training** for Biomedical Engineering Personnel on application and maintenance
- Valid Certificate of Calibration with Test Data

ITEM 18: TECHNICAL SPECIFICATION FOR CURRENT TRANSFORMER

- AC Current Range: 0 to 400.0 A
- DC Current Range: 0 to 400.0 A
- Voltage (AC/DC): up to 600.0 V
- Accuracy: 2.0% ± 5 digits for AC current
- Resistance: up to 40.00 kΩ
- Additional: Temperature measurement (-10.0°C to 400.0°C), capacitance, and frequency
- Three-year warranty
- Onsite training for Biomedical Engineering Personnel on application and maintenance
- Valid Certificate of Calibration with Test Data

ITEM 19: TECHNICAL SPECIFICATION FOR ASSORTED DECADE RESISTORS 1040

General Specifications

- **Resistance Range:** 1 Ω to 100 MΩ
- **Resolution / Steps:** 1 Ω steps using 8 decades
- **Residual Resistance:** Less than 250 mΩ
- **Temperature Coefficient:** Better than 50 ppm/°C
- **Indicators:** Color-coded thumbwheel digits for simple visual grouping (Megohms: red, Kilohms: yellow, Ohms: white)
- **Switch Mechanism:** Multiple gold contact arrangements to ensure low contact resistance and operational reliability

Electrical Ratings

- **Power Rating:** 1 W per resistor
- **Voltage Rating:** Maximum 250 V DC or 176 V AC RMS
- **Accuracy by Decade**
- **1 to 9 Ω:** ± 1%
- **10 to 90 Ω:** ± 0.5%
- **100 Ω to 100 kΩ:** ± 0.1%
- **1 MΩ:** ± 0.1%
- **10 MΩ:** ± 1%

Physical & Mechanical Specifications

- **Terminals:** Three 4 mm safety terminals (two active, one case earth/guard shield)
- **Casing:** Robust metal case fitted with an ergonomic, removable protective rubber boot for grip and added durability

ITEM 20: TECHNICAL SPECIFICATION FOR AC/DC VOLTAGE GENERATOR

- **DC Voltage:** 0.1V to 1000V (Accuracy: 0.0035 + 0.0005 at 10V).
- **AC Voltage:** 0.1V to 750V (up to 300 kHz).
- **Resistance:** 4-wire capabilities 100 Ω to 100 MΩ.
- **Current:** DC to 3A, AC to 3A.
- **Stability:** High-stability components make it suitable for long-term calibration applications
- **RTD Probe Adaptor (9717):** For temperature calibration.
- **Premium Test Lead Set (9796):** Ensures accuracy.
- Three-year warranty

- **Onsite training** for Biomedical Engineering Personnel on application and maintenance

- Valid Certificate of Calibration with Test Data

ITEM 21: TECHNICAL SPECIFICATION FOR PRESSURE TEST COUPLING NIPPLES

- **Maximum Pressure Rating:** Up to 15,000 psi (approx. 1,000 bar) for pneumatic and standard hydraulic quick-connect setups.
- **Thread Options (Device/Gauge Side):** Available in 1/4 NPT male/female, 1/2 NPT male/female, 1/4 BSP male/female, 1/2 BSP male/female, and M20 × 1.5 male/female.
- **Materials:** Constructed from durable **Stainless Steel (SS)** for high structural integrity and corrosion resistance.
- **Sealing Mechanism:** Employs precision O-ring seals (commonly Buna-N, F357, or PTFE) to ensure leak-tight seating at the threads.
- **Tooling:** Designed for finger-tight connections—requires zero thread sealants and eliminates the risk of over-tightening or damaging delicate pressure gauge

ITEM 23: TECHNICAL SPECIFICATION FOR DATA LOGGER

- **Extreme Heat Up to 1000 – 1832 and beyond.**
- If the *logger itself* must travel inside a furnace or oven, it must be housed in a specialized **thermal barrier**
- **Specialized Sensors:** The data logger is usually connected to specialized sensors (like Type-R or Type-B thermocouples) designed to withstand the 1500 - 1700 range, though the logger itself stays inside the protective barrier
- Loggers compatible with Type-K, Type-J, or Type-R thermocouples
- Available in 10 and 20 channel configurations and two case styles.
- Water resistant
- Communication via USB or Bluetooth
- Real time telemetry via USB, RF telemetry and Blue tooth

- Alkaline, NiMH and Lithium battery options.
- Application software
- Three-year warranty
- **Onsite training** for Biomedical Engineering Personnel on application and maintenance
- Valid Certificate of Calibration with Test Data

ITEM 24: TECHNICAL SPECIFICATIONS FOR ULTRASOUND MULTIPURPOSE TEST ANALYZER (PHANTOM)

General specifications	
Operating Temperature	15°C to 40°C
Storage Temperature	2°C to 40°C
Safety standards	IEC/EN 61010-1 3rd Edition; Pollution degree 2
Operating Humidity	10 %-95 % RH (non-condensing)
Operating Altitude	3000 M
Purpose	
Key Features	<ul style="list-style-type: none"> ➤ To ensure accurate screening, diagnosis and monitoring. ➤ To perform efficient QA testing of ultrasound systems and transducers ➤ To use across a variety of applications, including General Radiology, Musculoskeletal, Cardiology, Emergency, Pediatrics, Radiotherapy and Surgical ➤ Meet or exceed ACR, AIUM, EFSUMB, IAC, IEC, IPEM, WHO and other QA standards/guidance ➤ To provide advanced technology for measuring image quality of small parts and intra-cavity ultrasound scanning systems. ➤ To be designed for all QA testing of ultrasound systems and transducers in General Radiology, Musculoskeletal, Cardiology, Emergency, Pediatrics, Radiotherapy and Surgical applications.
Capabilities	
Should Offer	<ul style="list-style-type: none"> ➤ A near-linear response of attenuation-to-frequencies between 2 to 18 MHz. ➤ Response of attenuation-to-frequencies over 8 MHz to support accurate axial resolution and penetration depth representative of human tissue.
Phantom Specification	
	<ul style="list-style-type: none"> ➤ Attenuation Coefficient: 0.5 dB/cm/MHz ➤ Variation of Attenuation with Frequency: $f^{1.08}$ at 0.5 dB/cm/MHz ➤ HE Gel Freezing Point: < 0°C ➤ HE Gel Melting Point: >100°C ➤ Frequency Range: 2 - 18 MHz ➤ Speed of Sound: 1540 m/s

	<ul style="list-style-type: none"> ➤ Scanning Surface: Composite Film ➤ Pin Target Material: Nylon monofilament ➤ Case Material: Extruded ABS Plastic"
Target Specifications	
Cystic Targets	<ul style="list-style-type: none"> ➤ Diameters 2, 4, 6, and 10 mm ➤ Placement 3, 6, 8, and 14 cm deep ➤ Speed of Sound 1540±10m/s ➤ Attenuation Coefficient 0.05±0.01dB/cm/MHz
Grey Scale Targets	<ul style="list-style-type: none"> ➤ Dimensions 10 mm Diameter ➤ Placement 6 cm deep ➤ SOS 1540 ±10 m/s ➤ SOSTD 1.5 m/s/°C ➤ Contrast -6dB, +6dB, high scatter relative to background
Pin Targets	<ul style="list-style-type: none"> ➤ Diameter 0.1 mm ➤ Vertical Spacing 2 cm at 2 to 16 cm deep ➤ Horizontal Spacing 3 cm at 2 and 12 cm deep
Resolution Target Groups	<ul style="list-style-type: none"> ➤ Depth 3, 8, and 14 cm deep"
Tissue Mimicking Background Material	
Water-based gel with appearance of human tissue.	<ul style="list-style-type: none"> ➤ Speed of Sound 1540±10m/s ➤ Attenuation coefficient .0.7±0.05 dB/cm/MHz ➤ 0.5±0.05 dB/cm/MHz ➤ Nonlinearity parameter (B/A). 6.6±0.3..6.7
Service and Warranty	Free upgraded 24 months warranty
Manuals	MUST provide Technical and Operators manuals(Hard/Soft copies
Training	Onsite training for Biomedical Engineering Personnel on Application and Maintenance
CALIBRATION	<ul style="list-style-type: none"> ➤ Valid Certificate of calibration with test data ➤ Demonstrate Traceability to National / International Standards.

Performance Measures

- Image uniformity
- Artifact survey
- Axial and lateral resolution
- Horizontal and vertical distance
- Dead zone
- Depth of penetration
- Signal-to-noise ratio
- Anechoic and echogenic mass resolution
- Gray scale contrast resolution
- Measurement accuracy

Accessories

- Precision sono Transducer holder.
- Padded travel case with shoulder straps

ITEM 25a: ENGINEERS NAVY BLUE DUST COAT

ITEM 25b: LABORATORY COAT WHITE

Poly-Cotton Blend (e.g., 65% Polyester / 35% Cotton): The industry standard. It is easy to wash, doesn't shrink, and withstands daily wear.



ITEM 25c: MALE SAFETY SHOES

Goliath Safety Shoes - Dotted

SPECIFICATIONS:

1. Genuine Leather

The leather should be made of high-quality leather shoes thanks to their durability.

2. Breathable Inner Lining

The inner lining should be breathable, allowing air in to pass through your feet to prevent them from sweating and smelling.

3. Oil, Water, and Acid Resistant

When you're at work, you need shoes that keep your feet dry and safe. The shoe should have strong leather that protects your feet from any oil, water, or chemical spills. I.e. the shoes should be Moisture-wicking & washable lining.

4. Steel Toe

Should be steel-toed safety shoes offer protection from a wide range injuries including falling objects, slips and falls, lacerations/cuts, burns, and punctures.(Steel toe cap protection (200 joules impact tested).

5. Electric Insulating Property

The Safety Shoes should also have insulating properties to protect your feet from any electric shock.

6. Slip Resistant

The outsole of the safety shoes should have an interlocked tread pattern, offering firm contact on the floor to prevent you from slipping.

SAMPLE PHOTO



ITEM 25d: LADIES SAFETY SHOES

The **Ladies** Safety Shoe should have both steel toe protection and a breathable inner lining which provides both protection and comfort.

SPECIFICATIONS:

7. Genuine Leather

The leather should be made of high-quality leather shoes thanks to their durability.

8. Breathable Inner Lining

The inner lining should be breathable, allowing air in to pass through your feet to prevent them from sweating and smelling.

9. Oil, Water, and Acid Resistant

When you're at work, you need shoes that keep your feet dry and safe. The shoe should have strong leather that protects your feet from any oil, water, or chemical spills.

10. Steel Toe

Should be steel-toed safety shoes offer protection from a wide range injuries including falling objects, slips and falls, lacerations/cuts, burns, and punctures.

11. Electric Insulating Property

The Safety Shoes should also have insulating properties to protect your feet from any electric shock.

12. Slip Resistant

The outsole of the safety shoes should have an interlocked tread pattern, offering firm contact on the floor to prevent you from slipping.

SAMPLE PHOTO



ITEM 25e: BODY SUIT (TROUSER & JACKET) REFLECTOR

As per sample (to be provided during site visit and taking of individual measurements)

SECTION VI: SCHEDULE OF REQUIREMENTS FOR PATIENT LINEN

NB: Mandatory Compliance by All Contractors/Service Providers/Consultants Reference is made to the Public Procurement Regulatory Circular PPRA/6/5 Vol II (224) ; Circular No 01/2024 dated 30th August 2024. In order, to provide funds for the capacity development of persons involved in public procurement and asset disposal proceedings through mentoring, and technical assistance, the Cabinet Secretary, National Treasury and Economic Planning, pursuant to Sections 24(5)(d) and 180 of the Act, issued the Public Procurement Capacity Building Levy Order, 2023 (hereinafter referred to as 'The Levy Order, 2023') vide Legal Notice No. 206 of 6th November, 2023. Paragraph 3(1) of The Levy Order, 2023; provides that there shall be paid a Levy by a supplier on all procurement contracts signed between the supplier and a procuring entity, at the rate of zero point zero three per centum (0.03%) of the value of the signed contract, exclusive of applicable taxes.

No	ITEM DESCRIPTION	SPECIFICATIONS	Unit of issue	QTY	UNIT PRICE	TOTAL PRICE
1	Vital signs simulator	As Per Attached Specifications	NO	1		
2	Electrosurgical Analyzer	As Per Attached Specifications	NO	1		
3	Electrical safety Analyzer	As Per Attached Specifications	NO	1		
4	20-50 degrees Human Black Body source	As Per Attached Specifications	NO	1		
5	Biometer	As Per Attached Specifications	NO	1		
6	Biosafety cabinet testing tools	As Per Attached Specifications	Set	1		
7	Fetal simulator	As Per Attached Specifications	NO	1		
8	Current meter	As Per Attached Specifications	NO	1		
9	Tachometer	As Per Attached Specifications	NO	1		
10	Oxygen analyzer	As Per Attached Specifications	NO	1		
11	Thermohygrometer	As Per Attached Specifications	NO	1		
12	700 bar Hydraulic Test Pressure Kit	As Per Attached Specifications	NO	1		
13	Digital Pressure Gauge	As Per Attached Specifications	NO	1		
14	ECG Simulator	As Per Attached Specifications	NO	1		
15	Infusion Analyzer	As Per Attached Specifications	NO	1		
16	Defibrillator analyzer	As Per Attached Specifications	NO	1		
17	Digital Timer	As Per Attached Specifications	NO	1		
18	Current transformer	As Per Attached Specifications	NO	1		
19	1040 decade resistors	As Per Attached Specifications	NO	1		
20	AC/DC voltage generator	As Per Attached Specifications	NO	1		
21	Pressure Testing Coupling Nipples	As Per Attached Specifications	Set	1		

22	Calibration/Repair of Incubator Analyzer	As Per Attached Specifications	NO	1		
23	Data Logger	As Per Attached Specifications	No	1		
24	Ultrasound Analyzer	As Per Attached Specifications	No	1		
25	PPE					
25 a	Engineers Navy Blue Dust Coat	As per sample to be viewed	NO	1		
25 b	Engineers White Lab Coat	As per sample to be viewed	NO	1		
26 c	Male Safety Boots	As per sample to be viewed	NO	1		
25 d	Female Safety Boots	As per sample to be viewed	NO	1		
25 e	Body Suit (Trouser & Jacket) Reflector	As per sample to be viewed	NO	1		
27	Spartan MIG/250i Heavy Duty 240/315V Complete with cylinder	MIG Welding, TIG Welding and Spot Welding	NO	1		
28	Heavy Duty Drill	With Cord	NO	1		
29	Heavy Duty Drill	Cordless	No	1		
30	Ray Safe (Detector)	Ray Safe X2	No	1		
31	Electronic Diagnosis Microscope with display	Standard	No	1		
32	Soldering Station		No	1		
33	Allen Keys	Assorted Sizes - Imperial				
34	Allen Keys	Assorted sizes -Metric	Box	1		
35	Tool Kit with rachets	Tolsen	No	1		
36	Battery Analyzer	Fluke BT521	No	1		
37	Energy Meter	Fluke 1738	No	1		

Price Schedule: Goods Manufactured Outside Kenya, to be Imported

(Group C Tenders, goods to be imported)
Currencies in accordance with ITT 15

Date: _____

ITT No: _____

Alternative No: _____

Page N° _____ of _____

1	2	3	4	5	6	7	8	9		
Line Item	Description of Goods	Country of Origin	Delivery Date as defined by Incoterms	Quantity and physical unit	Unit price CIP <i>[insert place of destination]</i> in accordance with ITT 14.8(b)(i)	CIP Price per line item (Col. 5x6)	Price per line item for inland transportation and other services required in Kenya to convey the Goods to their final destination specified in TDS	Total Price item	per	Line
<i>[insert number of the item]</i>	<i>[insert name of good]</i>	<i>[insert country of origin of the Good]</i>	<i>[insert quoted Delivery Date]</i>	<i>[insert number of units to be supplied and name of the physical unit]</i>	<i>[insert unit price CIP per unit]</i>	<i>[insert total CIP price per line item]</i>	<i>[insert the corresponding price per line item]</i>	<i>[insert total price of the line item]</i>		
							Total Price			

Name of tenderer *[insert complete name of tenderer]* Signature of tenderer *[signature of person signing the Tender]* Date *[Insert Date]*

Price Schedule: Goods Manufactured Outside Kenya, already imported*

(Group C Tenders, Goods already imported)
Currencies in accordance with ITT 15

Date: _____
ITT No: _____
Alternative No: _____
Page N° _____ of _____

1	2	3	4	5	6	7	8	9	10	11	12
Line Item N°	Description of Goods	Country of Origin	Delivery Date as defined by Incoterms	Quantity and physical unit	Unit price including Custom Duties and Import Taxes paid, in accordance with ITT 14.8(c)(i)	Custom Duties and Import Taxes paid per unit in accordance with ITT 14.8(c)(ii), [to be supported by documents]	Unit Price net of custom duties and import taxes, in accordance with ITT 14.8 (c) (iii) (Col. 6 minus Col.7)	Price per line item net of Custom Duties and Import Taxes paid, in accordance with ITT 14.8(c)(i) (Col. 5×8)	Price per line item for inland transportation and other services required in Kenya to convey the goods to their final destination, as specified in TDS in accordance with ITT 14.8 (c)(v)	Sales and other taxes paid or payable per item if Contract is awarded (in accordance with ITT 14.8(c)(iv)	Total Price per line item (Col. 9+10)
<i>[insert number of the item]</i>	<i>[insert name of Goods]</i>	<i>[insert country of origin of the Good]</i>	<i>[insert quoted Delivery Date]</i>	<i>[insert number of units to be supplied and name of the physical unit]</i>	<i>[insert unit price per unit]</i>	<i>[insert custom duties and taxes paid per unit]</i>	<i>[insert unit price net of custom duties and import taxes]</i>	<i>[insert price per line item net of custom duties and import taxes]</i>	<i>[insert price per line item for inland transportation and other services required in Kenya]</i>	<i>[insert sales and other taxes payable per item if Contract is awarded]</i>	<i>[insert total price per line item]</i>
										Total Tender Price	

Name of tenderer *[insert complete name of tenderer]* Signature of tenderer *[signature of person signing the Tender]* Date *[insert date]*

* *[For previously imported Goods, the quoted price shall be distinguishable from the original import value of these Goods declared to customs and shall include any rebate or mark-up of the local agent or representative and all local costs except import duties and taxes, which have been and/or have to be paid by the Procuring Entity. For clarity, the tenderers are asked to quote the price including import duties, and additionally to provide the import duties and the price net of import duties which is the difference of those values.]*

Price Schedule: Goods Manufactured in Kenya

Kenya

(Group A and B Tenders)
Currencies in accordance with ITT 15

Date: _____
ITT No: _____
Alternative No: _____
Page N° _____ of _____

1	2	3	4	5	6	7	8	9	10	
Line Item	Description of Goods	Delivery Date	as	Quantity and physical unit	Unit price EXW	Total EXW price per line item (Col. 4x5)	Price per line item for inland transportation and other services required in Kenya to convey the Goods to their final destination	Cost of local labor, raw materials and components from within origin in Kenya % of Col. 5	Sales and other taxes payable per line item if Contract is awarded (in accordance with ITT 14.8(a)(ii))	Total Price per line item (Col. 6+7)
N°		defined by Incoterms		physical unit		line item (Col. 4x5)	transportation and other services required in Kenya to convey the Goods to their final destination	components from within origin in Kenya % of Col. 5	Contract is awarded (in accordance with ITT 14.8(a)(ii))	
[insert number of the item]	[insert name of Good]	[insert quoted Delivery Date]		[insert number of units to be supplied and name of the physical unit]	[insert EXW unit price]	[insert total EXW price per line item]	[insert the corresponding price per line item]	[Insert cost of local labor, raw material and components from within the Purchase's country as a % of the EXW price per line item]	[insert sales and other taxes payable per line item if Contract is awarded]	[insert price per item]
									Total Price	

Name of tenderer [insert complete name of tenderer] Signature of tenderer [signature of person signing the Tender] Date [insert date]

Price and Completion Schedule - Related Services

Currencies in accordance with ITT 15

Date:

ITT _____ No:

Alternative _____ No:

Page N° _____ of _____

1	2	3	4	5	6	7
Service N°	Description of Services (excludes inland transportation and other services required in Kenya to convey the goods to their final destination)	Country of Origin	Delivery Date at place of Final destination	Quantity and physical unit	Unit price	Total Price per Service (Col. 5*6 or estimate)
<i>[insert number of the Service]</i>	<i>[insert name of Services]</i>	<i>[insert country of origin of the Services]</i>	<i>[insert delivery date at place of final destination per Service]</i>	<i>[insert number of units to be supplied and name of the physical unit]</i>	<i>[insert unit price per item]</i>	<i>[insert total price per item]</i>
				Total Tender Price		

Name of tenderer *[insert complete name of tenderer]* Signature of tenderer *[signature of person signing the Tender]* Date *[insert date]*

FORM OF TENDER SECURITY-[Option 1–Demand Bank Guarantee]**Beneficiary:** _____**Request for Tenders No:**
_____**Date:** _____**TENDER GUARANTEE No.:** _____**Guarantor:** _____

1. We have been informed that _____ (here inafter called "the Applicant") has submitted or will submit to the Beneficiary its Tender (here inafter called" the Tender") for the execution of _____ under Request for Tenders No. _____ ("the ITT").
2. Furthermore, we understand that, according to the Beneficiary's conditions, Tenders must be supported by a Tender guarantee.
3. At the request of the Applicant, we, as Guarantor, hereby irrevocably undertake to pay the Beneficiary any sum or sums not exceeding in total an amount of _____ (_____) upon receipt by us of the Beneficiary's complying demand, supported by the Beneficiary's statement, whether in the demand itself or a separate signed document accompanying or identifying the demand, stating that either the Applicant:
 - (a) has withdrawn its Tender during the period of Tender validity set forth in the Applicant's Letter of Tender ("the Tender Validity Period"), or any extension thereto provided by the Applicant; or
 - b) having been notified of the acceptance of its Tender by the Beneficiary during the Tender Validity Period or any extension there to provided by the Applicant, (i) has failed to execute the contract agreement, or (ii) has failed to furnish the Performance.
4. This guarantee will expire: (a) if the Applicant is the successful Tenderer, upon our receipt of copies of the contract agreement signed by the Applicant and the Performance Security and, or (b) if the Applicant is not the successful Tenderer, upon the earlier of (i) our receipt of a copy of the Beneficiary's notification to the Applicant of the results of the Tendering process; or (ii) thirty days after the end of the Tender Validity Period.
5. Consequently, any demand for payment under this guarantee must be received by us at the office indicated above on or before that date.

[signature(s)]

Note: All italicized text is for use in preparing this form and shall be deleted from the final product

FORMAT OF TENDER SECURITY [Option 2–Insurance Guarantee]

TENDER GUARANTEE No.: _____

1. Whereas [*Name of the tenderer*] (hereinafter called “the tenderer”) has submitted its tender dated [*Date of submission of tender*] for the [*Name and/or description of the tender*] (hereinafter called “the Tender”) for the execution of _____ under Request for Tenders No. _____ (“the ITT”).
2. KNOW ALL PEOPLE by these presents that WE of [**Name of Insurance Company**] having our registered office at (hereinafter called “the Guarantor”), are bound unto [*Name of Procuring Entity*] (hereinafter called “the Procuring Entity”) in the sum of (Currency and guarantee amount) for which payment well and truly to be made to the said Procuring Entity, the Guarantor binds itself, its successors and assigns, jointly and severally, firmly by these presents.

Sealed with the Common Seal of the said Guarantor this ____ day of _____ 20 ____.

3. NOW, THEREFORE, THE CONDITION OF THIS OBLIGATION is such that if the Applicant:
 - a) has withdrawn its Tender during the period of Tender validity set forth in the Principal's Letter of Tender (“the Tender Validity Period”), or any extension thereto provided by the Principal; or
 - b) having been notified of the acceptance of its Tender by the Procuring Entity during the Tender Validity Period or any extension thereto provided by the Principal; (i) failed to execute the Contract agreement; or (ii) has failed to furnish the Performance Security, in accordance with the Instructions to tenderers (“ITT”) of the Procuring Entity's Tendering document.

then the guarantee undertakes to immediately pay to the Procuring Entity up to the above amount upon receipt of the Procuring Entity's first written demand, without the Procuring Entity having to substantiate its demand, provided that in its demand the Procuring Entity shall state that the demand arises from the occurrence of any of the above events, specifying which event(s) has occurred.

4. This guarantee will expire: (a) if the Applicant is the successful Tenderer, upon our receipt of copies of the contract agreement signed by the Applicant and the Performance Security and, or (b) if the Applicant is not the successful Tenderer, upon the earlier of (i) our receipt of a copy of the Beneficiary's notification to the Applicant of the results of the Tendering process; or (ii) twenty-eight days after the end of the Tender Validity Period.
5. Consequently, any demand for payment under this guarantee must be received by us at the office indicated above on or before that date.

[Date]

[Witness]

[Signature of the Guarantor]

[Seal]

Note: All italicized text is for use in preparing this form and shall be deleted from the final product.

FORM OF TENDER-SECURING DECLARATION

[The Bidder shall complete this Form in accordance with the instructions indicated]

Date:..... *[insert date (as day, month and year) of Tender Submission]*

Tender No.:..... *[Insert number of tendering process]*

To:..... *[insert complete name of*

Purchaser] I/We, the undersigned, declare that:

1. I/We understand that, according to your conditions, bids must be supported by a Tender-Securing Declaration.
2. I/We accept that I/we will automatically be suspended from being eligible for tendering in any contract with the Purchaser for the period of time of *[insert number of months or years]* starting on *[insert date]*, if we are in breach of our obligation(s) under the bid conditions, because we – (a) have withdrawn our tender during the period of tender validity specified by us in the Tendering Data Sheet; or (b) having been notified of the acceptance of our Bid by the Purchaser during the period of bid validity, (i) fail or refuse to execute the Contract, if required, or (ii) fail or refuse to furnish the Performance Security, in accordance with the instructions to tenders.
3. I/We understand that this Tender Securing Declaration shall expire if we are not the successful Tenderer(s), upon the earlier of:
 - a) our receipt of a copy of your notification of the name of the successful Tenderer; or
 - b) thirty days after the expiration of our Tender.
4. I/We understand that if I am/we are/in a Joint Venture, the Tender Securing Declaration must be in the name of the Joint Venture that submits the bid, and the Joint Venture has not been legally constituted at the time of bidding, the Tender Securing Declaration shall be in the names of all future partners as named in the letter of intent.

Signed:.....

Capacity / title (director or partner or sole proprietor, etc.)

Name:

Duly authorized to sign the bid for and on behalf of *[insert complete name of*

Tenderer]. Dated on day of..... *[Insert date of signing]*.

Seal or stamp.

MANUFACTURER’S AUTHORIZATION FORM

[The tenderer shall require the Manufacturer to fill in this Form in accordance with the instructions indicated. This letter of authorization should be on the letterhead of the Manufacturer and should be signed by a person with the proper authority to sign documents that are binding on the Manufacturer. The tenderer shall include it in its Tender, if so indicated in the TDS.]

Date:.....*[insert date (as day, month and year) of Tender submission]*

ITT No.:.....*[insert number of ITT*

process] Alternative No..... [insert identification No if this is a

Tender for an alternative]

To *[Insert complete name of Procuring*

Entity] WHEREAS

We..... *[insert complete name of Manufacturer], who are official manufacturers of..... [insert type of goods manufactured], having factories at [insert full address of Manufacturer's factories], do hereby authorize [insert complete name of tenderer] to submit a Tender the purpose of which is to provide the following Goods, manufactured by us..... [insert name and or brief description of the Goods], and to subsequently negotiate and sign the Contract.*

We hereby extend our full guarantee and warranty in accordance with Clause 28 of the General Conditions of Contract, with respect to the Goods offered by the above firm.

Signed.....*[Insert signature(s) of authorized representative(s) of the Manufacturer]*

Name:.....*[Insert complete name(s) of authorized representative(s) of the Manufacturer]*

Title *[Insert title]*

Dated on _____ day of _____, _____ *[insert date of signing]*

PART 2: SUPPLY REQUIREMENTS

Section V - Schedule of Requirements

Notes for Preparing the Schedule of Requirements

The Schedule of Requirements shall be included in the Tendering document by the Procuring Entity, and shall cover, at a minimum, a description of the goods and services to be supplied and the delivery schedule.

The objective of the Schedule of Requirements is to provide sufficient information to enable tenderers to prepare their Tenders efficiently and accurately, in particular, the Price Schedule, for which a form is provided in Section IV. In addition, the Schedule of Requirements, together with the Price Schedule, should serve as a basis in the event of quantity variation at the time of award of contract pursuant to ITT 42.1.

The date or period for delivery should be carefully specified, taking into account (a) the implications of delivery terms stipulated in the Instructions to tenderers pursuant to the *Incoterms* rules that “delivery” takes place when goods are delivered **to the final place of delivery**, and (b) the date prescribed herein from which the Procuring Entity's delivery obligations start (i.e., notice of award, contract signature, opening or confirmation of the letter of credit).

Section V - Schedule Requirements Technical Specifications

1. List of Related Services and Completion Schedule (NOT APPLICABLE)

[This table shall be filled in by the Procuring Entity. The Required Completion Dates should be realistic, and consistent with the required Goods Delivery Dates (as per Incoterms)].

Service	Description of Service	Quantity¹	Physical Unit	Place where Services shall be performed	Final Completion Date(s) of Services
<i>[insert Service No]</i>	<i>[insert description of Related Services]</i>	<i>[insert quantity of items to be supplied]</i>	<i>[insert physical unit for the items]</i>	<i>[insert name of the Place]</i>	<i>[insert required Completion Date(s)]</i>

¹If applicable

2. Technical Specifications

The purpose of the Technical Specifications (TS), is to define the technical characteristics of the Goods and Related Services required by the Procuring Entity. The Procuring Entity shall prepare the detailed TS consider that:

- i) The TS constitute the benchmarks against which the Procuring Entity will verify the technical responsiveness of Tenders and subsequently evaluate the Tenders. Therefore, well-defined TS will facilitate preparation of responsive Tenders by tenderers, as well as examination, evaluation, and comparison of the Tenders by the Procuring Entity.
- ii) The TS shall require that all goods and materials to be incorporated in the goods be new, unused, and of the most recent or current models, and that they incorporate all recent improvements in design and materials, unless provided for otherwise in the contract.
- iii) The TS shall make use of best practices. Samples of specifications from successful similar procurements in the same country or sector may provide a sound basis for drafting the TS.
- iv) The PPRA encourages the use of metric units.
- v) Standardizing technical specifications may be advantageous, depending on the complexity of the goods and the repetitiveness of the type of procurement. Technical Specifications should be broad enough to avoid restrictions on workmanship, materials, and equipment commonly used in manufacturing similar kinds of goods.
- vi) Standards for equipment, materials, and workmanship specified in the Tendering document shall not be restrictive. Recognized international standards should be specified as much as possible. Reference to brand names, catalogue numbers, or other details that limit any materials or items to a specific manufacturer should be avoided as far as possible. Where unavoidable, such item description should always be followed by the words "or substantially equivalent." When other particular standards or codes of practice are referred to in the TS, whether from the Procuring Entity's or from other eligible countries, a statement should follow other authoritative standards that ensure at least a substantially equal quality, then the standards mentioned in the TS will also be acceptable.
- vii) Reference to brand names and catalogue numbers should be avoided as far as possible; where unavoidable the words "or at least equivalent" shall always follow such references.
- viii) Technical Specifications shall be fully descriptive of the requirements in respect of, but not limited to, the following:
 - a) Standards of materials and workmanship required for the production and manufacturing of the Goods.
 - b) Any sustainable procurement technical requirements shall be clearly specified.

To encourage tenderers' innovation in addressing sustainable procurement requirements, as long as the Tender evaluation criteria specify the mechanism for monetary adjustments for the purpose of Tender comparisons, tenderers may be invited to offer Goods that exceeds the specified minimum sustainable procurement requirements.

- i) Detailed tests required (type and number).
- ii) Other additional work and/or Related Services required to achieve full delivery/completion.
- iii) Detailed activities to be performed by the Supplier, and participation of the Procuring Entity thereon.
- iv) List of detailed functional guarantees covered by the Warranty and the specification of the liquidated damages to be applied in the event that such guarantees are not met.

The TS shall specify all essential technical and performance characteristics and requirements, including guaranteed or acceptable maximum or minimum values, as appropriate. Whenever necessary, the Procuring Entity shall include an additional ad-hoc Tendering form (to be an Attachment to the Letter of Tender), where the tenderer shall provide detailed information on such technical performance characteristics in respect to the corresponding acceptable or guaranteed values.

When the Procuring Entity requests that the tenderer provides in its Tender a part or all of the Technical Specifications, technical schedules, or other technical information, the Procuring Entity shall specify in detail the nature and extent of the required information and the manner in which it has to be presented by the tenderer in its Tender.

If a summary of the Technical Specifications(TS) has to be provided, the Procuring Entity shall insert information in the table below. The tenderer shall prepare a similar table to justify compliance with the requirements.

Summary of Technical Specifications: The Goods and Related Services shall comply with following Technical Specifications and Standards:

Item No	Name of Goods or Related Service	Technical Specifications and Standards
[insert item No]	[insert name]	[insert TS and Standards]

Detailed Technical Specifications and Standards *[insert whenever necessary]. [Insert detailed description of TS]*

3. Drawings

This Tendering document includes*[Insert “the following” or “no”] drawings. [If documents shall be included, insert the following List of Drawings].*

List of Drawings		
Drawing No.	Drawing Name	Purpose

4. Inspections and Tests

The following inspections and tests shall be performed*[Insert list of inspections and tests]*

PART 3 - CONDITIONS OF CONTRACT AND CONTRACT FORMS

SECTION VI - GENERAL CONDITIONS OF CONTRACT

1. Definitions

In the Conditions of Contract (“these Conditions”), which include Special Conditions, Parts A and B, and these General Conditions, the following words and expressions shall have the meanings stated. Words indicating persons or parties include corporations and other legal entities, except where the context requires otherwise.

- a) “Contract” means the Contract Agreement entered into between the Procuring Entity and the Supplier, together with the Contract Documents referred to therein, including all attachments, appendices, and all documents incorporated by reference therein.
- b) “Contract Documents” means the documents listed in the Contract Agreement, including any amendments thereto.
- c) “Contract Price” means the price payable to the Supplier as specified in the Contract Agreement, subject to such additions and adjustments thereto or deductions therefrom, as may be made pursuant to the Contract.
- d) “Day” means calendar day.
- e) “Completion” means the fulfilment of the Related Services by the Supplier in accordance with the terms and conditions set forth in the Contract.
- f) “GCC” means the General Conditions of Contract.
- g) “Goods” means all of the commodities, raw material, machinery and equipment, and/or other materials that the Supplier is required to supply to the Procuring Entity under the Contract.
- h) “Procuring Entity” means the Procuring Entity purchasing the Goods and Related Services, as **specified in the SCC.**
- i) “Related Services” means the services incidental to the supply of the goods, such as insurance, delivery, installation, commissioning, training and initial maintenance and other such obligations of the Supplier under the Contract.
- j) “SCC” means the Special Conditions of Contract.
- k) “Subcontractor” means any person, private or government entity, or a combination of the above, to whom any part of the Goods to be supplied or execution of any part of the Related Services is subcontracted by the Supplier.
- l) “Supplier” means the person, private or government entity, or a combination of the above, whose Tender to perform the Contract has been accepted by the Procuring Entity and is named as such in the Contract Agreement.
- m) “**Base Date**” means a date 30 day prior to the submission of tenders.
- n) “**Laws**” means all national legislation, statutes, ordinances, and regulations and by-laws of any legally constituted public authority.
- o) “**Letter of Acceptance**” means the letter of formal acceptance, signed by the contractor. Procuring Entity, including any annexed memoranda comprising agreements between and signed by both Parties.
- p) “**Procuring Entity**” means the Entity named in the Special Conditions of Contract.

2. Interpretation

If the context so requires it, singular means plural and vice versa.

Incoterms

- a) Unless inconsistent with any provision of the Contract, the meaning of any trade term and the rights and obligations of parties thereunder shall be as prescribed by Incoterms specified **in the SCC.**
- b) The terms EXW and CIP and other similar terms, when used, shall be governed by the rules

prescribed in the current edition of Incoterms specified in the **SCC** and published by the International Chamber of Commerce in Paris, France.

3. **Contract Documents**

Subject to the order of precedence set forth in the Contract Agreement, all documents forming the Contract (and all parts thereof) are intended to be correlative, complementary, and mutually explanatory. The Contract Agreement shall be read as a whole. The documents forming the Contract shall be interpreted in the following order of priority:

- a) the Contract Agreement,
- b) the Letter of Acceptance,
- c) the General Conditions of Contract
- d) Special Conditions of Contract
- e) the Form of Tender,
- f) the Specifications and Schedules of the Drawings (if any), and
- g) the Schedules of Requirements, Price Schedule and any other documents forming part of the Contract.

4. **Fraud and Corruption**

The supplier shall comply with anti-corruption laws and guidelines and the prevailing sanctions, policies and procedures as set forth in the Laws of Kenya.

The Supplier shall disclose any commissions, gratuity or fees that may have been paid or are to be paid to agents or any other person with respect to the Tendering process or execution of the Contract. The information disclosed must include at least the name and address of the agent or other party, the amount and currency, and the purpose of the commission, gratuity or fee.

Entire Agreement

4.3.1 The Contract constitutes the entire agreement between the Procuring Entity and the Supplier and supersedes all communications, negotiations and agreements (whether written or oral) of the parties with respect thereto made prior to the date of Contract.

Amendment

No amendment or other variation of the Contract shall be valid unless it is in writing, is dated, expressly refers to the Contract, and is signed by a duly authorized representative of each party thereto.

Non-waiver

- a) Subject to GCC Sub-Clause 4.5(b) below, no relaxation, forbearance, delay, or indulgence by either party in enforcing any of the terms and conditions of the Contract or the granting of time by either party to the other shall prejudice, affect, or restrict the rights of that party under the Contract, neither shall any waiver by either party of any breach of Contract operate as waiver of any subsequent or continuing breach of Contract.
- b) Any waiver of a party's rights, powers, or remedies under the Contract must be in writing, dated, and signed by an authorized representative of the party granting such waiver, and must specify the right and the extent to which it is being waived.

Severability

If any provision or condition of the Contract is prohibited or rendered invalid or unenforceable, such prohibition, invalidity or unenforceability shall not affect the validity or enforceability of any other provisions and conditions of the Contract.

5. **Language**

The Contract as well as all correspondence and documents relating to the Contract exchanged by the Supplier and the Procuring Entity, shall be written in the **English Language**. Supporting documents and printed literature that are part of the Contract may be in another language provided they are accompanied by an accurate and certified translation of the relevant passages in the **English Language**, in which case, for purposes of interpretation of the Contract, the English language is translation shall govern.

The Supplier shall bear all costs of translation to the governing language and all risks of the accuracy of

such translation, for documents provided by the Supplier.

6. Joint Venture, Consortium or Association

6.1 If the Supplier is a joint venture, consortium, or association, all of the parties shall be jointly and severally liable to the Procuring Entity for the fulfilment of the provisions of the Contract and shall designate one member of the joint venture, consortium, or association to act as a leader with authority to bind the joint venture, consortium, or association. The composition or the constitution of the joint venture, consortium, or association shall not be altered without the prior written consent of the Procuring Entity.

7. Eligibility

The Supplier and its Subcontractors shall have the nationality of an eligible country. A Supplier or Sub-contractor shall be deemed to have the nationality of a country if it is a citizen or constituted, incorporated, or registered, and operates in conformity with the provisions of the laws of that country.

All Goods and Related Services to be supplied under the Contract shall have their origin in Eligible Countries. For the purpose of this Clause, origin means the country where the goods have been grown, mined, cultivated, produced, manufactured, or processed; or through manufacture, processing, or assembly, another commercially recognized article results that differs substantially in its basic characteristics from its components.

The Tenderer, if a Kenyan firm, must submit with its tender a valid tax compliance certificate from the Kenya Revenue Authority.

8. Notices

Any notice given by one party to the other pursuant to the Contract shall be in writing to the address specified in the SCC. The term "in writing" means communicated in written form with proof of receipt.

A notice shall be effective when delivered or on the notice's effective date, whichever is later.

9. Governing Law

The Contract shall be governed by and interpreted in accordance with the laws of Kenya.

Throughout the execution of the Contract, the Supplier shall comply with the import of goods and services prohibitions in Kenya:

- a) where, as a matter of law, compliance or official regulations, Kenya prohibits commercial relations with that country or any import of goods from that country or any payments to any country, person, or entity in that country; or
- b) by an act of compliance with a decision of the United Nations Security Council taken under Chapter VII of the Charter of the United Nations, Kenya prohibits any import of goods from that country or any payments to any country, person, or entity.

10. Settlement of Disputes

The Procuring Entity and the Supplier shall make every effort to resolve amicably by direct negotiation any disagreement or dispute arising between them under or in connection with the Contract.

If, after thirty (30) days, the parties have failed to resolve their dispute or difference by such mutual consultation, then either the Procuring Entity or the Supplier may give notice to the other party of its intention to commence arbitration, as hereinafter provided, as to the matter in dispute, and no arbitration in respect of this matter may be commenced unless such notice is given. Any dispute or difference in respect of which a notice of intention to commence arbitration has been given in accordance with this Clause shall be finally settled by arbitration. Arbitration may be commenced prior to or after delivery of the Goods under the Contract.

Arbitration proceedings shall be conducted as follows:

Any claim or dispute between the Parties arising out of or in connection with the Contract not settled amicably in accordance with Sub-Clause 10.1 shall be finally settled by arbitration.

No arbitration proceedings shall be commenced on any claim or dispute where notice of a claim or dispute has not been given by the applying party within thirty days of the occurrence or discovery of the matter or issue giving rise to the dispute.

Notwithstanding the issue of a notice as stated above, the arbitration of such a claim or dispute shall not commence unless an attempt has in the first instance been made by the parties to settle such claim or dispute amicably with or without the assistance of third parties. Proof of such attempt shall be required.

The Arbitrator shall, without prejudice to the generality of his powers, have powers to direct such measurements, computations, or valuations as may in his opinion be desirable in order to determine the

rights of the parties and assess and award any sums which ought to have been the subject of or included in any due payments.

Neither Party shall be limited in the proceedings before the arbitrators to the evidence, or to the reasons for the dispute given in its notice of a claim or dispute.

Arbitration may be commenced prior to or after delivery of the goods. The obligations of the Parties shall not be altered by reason of any arbitration being conducted during the progress of the delivery of goods.

The terms of the remuneration of each or all the members of Arbitration shall be mutually agreed upon by the Parties when agreeing the terms of appointment. Each Party shall be responsible for paying one-half of this remuneration.

Arbitration Proceedings

Arbitration proceedings with national suppliers will be conducted in accordance with the Arbitration Laws of Kenya. In case of any claim or dispute, such claim or dispute shall be notified in writing by either party to the other with a request to submit it to arbitration and to concur in the appointment of an Arbitrator within thirty days of the notice. The dispute shall be referred to the arbitration and final decision of a person or persons to be agreed between the parties. Failing agreement to concur in the appointment of an Arbitrator, the Arbitrator shall be appointed, on the request of the applying party, by the Chairman or Vice Chairman of any of the following professional institutions;

- (i) Kenya National Chamber of Commerce
- (ii) Chartered Institute of Arbitrators (Kenya Branch)
- (iii) The Law Society of Kenya

The institution written to first by the aggrieved party shall take precedence over all other institutions.

Alternative Arbitration Proceedings

Alternatively, the Parties may refer the matter to the Nairobi Centre for International Arbitration (NCIA) which offers a neutral venue for the conduct of national and international arbitration with commitment to providing institutional support to the arbitral process.

Arbitration with Foreign Suppliers

Arbitration with foreign suppliers shall be conducted in accordance with the arbitration rules of the United Nations Commission on International Trade Law (UNCITRAL); or with proceedings administered by the International Chamber of Commerce (ICC) and conducted under the ICC Rules of Arbitration; by one or more arbitrators appointed in accordance with said arbitration rules.

The place of arbitration shall be a location specified in the **SCC**; and the arbitration shall be conducted in the language for communications defined in Sub-Clause 1.4 [Law and Language].

Alternative Arbitration Proceedings

Alternatively, the Parties may refer the matter to the Nairobi Centre for International Arbitration (NCIA) which offers a neutral venue for the conduct of national and international arbitration with commitment to providing institutional support to the arbitral process.

Failure to Comply with Arbitrator's Decision

The award of such Arbitrator shall be final and binding upon the parties.

10.6.1 In the event that a Party fails to comply with a final and binding Arbitrator's decision, then the other Party may, without prejudice to any other rights it may have, refer the matter to a competent court of law.

Contract operations continue

Notwithstanding any reference to arbitration herein,

- a) the parties shall continue to perform their respective obligations under the Contract unless they otherwise agree; and
- b) the Procuring Entity shall pay the Supplier any monies due the Supplier.

11. Inspections and Audit by the Procuring Entity

The Supplier shall keep, and shall cause its Subcontractors to keep, accurate and systematic accounts and records in respect of the Goods in such form and details as will clearly identify relevant time, changes and costs.

Pursuant to paragraph 2.2 of Instruction to Tenderers, the Supplier shall permit and shall cause its subcontractors to permit, the Procuring Entity and/or persons appointed by the Procuring Entity or by other statutory bodies of the Government to inspect the Site and/or the accounts and records relating to the procurement process, selection and/or contract execution, and to have such accounts and records audited by auditors appointed by the Procuring Entity. The Supplier's and its Subcontractors' attention is drawn to Sub-Clause 3.1 which provides, inter alia, that acts intended to materially impede the exercise of the Procuring Entity's inspection and audit rights constitute a prohibited practice subject to contract termination, as well as to a determination of ineligibility.

12. Scope of Supply

12.1 The Goods and Related Services to be supplied shall be as specified in the Schedule of Requirements.

13. Delivery and Documents

13.1 Subject to GCC Sub-Clause 33.1, the delivery of the Goods and completion of the Related Services shall be in accordance with the List of Goods and Delivery Schedule specified in the Supply Requirements. The details of shipping and other documents to be furnished by the Supplier are specified in the **SCC**.

14. Supplier's Responsibilities

14.1 The Supplier shall supply all the Goods and Related Services included in the Scope of Supply in accordance with GCC Clause 12, and the Delivery and Completion Schedule, as per GCC Clause 13.

15. Contract Price

Prices charged by the Supplier for the Goods supplied and the Related Services performed under the Contract shall not vary from the prices quoted by the Supplier in its Tender, with the exception of any price adjustments authorized in the **SCC**.

Where the contract price is different from the corrected tender price, in order to ensure the supplier is not paid less or more relative to the contract price (*which would be the tender price*), any partial payment valuation based on rates in the schedule of prices in the Tender, will be adjusted by a plus or minus percentage. The percentage already worked out during tender evaluation is worked out as follows: (*corrected tender price – tender price*)/*tender price* X 100.

16. Terms of Payment

The Supplier shall request for payment by submitting invoice(s), delivery note(s) and any other relevant documents as specified in the **SCC** to the Procuring Entity.

Payments shall be made promptly by the Procuring Entity, but not later than thirty (30) days after submission of an invoice by the Supplier, and after the Procuring Entity has accepted it.

Where a Procuring Entity Rejects Goods and Related Services, in part or wholly, the procuring Entity shall promptly inform the Supplier to collect, replace or rectify as appropriate and give reasons for rejection. The Supplier shall submit a fresh invoice, delivery note and any other relevant documents as specified in the **SCC**.

The currencies in which payments shall be made to the Supplier under this Contract shall be those in which the Tender price is expressed.

In the event that the Procuring Entity fails to pay the Supplier any payment by its due date or within the period set forth in the **SCC**, the Procuring Entity may pay to the Supplier interest on the amount of such delayed payment at the rate shown in the **SCC**, for the period of delay until payment has been made in full, whether before or after judgment or arbitration award.

17. Taxes and Duties

17.1 The Supplier shall be entirely responsible for all taxes, duties, license fees, and other such levies incurred to deliver the Goods and Related Services to the Procuring Entity at the final delivery point.

17.3 If any tax exemptions, reductions, allowances or privileges may be available to the Supplier in Kenya, the Supplier shall inform the Procuring Entity and the Procuring Entity shall use its best efforts to enable the Supplier to benefit from any such tax savings to the maximum allowable extent.

18. Performance Security

If required as specified in the **SCC**, the Supplier shall, within twenty-eight (28) days of the notification of contract award, provide a performance security for the performance of the Contract in the amount specified in the **SCC**.

The proceeds of the Performance Security shall be payable to the Procuring Entity as compensation for any loss resulting from the Supplier's failure to complete its obligations under the Contract.

As specified in **the SCC**, the Performance Security, if required, shall be denominated in the currency(ies) of the Contract, or in a freely convertible currency acceptable to the Procuring Entity; and shall be in one of the formats stipulated by the Procuring Entity in **the SCC**, or in another format acceptable to the Procuring Entity.

The Performance Security shall be discharged by the Procuring Entity and returned to the Supplier not later than thirty (30) days following the date of Completion of the Supplier's performance obligations under the Contract, including any warranty obligations, unless specified otherwise in the **SCC**.

19. Copyright

19.1 The copyright in all drawings, documents, and other materials containing data and information furnished to the Procuring Entity by the Supplier herein shall remain vested in the Supplier, or, if they are furnished to the Procuring Entity directly or through the Supplier by any third party, including suppliers of materials, the copyright in such materials shall remain vested in such third party.

20. Confidential Information

The Procuring Entity and the Supplier shall keep confidential and shall not, without the written consent of the other party hereto, divulge to any third party any documents, data, or other information furnished directly or indirectly by the other party hereto in connection with the Contract, whether such information has been furnished prior to, during or following completion or termination of the Contract. Notwithstanding the above, the Supplier may furnish to its Sub-Supplier such documents, data, and other information it receives from the Procuring Entity to the extent required for the Sub Supplier to perform its work under the Contract, in which event the Supplier shall obtain from such Sub Supplier undertaking of confidentiality similar to that imposed on the Supplier under GCC Clause 20.

The Procuring Entity shall not use such documents, data, and other information received from the Supplier for any purposes unrelated to the contract. Similarly, the Supplier shall not use such documents, data, and

other information received from the Procuring Entity for any purpose other than the performance of the Contract.

The obligation of a party under GCC Sub-Clauses 20.1 and 20.2 above, however, shall not apply to information that:

- a) the Procuring Entity or Supplier need to share with other arms of Government or other bodies participating in the financing of the Contract; such parties shall be disclosed in **the SCC**;
- b) now or hereafter enters the public domain through no fault of that party;
- c) can be proven to have been possessed by that party at the time of disclosure and which was not previously obtained, directly or indirectly, from the other party; or
- d) otherwise lawfully becomes available to that party from a third party that has no obligation of confidentiality.

The above provisions of GCC Clause 20 shall not in any way modify any undertaking of confidentiality given by either of the parties hereto prior to the date of the Contract in respect of the Supply or any part thereof.

The provisions of GCC Clause 20 shall survive completion or termination, for whatever reason, of the Contract.

21. Subcontracting

The Supplier shall notify the Procuring Entity in writing of all subcontracts awarded under the Contract if not already specified in the Tender. Such notification, in the original Tender or later shall not relieve the Supplier from any of its obligations, duties, responsibilities, or liability under the Contract.

Subcontracts shall comply with the provisions of GCC Clauses 3 and 7.

22. Specifications and Standards

Technical Specifications and Drawings

- a) The Goods and Related Services supplied under this Contract shall conform to the technical specifications and standards mentioned in Section VI, Schedule of Requirements and, when no applicable standard is mentioned, the standard shall be equivalent or superior to the official standards whose application is appropriate to the Goods' country of origin.
- b) The Supplier shall be entitled to disclaim responsibility for any design, data, drawing, specification or other document, or any modification thereof provided or designed by or on behalf of the Procuring Entity, by giving a notice of such disclaimer to the Procuring Entity.
- c) Wherever references are made in the Contract to codes and standards in accordance with which it shall be executed, the edition or the revised version of such codes and standards shall be those specified in the Schedule of Requirements. During Contract execution, any changes in any such codes and standards shall be applied only after approval by the Procuring Entity and shall be treated in accordance with GCC Clause 33.

23. Packing and Documents

The Supplier shall provide such packing of the Goods as is required to prevent their damage or deterioration during transit to their final destination, as indicated in the Contract. During transit, the packing shall be sufficient to withstand, without limitation, rough handling and exposure to extreme temperatures, salt and precipitation, and open storage. Packing case size and weights shall take into consideration, where appropriate, the remoteness of the goods' final destination and the absence of heavy handling facilities at all points in transit.

The packing, marking, and documentation within and outside the packages shall comply strictly with such special requirements as shall be expressly provided for in the Contract, including additional requirements, if any, specified **in the SCC**, and in any other instructions ordered by the Procuring Entity.

24. Insurance

24.1 Unless otherwise specified in the **SCC**, the Goods supplied under the Contract shall be fully insured—in a freely convertible currency from an eligible country—against loss or damage incidental to manufacture or acquisition, transportation, storage, and delivery, in accordance with the applicable Incoterms or in the manner specified in the **SCC**.

25. Transportation and Incidental Services

Unless otherwise specified in the **SCC**, responsibility for arranging transportation of the Goods shall be in accordance with the specified Incoterms.

The Supplier may be required to provide any or all of the following services, including additional services, if any, specified in **SCC**:

- a) performance or supervision of on-site assembly and/or start-up of the supplied Goods;
- b) furnishing of tools required for assembly and/or maintenance of the supplied Goods;
- c) furnishing of a detailed operations and maintenance manual for each appropriate unit of the supplied Goods;
- d) performance or supervision or maintenance and/or repair of the supplied Goods, for a period of time agreed by the parties, provided that this service shall not relieve the Supplier of any warranty obligations under this Contract; and
- e) training of the Procuring Entity's personnel, at the Supplier's plant and/or on-site, in assembly, start-up, operation, maintenance, and/or repair of the supplied Goods.

Prices charged by the Supplier for incidental services, if not included in the Contract Price for the Goods, shall be agreed upon in advance by the parties and shall not exceed the prevailing rates charged to other parties by the Supplier for similar services

26. Inspections and Tests

The Supplier shall at its own expense and at no cost to the Procuring Entity carry out all such tests and/or inspections of the Goods and Related Services as are specified in the **SCC**.

The inspections and tests may be conducted on the premises of the Supplier or its Subcontractor, at point of delivery, and/or at the Goods' final destination, or in another place in Kenya as specified in the **SCC**. Subject to GCC Sub-Clause 26.3, if conducted on the premises of the Supplier or its Subcontractor, all reasonable facilities and assistance, including access to drawings and production data, shall be furnished to the inspectors at no charge to the Procuring Entity.

The Procuring Entity or its designated representative shall be entitled to attend the tests and/or inspections referred to in GCC Sub-Clause 26.2, provided that the Procuring Entity bear all of its own costs and expenses incurred in connection with such attendance including, but not limited to, all travelling and board and lodging expenses.

Whenever the Supplier is ready to carry out any such test and inspection, it shall give a reasonable advance notice, including the place and time, to the Procuring Entity. The Supplier shall obtain from any relevant third party or manufacturer any necessary permission or consent to enable the Procuring Entity or its designated representative to attend the test and/or inspection.

The Procuring Entity may require the Supplier to carry out any test and/or inspection not required by the Contract but deemed necessary to verify that the characteristics and performance of the Goods comply with the technical specifications codes and standards under the Contract, provided that the Supplier's reasonable costs and expenses incurred in the carrying out of such test and/or inspection shall be added to the Contract Price. Further, if such test and/or inspection impedes the progress of manufacturing and/or the Supplier's performance of its other obligations under the Contract, due allowance will be made in respect of the Delivery Dates and Completion Dates and the other obligations so affected.

The Supplier shall provide the Procuring Entity with a report of the results of any such test and/or inspection.

The Procuring Entity may reject any Goods or any part thereof that fail to pass any test and/or inspection or do not conform to the specifications. The Supplier shall either rectify or replace such rejected Goods or parts thereof or make alterations necessary to meet the specifications at no cost to the Procuring Entity, and shall repeat the test and/or inspection, at no cost to the Procuring Entity, upon giving a notice pursuant to GCC Sub-Clause 26.4.

The Supplier agrees that neither the execution of a test and/or inspection of the Goods or any part thereof, nor the attendance by the Procuring Entity or its representative, nor the issue of any report pursuant to GCC Sub-Clause 26.6, shall release the Supplier from any warranties or other obligations under the Contract.

27. Liquidated Damages

27.1 Except as provided under GCC Clause 32, if the Supplier fails to deliver any or all of the Goods by the Date(s) of delivery or perform the Related Services within the period specified in the Contract, the Procuring Entity may without prejudice to all its other remedies under the Contract, deduct from the Contract Price, as liquidated damages, a sum equivalent to the percentage specified in the **SCC** of the delivered price of the delayed Goods or unperformed Services for each week or part thereof of delay until actual delivery or performance, up to a maximum deduction of the percentage specified in those **SCC**. Once the maximum is reached, the Procuring Entity may terminate the Contract pursuant to GCC Clause 35.

28. Warranty

The Supplier warrants that all the Goods are new, unused, and of the most recent or current models, and that they incorporate all recent improvements in design and materials, unless provided otherwise in the Contract.

Subject to GCC Sub-Clause 22.1(b), the Supplier further warrants that the Goods shall be free from defects arising from any act or omission of the Supplier or arising from design, materials, and workmanship, under normal use in the conditions prevailing in the country of final destination.

Unless otherwise specified in the **SCC**, the warranty shall remain valid for twelve (12) months after the Goods, or any portion thereof as the case may be, have been delivered to and accepted at the final destination indicated in the **SCC**, or for eighteen (18) months after the date of shipment from the port or place of loading in the country of origin, whichever period concludes earlier.

The Procuring Entity shall give notice to the Supplier stating the nature of any such defects together with all available evidence thereof, promptly following the discovery thereof. The Procuring Entity shall afford all reasonable opportunity for the Supplier to inspect such defects.

Upon receipt of such notice, the Supplier shall, within the period specified in the **SCC**, expeditiously repair or replace the defective Goods or parts thereof, at no cost to the Procuring Entity.

If having been notified, the Supplier fails to remedy the defect within the period specified in the **SCC**, the Procuring Entity may proceed to take within a reasonable period such remedial action as may be necessary, at the Supplier's risk and expense and without prejudice to any other rights which the Procuring Entity may have against the Supplier under the Contract.

29. Patent Indemnity

The Supplier shall, subject to the Procuring Entity's compliance with GCC Sub-Clause 29.2, indemnify and hold harmless the Procuring Entity and its employees and officers from and against any and all suits, actions or administrative proceedings, claims, demands, losses, damages, costs, and expenses of any nature, including attorney's fees and expenses, which the Procuring Entity may suffer as a result of any infringement or alleged infringement of any patent, utility model, registered design, trademark, copyright, or other intellectual property right registered or otherwise existing at the date of the Contract by reason of:

- a) the installation of the Goods by the Supplier or the use of the Goods in the country where the Site is located; and
- b) the sale in any country of the products produced by the Goods.

Such indemnity shall not cover any use of the Goods or any part thereof other than for the purpose indicated by or to be reasonably inferred from the Contract, neither any infringement resulting from the use of the Goods or any part thereof, or any products produced thereby in association or combination with any other equipment, plant, or materials not supplied by the Supplier, pursuant to the Contract.

If any proceedings are brought or any claim is made against the Procuring Entity arising out of the matters referred to in GCC Sub-Clause 29.1, the Procuring Entity shall promptly give the Supplier a notice thereof, and the Supplier may at its own expense and in the Procuring Entity's name conduct such proceedings or claim and any negotiations for the settlement of any such proceedings or claim.

If the Supplier fails to notify the Procuring Entity within twenty-eight (28) days after receipt of such notice that it intends to conduct any such proceedings or claim, then the Procuring Entity shall be free to conduct the same on its own behalf.

The Procuring Entity shall, at the Supplier's request, afford all available assistance to the Supplier in

conducting such proceedings or claim, and shall be reimbursed by the Supplier for all reasonable expenses incurred in so doing.

The Procuring Entity shall indemnify and hold harmless the Supplier and its employees, officers, and Subcontractors from and against any and all suits, actions or administrative proceedings, claims, demands, losses, damages, costs, and expenses of any nature, including attorney's fees and expenses, which the Supplier may suffer as a result of any infringement or alleged infringement of any patent, utility model, registered design, trademark, copyright, or other intellectual property right registered or otherwise existing at the date of the Contract arising out of or in connection with any design, data, drawing, specification, or other documents or materials provided or designed by or on behalf of the Procuring Entity.

30. Limitation of Liability

Except in cases of criminal negligence or willful misconduct,

a) the Supplier shall not be liable to the Procuring Entity, whether in contract, tort, or otherwise, for any indirect or consequential loss or damage, loss of use, loss of production, or loss of profits or interest costs, provided that this exclusion shall not apply to any obligation of the Supplier to pay liquidated damages to the Procuring Entity, and

b) the aggregate liability of the Supplier to the Procuring Entity, whether under the Contract, in tort or otherwise, shall not exceed the total Contract Price, provided that this limitation shall not apply to the cost of repairing or replacing defective equipment, or to any obligation of the supplier to indemnify the Procuring Entity with respect to patent infringement.

31. Change in Laws and Regulations

31.1 Unless otherwise specified in the Contract, if after the date of 30 days prior to date of Tender submission, any law, regulation, ordinance, order or bylaw having the force of law is enacted, promulgated, abrogated, or changed in Kenya (which shall be deemed to include any change in interpretation or application by the competent authorities) that subsequently affects the Delivery Date and/or the Contract Price, then such Delivery Date and/or Contract Price shall be correspondingly increased or decreased, to the extent that the Supplier has thereby been affected in the performance of any of its obligations under the Contract. Notwithstanding the foregoing, such additional or reduced cost shall not be separately paid or credited if the same has already been accounted for in the price adjustment provisions where applicable, in accordance with GCC Clause 15.

32. Force Majeure

The Supplier shall not be liable for forfeiture of its Performance Security, liquidated damages, or termination for default if and to the extent that its delay in performance or other failure to perform its obligations under the Contract is the result of an event of Force Majeure.

For purposes of this Clause, "Force Majeure" means an event or situation beyond the control of the Supplier that is not foreseeable, is unavoidable, and its origin is not due to negligence or lack of care on the part of the Supplier. Such events may include, but not be limited to, acts of the Procuring Entity in its sovereign capacity, wars or revolutions, fires, floods, epidemics, quarantine restrictions, and freight embargoes.

If a Force Majeure situation arises, the Supplier shall promptly notify the Procuring Entity in writing of such condition and the cause thereof. Unless otherwise directed by the Procuring Entity in writing, the Supplier shall continue to perform its obligations under the Contract as far as is reasonably practical, and shall seek all reasonable alternative means for performance not prevented by the Force Majeure event.

33. Change Orders and Contract Amendments

The Procuring Entity may at any time order the Supplier through notice in accordance GCC Clause 8, to make changes within the general scope of the Contract in any one or more of the following:

- a) drawings, designs, or specifications, where Goods to be furnished under the Contract are to be specifically manufactured for the Procuring Entity;
- b) the method of shipment or packing;
- c) the place of delivery; and
- d) the Related Services to be provided by the Supplier.

If any such change causes an increase or decrease in the cost of, or the time required for, the Supplier's

performance of any provisions under the Contract, an equitable adjustment shall be made in the Contract Price or in the Delivery/Completion Schedule, or both, and the Contract shall accordingly be amended. Any claims by the Supplier for adjustment under this Clause must be asserted within twenty-eight (28) days from the date of the Supplier's receipt of the Procuring Entity's change order.

Prices to be charged by the Supplier for any Related Services that might be needed but which were not included in the Contract shall be agreed upon in advance by the parties and shall not exceed the prevailing rates charged to other parties by the Supplier for similar services.

Value Engineering: The Supplier may prepare, at its own cost, a value engineering proposal at any time during the performance of the contract. The value engineering proposal shall, at a minimum, include the following;

- a) the proposed change(s), and a description of the difference to the existing contract requirements;
- b) a full cost/benefit analysis of the proposed change(s) including a description and estimate of costs (including life cycle costs) the Procuring Entity may incur in implementing the value engineering proposal; and
- c) a description of any effect(s) of the change on performance/functionality.

The Procuring Entity may accept the value engineering proposal if the proposal demonstrates benefits that:

- a) accelerates the delivery period; or
- b) reduces the Contract Price or the life cycle costs to the Procuring Entity; or
- c) improves the quality, efficiency or sustainability of the Goods; or
- d) yields any other benefits to the Procuring Entity, without compromising the necessary functions of the Facilities.

If the value engineering proposal is approved by the Procuring Entity and results in:

- a) a reduction of the Contract Price; the amount to be paid to the Supplier shall be the percentage specified in the SCC of the reduction in the Contract Price; or
- b) an increase in the Contract Price; but results in a reduction in life cycle costs due to any benefit described in
(a) to (d) above, the amount to be paid to the Supplier shall be the full increase in the Contract Price.

Subject to the above, no variation in or modification of the terms of the Contract shall be made except by written amendment signed by the parties.

34. Extensions of Time

If at any time during performance of the Contract, the Supplier or its subcontractors should encounter conditions impeding timely delivery of the Goods or completion of Related Services pursuant to GCC Clause 13, the Supplier shall promptly notify the Procuring Entity in writing of the delay, its likely duration, and its cause. As soon as practicable after receipt of the Supplier's notice, the Procuring Entity shall evaluate the situation and may at its discretion extend the Supplier's time for performance, in which case the extension shall be ratified by the parties by amendment of the Contract.

Except in case of Force Majeure, as provided under GCC Clause 32, a delay by the Supplier in the performance of its Delivery and Completion obligations shall render the Supplier liable to the imposition of liquidated damages pursuant to GCC Clause 26, unless an extension of time is agreed upon, pursuant to GCC Sub-Clause 34.1.

35. Termination

Termination for Default

- a) The Procuring Entity, without prejudice to any other remedy for breach of Contract, by written notice of default sent to the Supplier, may terminate the Contract in whole or in part:
 - i) if the Supplier fails to deliver any or all of the Goods within the period specified in the Contract, or within any extension thereof granted by the Procuring Entity pursuant to GCC Clause 34;
 - ii) if the Supplier fails to perform any other obligation under the Contract; or
 - iii) if the Supplier, in the judgment of the Procuring Entity has engaged in Fraud and Corruption, as defined in paragraph 2.2 a of the Appendix to the GCC, in competing for or in executing the Contract.
- b) In the event the Procuring Entity terminates the Contract in whole or in part, pursuant to GCC

Clause 35.1(a), the Procuring Entity may procure, upon such terms and in such manner as it deems appropriate, Goods or Related Services similar to those undelivered or not performed, and the Supplier shall be liable to the Procuring Entity for any additional costs for such similar Goods or Related Services. However, the Supplier shall continue performance of the Contract to the extent not terminated.

Termination for Insolvency.

The Procuring Entity may at any time terminate the Contract by giving notice to the Supplier if the Supplier becomes bankrupt or otherwise insolvent. In such event, termination will be without compensation to the Supplier, provided that such termination will not prejudice or affect any right of action or remedy that has accrued or will accrue thereafter to the Procuring Entity

Termination for Convenience.

- a) The Procuring Entity, by notice sent to the Supplier, may terminate the Contract, in whole or in part, at any time for its convenience. The notice of termination shall specify that termination is for the Procuring Entity's convenience, the extent to which performance of the Supplier under the Contract is terminated, and the date upon which such termination becomes effective.
- b) The Goods that are complete and ready for shipment within twenty-eight (28) days after the Supplier's receipt of notice of termination shall be accepted by the Procuring Entity at the Contract terms and prices. For the remaining Goods, the Procuring Entity may elect:
 - i) to have any portion completed and delivered at the Contract terms and prices; and/or
 - ii) to cancel the remainder and pay to the Supplier an agreed amount for partially completed Goods and Related Services and for materials and parts previously procured by the Supplier.

36. Assignment

- 36.1 Neither the Procuring Entity nor the Supplier shall assign, in whole or in part, their obligations under this Contract, except with prior written consent of the other party.

37. Export Restriction

Notwithstanding any obligation under the Contract to complete all export formalities, any export restrictions attributable to the Procuring Entity, to Kenya, or to the use of the products/goods, systems or services to be supplied, which arise from trade regulations from a country supplying those products/goods, systems or services, and which substantially impede the Supplier from meeting its obligations under the Contract, shall release the Supplier from the obligation to provide deliveries or services, always provided, however, that the Supplier can demonstrate to the satisfaction of the Procuring Entity that it has completed all formalities in a timely manner, including applying for permits, authorizations and licenses necessary for the export of the products/goods, systems or services under the terms of the Contract. Termination of the Contract on this basis shall be for the Procuring Entity's convenience pursuant to Sub-Clause 35.3.

SECTION VII - SPECIAL CONDITIONS OF CONTRACT

The following Special Conditions of Contract (SCC) shall supplement and/or amend the General Conditions of Contract (GCC). Whenever there is a conflict, the provisions herein shall prevail over those in the GCC.

[The Procuring Entity shall select insert the appropriate wording using the samples below or other acceptable wording, and delete the text in italics].

SECTION VII - SPECIAL CONDITIONS OF CONTRACT

The following Special Conditions of Contract (SCC) shall supplement and / or amend the General Conditions of Contract (GCC). Whenever there is a conflict, the provisions herein shall prevail over those in the GCC.

[The Procuring Entity shall select insert the appropriate wording using the samples below or other acceptable wording, and delete the text in italics]

Number of GC Clause	Amendments of, and Supplements to, Clauses in the General Conditions of Contract
GCC 1.1(h)	The Procuring Entity is: Kenyatta National Hospital
GCC 4.2 (a)	The meaning of the trade terms shall be as prescribed by Incoterms. If the meaning of any trade term and the rights and obligations of the parties there under shall not be as prescribed by Incoterms, they shall be as prescribed by: <i>[exceptional; refer to other internationally accepted trade terms]</i>
GCC 4.2 (b)	The version edition of Incoterms shall be <i>INCOTERMS 2015</i>
GCC 8.1	For notices , the Procuring Entity's address shall be: Attention: <i>To Chief executive officer]</i> Postal Address: [20723-00202 Nairobi Kenya] Physical Address: Nairobi City county Upperhill off Hospital, Kenyatta National Hospital Administration block, supply Chain Management Entrance. Telephone: [2726300-9] Electronic mail address: procurement@knh.or.ke
GCC 10.4.2	The place of arbitration shall be Nairobi Kenya
GCC 13.1	Details of Shipping and other Documents to be furnished by the Supplier are <i>[insert the required documents, such as a negotiable bill of lading, a non-negotiable sea way bill, an airway bill, a railway consignment note, a road consignment note, insurance certificate, Manufacturer's or Supplier's warranty certificate, inspection certificate issued by nominated inspection agency, Supplier's factory shipping details etc.]</i> . The above documents shall be received by the Procuring Entity before arrival of the Goods and, if not received, the Supplier will be responsible for any consequent expenses.
GCC 15.1	The prices charged for the Goods supplied and the related Services performed - Shall not be adjustable. If prices are adjustable, the following method shall be used to calculate the price adjustment ..N/A
GCC 16.5	The payment-delay period after which the Procuring Entity shall pay interest to the supplier shall be 90 days. The interest rate that shall be applied is ...N/A
GCC 18.1	A Performance Security of 1% Shall be required
GCC 18.3	If required, the Performance Security shall be in the form of a demand guarantee If required, the Performance security shall be denominated in <i>[insert "a freely convertible currency acceptable to the Procuring Entity" or "the currencies of payment of the Contract, in accordance with their portions of the Contract Price"]</i>
GCC 18.4 55	Discharge of the Performance Security shall take place: <i>[insert date if different from the one indicated in sub clause GCC 18.4]</i>

GCC 23.2	The packing, marking and documentation within and outside the packages shall be: <i>[insert in detail the type of packing required, the markings in the packing and all documentation required]</i>
GCC 24.1	The insurance coverage shall be as specified in the Incoterms. If not in accordance with Incoterms, insurance shall be as follows: <i>[insert specific insurance provisions agreed upon, including coverage, currency and amount]</i>
GCC 25.1	Responsibility for transportation of the Goods shall be as specified in the Incoterms.
	If not in accordance with Incoterms, responsibility for transportations shall be as follows: <i>[insert “The Supplier is required under the Contract to transport the Goods to a specified place of final destination within Kenya, defined as the Project Site, transport to such place of destination in Kenya, including insurance and storage, as shall be specified in the Contract, shall be arranged by the Supplier, and related costs shall be included in the Contract Price”; or any other agreed upon trade terms (specify the respective responsibilities of the Procuring Entity and the Supplier)]</i>
GCC 25.2	Incidental services to be provided are: <i>[Selected services covered under GCC Clause 25.2 and/or other should be specified with the desired features. The price quoted in the Tender price or agreed with the selected Supplier shall be included in the Contract Price.]</i>
GCC 26.1	The inspections and tests shall be: <i>[insert nature, frequency, procedures for carrying out the inspections and tests]</i>
GCC 26.2	The Inspections and tests shall be conducted at: <i>[insert name(s) of location(s)]</i>
GCC 27.1	The liquidated damage shall be: <i>[insert number]</i> % per week
GCC 27.1	The maximum amount of liquidated damages shall be: <i>[insert number]</i> %
GCC 28.3	<p>The period of validity of the Warranty shall be: <i>[insert number]</i> days For purposes of the Warranty, the place(s) of final destination(s) shall be: <i>[insert name(s) of location(s)]</i></p> <p>Sample provision</p> <p>GCC 28.3—In partial modification of the provisions, the warranty period shall be _____ hours of operation or _____ months from date of acceptance of the Goods or (_____) months from the date of shipment, whichever occurs earlier. The Supplier shall, in addition, comply with the performance and/or consumption guarantees specified under the Contract. If, for reasons attributable to the Supplier, these guarantees are not attained in whole or in part, the Supplier shall, at its discretion, either:</p> <p>(a) make such changes, modifications, and/or additions to the Goods or any part thereof as may be necessary in order to attain the contractual guarantees specified in the Contract at its own cost and expense and to carry out further performance tests in accordance with GCC 26.7,</p> <p>or</p> <p>(b) pay liquidated damages to the Procuring Entity with respect to the failure to meet the contractual guarantees. The rate of these liquidated damages shall be (_____).</p> <p><i>[The rate should be higher than the adjustment rate used in the Tender evaluation under TDS 34.6(f)]</i></p> <p>The period for repair or replacement shall be: <i>[insert number(s)]</i> days.</p>
GCC 28.5, GCC 28.6	
GCC 33.6	<p>If the value engineering proposal is approved by the Procuring Entity the amount to be paid to the Supplier shall be ___% (insert appropriate percentage).</p> <p>The percentage is normally up to 50%) of the reduction in the Contract Price.</p>

SECTION VIII - CONTRACT FORMS

This Section contains forms which, once completed, will form part of the Contract. The forms for Performance Security and Advance Payment Security, when required, shall only be completed by the successful tenderer after contract award.

FORM No. 1: NOTIFICATION OF INTENTION TO AWARD

This Notification of Intention to Award shall be sent to each Tenderer that submitted a Tender. Send this Notification to the Tenderer's Authorized Representative named in the Tender Information Form on the format below.

FORMAT

1. For the attention of Tenderer's Authorized Representative

- I) Name: _____ [insert Authorized Representative's name]
- ii) Address: _____ [insert Authorized Representative's Address]
- iii) Telephone: _____ [insert Authorized Representative's telephone/fax numbers]
- iv) Email Address: _____ [insert Authorized Representative's email address]

[IMPORTANT: insert the date that this Notification is transmitted to Tenderers. The Notification must be sent to all Tenderers simultaneously. This means on the same date and as close to the same time as possible.]

2. Date of transmission: _____ [email] on [date] _____ (local time)

This Notification is sent by _____ (Name and designation)

3. Notification of Intention to Award

- I) Employer: _____ [insert the name of the Employer]
- ii) Project: _____ [insert name of project]
- iii) Contract title: _____ [insert the name of the contract]
- iv) Country: _____ [insert country where ITT is issued]
- v) ITT No: _____ [insert ITT reference number from Procurement Plan]

This Notification of Intention to Award (Notification) notifies you of our decision to award the above contract. The transmission of this Notification begins the Standstill Period. During the Standstill Period, you may:

4. Request a debriefing in relation to the evaluation of your tender

Submit a Procurement-related Complaint in relation to the decision to award the contract.

- a) The successful tenderer
 - i) Name of successful Tender _____
 - ii) Address of the successful Tender _____
 - iii) Contract price of the successful Tender Kenya Shillings _____ (in words _____)
- b) Other Tenderers

Names of all Tenderers that submitted a Tender. If the Tender's price was evaluated include the evaluated price as well as the Tender price as read out. For Tenders not evaluated, give one main reason the Tender was unsuccessful.

S/No.	Name of Tender	Tender Price as read out	Tender's evaluated price (Note a)	One Reason Why Not Evaluated
1				
2				
3				
4				
5				

(Note a) State NE if not evaluated

5. How to request a debriefing?

- a) DEADLINE: The deadline to request a debriefing expires at midnight on *[insert date]* (local time).
- b) You may request a debriefing in relation to the results of the evaluation of your Tender. If you decide to request a debriefing your written request must be made within three (5) Business Days of receipt of this Notification of Intention to Award.
- c) Provide the contract name, reference number, name of the Tenderer, contact details; and address the request for debriefing as follows:
 - I) Attention: _____ *[insert full name of person, if applicable]*
 - ii) Title/position: _____ *[insert title/position]*
 - ii) Agency: _____ *[insert name of Employer]*
 - iii) Email address: _____ *[insert email address]*
- d) If your request for a debriefing is received within the 3 Days deadline, we will provide the debriefing within five (3) Business Days of receipt of your request. If we are unable to provide the debriefing within this period, the Standstill Period shall be extended by five (3) Days after the date that the debriefing is provided. If this happens, we will notify you and confirm the date that the extended Standstill Period will end.
- e) The debriefing may be in writing, by phone, video conference call or in person. We shall promptly advise you in writing how the debriefing will take place and confirm the date and time.
- f) If the deadline to request a debriefing has expired, you may still request a debriefing. In this case, we will provide the debriefing as soon as practicable, and normally no later than fifteen (15) Days from the date of publication of the Contract Award Notice.

6. How to make a complaint?

- a) Period: Procurement-related Complaint challenging the decision to award shall be submitted by midnight, *[insert date]* (local time).
- b) Provide the contract name, reference number, name of the Tenderer, contact details; and address the Procurement-related Complaint as follows:
 - I) Attention: _____ *[insert full name of person, if applicable]*
 - ii) Title/position: _____ *[insert title/position]*
 - iii) Agency: _____ *[insert name of Employer]*
 - iv) Email address: _____ *[insert email address]*
- c) At this point in the procurement process, you may submit a Procurement-related Complaint challenging the decision to award the contract. You do not need to have requested, or received, a debriefing before making this complaint. Your complaint must be submitted within the Standstill Period and received by us before the Standstill Period ends.
- d) Further information: For more information refer to the Public Procurement and Disposals Act 2015 and its Regulations available from the Website www.ppra.go.ke or email complaints@ppra.go.ke.

You should read these documents before preparing and submitting your complaint.

- e) There are four essential requirements:
 - i) You must be an ‘interested party’. In this case, that means a Tenderer who submitted a Tender in this tendering process, and is the recipient of a Notification of Intention to Award.
 - ii) The complaint can only challenge the decision to award the contract.
 - iii) You must submit the complaint within the period stated above.
 - iv) You must include, in your complaint, all of the information required to support your complaint.

7. Standstill Period

- i) **DEADLINE:** The Standstill Period is due to end at midnight on [*insert date*] (local time).
- ii) The Standstill Period lasts ten (14) Days after the date of transmission of this Notification of Intention to Award.
- iii) The Standstill Period may be extended as stated in paragraph Section 5 (d) above.

If you have any questions regarding this Notification please do not hesitate to contact us.

On behalf of the Employer:

Signature: _____

Name: _____ **Title/position:** _____

FORM NO. 2 - REQUEST FOR REVIEW

FORM FOR REVIEW(r.203(1))

PUBLIC PROCUREMENT ADMINISTRATIVE REVIEW BOARD

APPLICATION NO.....OF.....20.....

BETWEEN

.....**APPLICANT**

AND

.....**RESPONDENT (Procuring Entity)**

Request for review of the decision of the..... (Name of the Procuring Entity ofdated the...day of20.....in the matter of Tender No.....of20..... for (Tender description).

REQUEST FOR REVIEW

I/We.....,the above named Applicant(s), of address: Physical address..... P. O. Box No..... Tel. No.....Email....., hereby request the Public Procurement Administrative Review Board to review the whole/part of the above mentioned decision on the following grounds , namely:

- 1.
- 2.

By this memorandum, the Applicant requests the Board for an order/orders that: 1.

- 2.

SIGNED(Applicant) Dated on.....day of/...20.....

FOR OFFICIAL USE ONLY Lodged with the Secretary Public Procurement Administrative Review Board on.....day of20.....

SIGNED

Board Secretary

FORM NO. 3 LETTER OF AWARD

[Use letter head paper of the Procuring Entity]

_____ *[Date]*

To: _____ *[name and address of the Supplier]*

Subject: _____ *Notification of Award Contract No.*

This is to notify you that your Tender dated _____ *[insert date]* for execution of the _____ *[insert name of the contract and identification number, as given in the SCC]* for the Accepted Contract Amount of _____ *[insert amount in numbers and words and name of currency]*, as corrected and modified in accordance with the Instructions to tenderers is hereby accepted by our Agency.

You are requested to furnish the Performance Security within 30 days in accordance with the Conditions of Contract, using for that purpose the of the Performance Security Form included in Section X, Contract Forms, of the Tendering document.

Authorized Signature: _____ Name an

Attachment: Contract Agreement

FORM NO. 4 - CONTRACT AGREEMENT

[The successful tenderer shall fill in this form in accordance with the instructions indicated]

THIS AGREEMENT made the _____ *[insert: number]* day of _____ *[insert: month]*, *[insert: year]*. BETWEEN (1) _____ *[insert complete name of Procuring Entity]* and having its principal place of business at *[insert: address of Procuring Entity]* (hereinafter called "Procuring Entity"), of the one part; and (2) _____ *[insert name of Supplier]*, a corporation incorporated under the laws of *[insert: country of Supplier]* and having its principal place of business at _____ *[insert: address of Supplier]* (hereinafter called "the Supplier"), of the other part.

1. WHEREAS the Procuring Entity invited Tenders for certain Goods and ancillary services, viz., _____ *[insert]*
 - i) In this Agreement words and expressions shall have the same meanings as are _____ respectively assigned to them in the Contract documents referred to.
 - ii) The following documents shall be deemed to form and be read and construed as part of this Agreement. This Agreement shall prevail over all other contract documents.
 - a) the Letter of Acceptance
 - b) the Letter of Tender
 - c) the Addenda Nos. _____ (if any)
 - d) Special Conditions of Contract
 - e) General Conditions of Contract
 - f) the Specification (including Schedule of Requirements and Technical Specifications)
 - g) the completed Schedules (including Price Schedules)
 - h) any other document listed in GCC as forming part of the Contract
 - iii) In consideration of the payments to be made by the Procuring Entity to the Supplier as specified in this Agreement, the Supplier hereby covenants with the Procuring Entity to provide the Goods and Services and to remedy defects therein in conformity in all respects with the provisions of the Contract.
2. The Procuring Entity hereby covenants to pay the Supplier in consideration of the provision of the Goods and Services and the remedying of defects therein, the Contract Price or such other sum as may become payable under the provisions of the Contract at the times and in the manner prescribed by the Contract.
3. IN WITNESS whereof the parties hereto have caused this Agreement to be executed in accordance with the laws of Kenya on the day, month and year indicated above.

For and on behalf of the Procuring Entity

Signed: _____ *[insert signature]*

in the capacity of _____ *[insert title or other appropriate designation]* In the presence of _____
_____ *[insert identification of official witness]* **For and on behalf of the Supplier**

Signed: _____ *[insert signature of authorized representative(s) of the Supplier]* in the capacity of _____
_____ *[insert title or other appropriate designation]* in the presence of _____
_____ *[insert identification of official witness]*

FORM NO. 5 - PERFORMANCE SECURITY [Option 1 - Unconditional Demand Bank Guarantee]

[Guarantor letterhead]

Beneficiary: _____ *[insert name and Address of Employer]*

Date: _____ *[Insert date of issue]*

Guarantor: _____ *[Insert name and address of place of issue, unless indicated in the letterhead]*

1. We have been informed that _____ (hereinafter called "the Contractor") has entered into Contract No. _____ dated _____ with *(name of Employer)* _____ (the Employer as the Beneficiary), for the execution of _____ (hereinafter called "the Contract").
2. Furthermore, we understand that, according to the conditions of the Contract, a performance guarantee is required.
3. At the request of the Contractor, we as Guarantor, hereby irrevocably undertake to pay the Beneficiary any sum or sums not exceeding in total an amount of ___ *(in words)*,¹ such sum being payable in the types and proportions of currencies in which the Contract Price is payable, upon receipt by us of the Beneficiary's complying demand supported by the Beneficiary's statement, whether in the demand itself or in a separate signed document accompanying or identifying the demand, stating that the Applicant is in breach of its obligation(s) under the Contract, without the Beneficiary needing to prove or to show grounds for your demand or the sum specified therein.
4. This guarantee shall expire, no later than the Day of, 2.....², and any demand for payment under it must be received by us at the office indicated above on or before that date.
5. The Guarantor agrees to a one-time extension of this guarantee for a period not to exceed *[six months] [one year]*, in response to the Beneficiary's written request for such extension, such request to be presented to the Guarantor before the expiry of the guarantee."

[Name of Authorized Official, signature(s) and seals/stamps]

Note: All italicized text (including footnotes) is for use in preparing this form and shall be deleted from the final product.

FORM No. 6 - PERFORMANCE SECURITY [Option 2- Performance Bond]

[Note: Procuring Entities are advised to use Performance Security – Unconditional Demand Bank Guarantee instead of Performance Bond due to difficulties involved in calling Bond holder to action]

[Guarantor letterhead or SWIFT identifier code]

Beneficiary: _____ *[insert name and Address of Employer]* **Date:** _____ *[Insert date of issue]*

PERFORMANCE BOND No.: _____

Guarantor: _____ *[Insert name and address of place of issue, unless indicated in the letterhead]*

1. By this Bond _____ as Principal (hereinafter called “the Contractor”) and _____ as Surety (hereinafter called “the Surety”), are held and firmly bound unto _____ as Oblige (hereinafter called “the Employer”) in the amount of _____ for the payment of which sum well and truly to be made in the types and proportions of currencies in which the Contract Price is payable, the Contractor and the Surety bind themselves, their heirs, executors, administrators, successors and assigns, jointly and severally, firmly by these presents.
2. WHEREAS the Contractor has entered into a written Agreement with the Employer dated the _____ day of _____, 20____, for _____ in accordance with the documents, plans, specifications, and amendments thereto, which to the extent herein provided for, are by reference made part hereof and are hereinafter referred to as the Contract.
3. NOW, THEREFORE, the Condition of this Obligation is such that, if the Contractor shall promptly and faithfully perform the said Contract (including any amendments thereto), then this obligation shall be null and void; otherwise, it shall remain in full force and effect. Whenever the Contractor shall be, and declared by the Employer to be, in default under the Contract, the Employer having performed the Employer's obligations thereunder, the Surety may promptly remedy the default, or shall promptly:
 - 1) complete the Contract in accordance with its terms and conditions; or
 - 2) obtain a tender or tenders from qualified tenderers for submission to the Employer for completing the Contract in accordance with its terms and conditions, and upon determination by the Employer and the Surety of the lowest responsive Tenderers, arrange for a Contract between such Tenderer, and Employer and make available as work progresses (even though there should be a default or a succession of defaults under the Contract or Contracts of completion arranged under this paragraph) sufficient funds to pay the cost of completion less the Balance of the Contract Price; but not exceeding, including other costs and damages for which the Surety may be liable hereunder, the amount set forth in the first paragraph hereof. The term “Balance of the Contract Price,” as used in this paragraph, shall mean the total amount payable by Employer to Contractor under the Contract, less the amount properly paid by Employer to Contractor; or
 - 3) pay the Employer the amount required by Employer to complete the Contract in accordance with its terms and conditions up to a total not exceeding the amount of this Bond.
4. The Surety shall not be liable for a greater sum than the specified penalty of this Bond.
5. Any suit under this Bond must be instituted before the expiration of one year from the date of the issuing of the Taking-Over Certificate. No right of action shall accrue on this Bond to or for the use of any person or corporation other than the Employer named herein or the heirs, executors, administrators, successors, and assigns of the Employer.
6. In testimony whereof, the Contractor has hereunto set his hand and affixed his seal, and the Surety has caused these presents to be sealed with his corporate seal duly attested by the signature of his legal representative, this day _____ of _____ 20_____.

SIGNED ON _____ on behalf of _____

By _____ in the capacity of _____

In the presence of _____

SIGNED ON _____ on behalf of _____

By _____ in the capacity of _____

In the presence of _____

FORM NO. 7 - ADVANCE PAYMENT SECURITY [Demand Bank Guarantee]

[Guarantor letterhead]

Beneficiary: _____ [Insert
name and Address of Employer]

Date: _____ [Insert date of issue]

ADVANCE PAYMENT GUARANTEE No.: _____ [Insert guarantee reference
number]

Guarantor: [Insert name and address of place of issue, unless indicated in the letterhead]

1. We have been informed that _____ (hereinafter called "the Contractor") has entered into Contract No. _____ dated _____ with the Beneficiary, for the execution of _____ (hereinafter called "the Contract").

2. Furthermore, we understand that, according to the conditions of the Contract, an advance payment in the sum (in words) is to be made against an advance payment guarantee.

3. At the request of the Contractor, we as Guarantor, hereby irrevocably undertake to pay the Beneficiary any sum or sums not exceeding in total an amount of _____ (in words _____) ¹ upon receipt by us of the Beneficiary's complying demand supported by the Beneficiary's statement, whether in the demand itself or in a separate signed document accompanying or identifying the demand, stating either that the Applicant:

- (a) has used the advance payment for purposes other than the costs of mobilization in respect of the goods; or
- (b) has failed to repay the advance payment in accordance with the Contract conditions, specifying the amount which the Applicant has failed to repay.

4. A demand under this guarantee may be presented as from the presentation to the Guarantor of a certificate from the Beneficiary's bank stating that the advance payment referred to above has been credited to the Contractor on its account number _____ at -----.

5. The maximum amount of this guarantee shall be progressively reduced by the amount of the advance payment repaid by the Contractor as specified in copies of interim statements or payment certificates which shall be presented to us. This guarantee shall expire, at the latest, upon our receipt of a copy of the interim payment certificate indicating that ninety (90) percent of the Accepted Contract Amount, less provisional sums, has been certified for payment, or on the ___ day of _____, 2nd,² whichever is earlier. Consequently, any demand for payment under this guarantee must be received by us at this office on or before that date.

6. The Guarantor agrees to a one-time extension of this guarantee for a period not to exceed [six months] [one year], in response to the Beneficiary's written request for such extension, such request to be presented to the Guarantor before the expiry of the guarantee.

[Name of Authorized Official, signature(s) and seals/stamps]

Note: All italicized text (including footnotes) is for use in preparing this form and shall be deleted from the final product.

¹The Guarantor shall insert an amount representing the amount of the advance payment and denominated either in the currency of the advance payment as specified in the Contract.

² Insert the expected expiration date of the Time for Completion. The Employer should note that in the event of an extension of the time for completion of the Contract, the Employer would need to request an extension of this guarantee from the Guarantor. Such request must be in writing and must be made prior to the expiration date established in the guarantee.

FORM NO. 4 BENEFICIAL OWNERSHIP DISCLOSURE FORM
(Amended and issued pursuant to PPRA CIRCULAR No. 02/2022)

INSTRUCTIONS TO TENDERERS: DELETE THIS BOX ONCE YOU HAVE COMPLETED THE FORM

This Beneficial Ownership Disclosure Form ("Form") is to be completed by the successful tenderer pursuant to Regulation 13 (2A) and 13 (6) of the Companies (Beneficial Ownership Information) Regulations, 2020. In case of joint venture, the tenderer must submit a separate Form for each member. The beneficial ownership information to be submitted in this Form shall be current as of the date of its submission.

For the purposes of this Form, a Beneficial Owner of a Tenderer is any natural person who ultimately owns or controls the legal person (tenderer) or arrangements or a natural person on whose behalf a transaction is conducted, and includes those persons who exercise ultimate effective control over a legal person (Tenderer) or arrangement.

Tender Reference No.: _____ [insert identification no]

Name of the Tender Title/Description: _____ [insert name of the assignment] to:
 _____ [insert complete name of Procuring Entity]

In response to the requirement in your notification of award dated _____ [insert date of notification of award] to furnish additional information on beneficial ownership: _____ [select one option as applicable and delete the options that are not applicable]

I) We here by provide the following beneficial ownership information.

Details of beneficial ownership

	Details of all Beneficial Owners	% of shares a person holds in the company Directly or indirectly	% of voting rights a person holds in the company	Whether a person directly or indirectly holds a right to appoint or remove a member of the board of directors of the company or an equivalent governing body of the Tenderer (Yes / No)	Whether a person directly or indirectly exercises significant influence or control over the Company (tenderer) (Yes / No)
1.	Full Name	Directly -----% of shares	Directly..... % of voting rights	1. Having the right to appoint a majority of the board of the directors or an equivalent governing body of the Tenderer: Yes -----No---- 2. Is this right held directly or indirectly?: Direct..... Indirect.....	1. Exercises significant influence or control over the Company body of the Company (tenderer) Yes ----- No-- -- 2. Is this influence or control exercised directly or indirectly? Direct.....
	National identity card number or Passport number				
	Personal Identification Number (where applicable)	Indirectly -----% of shares	Indirectly----- % of voting rights		
	Nationality				
	Date of birth [dd/mm/yyyy]				
	Postal address				
	Residential address				
	Telephone number				
	Email address				
Occupation or profession					

Details of all Beneficial Owners		% of shares a person holds in the company Directly or indirectly	% of voting rights a person holds in the company	Whether a person directly or indirectly holds a right to appoint or remove a member of the board of directors of the company or an equivalent governing body of the Tenderer (Yes / No)	Whether a person directly or indirectly exercises significant influence or control over the Company (tenderer) (Yes / No)
					Indirect..... ...
2.	Full Name	Directly-----% of shares	Directly.....% of voting rights	1. Having the right to appoint a majority of the board of the directors or an equivalent governing body of the Tenderer: Yes ----No---- 2. Is this right held directly or indirectly?: Direct..... Indirect.....	1. Exercises significant influence or control over the Company body of the Company (tenderer) Yes ---- No-- -- 2. Is this influence or control exercised directly or indirectly? Direct..... Indirect.....
	National identity card number or Passport number				
	Personal Identification Number (where applicable)	Indirectly-----% of shares	Indirectly----- % of voting rights		
	Nationality(ies)				
	Date of birth [dd/mm/yyyy]				
	Postal address				
	Residential address				
	Telephone number				
	Email address				
	Occupation or profession				
3.					
e.t					
.c					

II) Am fully aware that beneficial ownership information above shall be reported to the Public Procurement Regulatory Authority together with other details in relation to contract awards and shall be maintained in the Government Portal, published and made publicly available pursuant to Regulation 13(5) of the Companies (Beneficial Ownership Information) Regulations, 2020. (Notwithstanding this paragraph Personally Identifiable Information in line with the Data Protection Act shall not be published or made public). *Note that Personally Identifiable Information (PII) is defined as any information that can be used to distinguish one person from another and can be used to deanonymize previously anonymous data. This information includes National identity card number or Passport number, Personal Identification Number, Date of birth, Residential address, email address and Telephone number.*

III) In determining who meets the threshold of who a beneficial owner is, the Tenderer must consider a natural person who in relation to the company:

- (a) holds at least ten percent of the issued shares in the company either directly or indirectly;
- (b) exercises at least ten percent of the voting rights in the company either directly or indirectly;
- (c) holds a right, directly or indirectly, to appoint or remove a director of the company; or
- (d) exercises significant influence or control, directly or indirectly, over the company.

IV) What is stated to herein above is true to the best of my knowledge, information and belief.

Name of the Tenderer[insert complete name of the Tenderer] _____*

*Name of the person duly authorized to sign the Tender on behalf of the Tenderer: ** [insert complete name of person duly authorized to sign the Tender]*

Designation of the person signing the Tender [insert complete title of the person signing the Tender]

Signature of the person named above..... [insert signature of person whose name and capacity are shown above]

Date this [insert date of signing] day of.....[Insert month], [insert year]



KENYATTA NATIONAL HOSPITAL RECOMMENDATION FORM

TENDER NO: KNH/T/...../2026-2027 - SUPPLY AND DELIVERY OF CALIBRATION TOOLS (Reserved for Youth, Women and Persons Living with Disability)

Bidder Name.....

Address:

This is to confirm that the above mentioned bidder has supplied and delivered similar items to our institution satisfactorily.

Name of Institution.....	Name of Institution.....	Name of Institution.....
Contact Person.....	Contact Person.....	Contact Person.....
Designation.....	Designation.....	Designation.....
Phone Number.....	Phone Number.....	Phone Number.....
Sign.....	Sign.....	Sign.....
Stamp.....	Stamp.....	Stamp.....

